

DPS - DMT

From: Driscoll, Rob
Sent: Wednesday, January 06, 2016 9:57 AM
To: DPS - DMT
Subject: Comm. Log (Fair Haven/Rutland VSP)

1/6/16 – Contacted Fair Haven PD and asked to have a note placed on the DMT indicating it is out of service. Also left a voicemail for Blake Cushing with Rutland VSP stating that the instrument is out of service until further notice due to an oscillating detector voltage.

Robert Driscoll
Vermont Forensics Laboratory
45 State Drive
Waterbury, VT 05671
Electronics Technician
Rob.Driscoll@Vermont.gov
Desk: (802) 241-5817
Cell: (802) 585-0123