

## DPS - DMT

---

**From:** Driscoll, Rob  
**Sent:** Tuesday, February 16, 2016 9:20 AM  
**To:** DPS - DMT  
**Subject:** COMM LOG 100137

Contacted Grand Isle County SO after reviewing the Feb. RPC and found the Lot number entered as 12-25-100. Spoke with Dustin Abell (DMT Supervisor). He checked the reports to verify that no tests were run between the RPC (2/13/16) and now (2/16/16). He is going to run a Solution change protocol to enter the correct lot number (15-25-100).

Robert Driscoll  
Vermont Forensics Laboratory  
45 State Drive  
Waterbury, VT 05671  
Electronics Technician  
[Rob.Driscoll@Vermont.gov](mailto:Rob.Driscoll@Vermont.gov)  
Desk: (802) 241-5817  
Cell: (802) 585-0123