

DPS - DMT

From: Driscoll, Rob
Sent: Tuesday, April 26, 2016 10:46 AM
To: DPS - DMT
Subject: Call Log 4/26/16

While checking DM Host, I found that Windham CSO – Vernon had performed a simulator solution change with an expired bottle of simulator solution (lot # 15-05-100). I called Holly Ellis and asked her to place the instrument out of service until which time a solution change could be completed using unexpired simulator solution. According to Host, there were no tests performed after the invalid solution change.

Robert Driscoll
Vermont Forensics Laboratory
45 State Drive
Waterbury, VT 05671
Electronics Technician
Rob.Driscoll@Vermont.gov
Desk: (802) 241-5817
Cell: (802) 585-0123