

DPS - DMT

From: Driscoll, Rob
Sent: Wednesday, May 11, 2016 8:53 AM
To: DPS - DMT
Subject: Call Log 5/9/16 Orange County SO

Called Orange County Sheriff's Office and left a message with Holly Ellis (one of the DMT supervisors), stating that we had not seen a solution change as being completed on Vernon's DMT (100141) and to be certain the instrument remains out of service until which time a solution change is completed.

Robert Driscoll
Vermont Forensics Laboratory
45 State Drive
Waterbury, VT 05671
Electronics Technician
Rob.Driscoll@Vermont.gov
Desk: (802) 241-5817
Cell: (802) 585-0123