

## DPS - DMT

---

**From:** Driscoll, Rob  
**Sent:** Monday, October 05, 2015 8:18 AM  
**To:** DPS - DMT  
**Subject:** Comm Log (Essex County SO)

-Found RPC was completed with expired simulator solution (Lot # 14-10-100).  
-Checked DM Host to verify no tests had been ran since the RPC date (10/3/15).  
-Called Essex County SO and explained that Gil Rainault had used an expired simulator solution and to put the instrument out of service immediately. Their DMT supervisor is out of town on vacation, but they may have a DMT supervisor from another agency complete another RPC with non-expired solution. At this point the instrument is out of service and we are awaiting a passing RPC from the agency.

Robert Driscoll  
Vermont Forensics Laboratory  
103 South Main Street  
Waterbury, VT 05671  
Electronics Technician  
[Rob.Driscoll@Vermont.gov](mailto:Rob.Driscoll@Vermont.gov)  
Desk: (802) 241-5817  
Cell: (802) 585-0123