

DPS - DMT

From: Driscoll, Rob
Sent: Monday, November 16, 2015 9:04 AM
To: DPS - DMT
Subject: COMM LOG (Bennington CSO)

11/16 – Found that wrong concentration was entered for the simulator solution during the RPC performed on 11/13. Tried several contacts at the agency and received voicemail. Left a message with Dispatch to immediately remove the instrument from service and contact me as soon as possible.

Robert Driscoll
Vermont Forensics Laboratory
45 State Drive
Waterbury, VT 05671
Electronics Technician
Rob.Driscoll@Vermont.gov
Desk: (802) 241-5817
Cell: (802) 585-0123