

DPS - DMT

From: Dukette, Jeffery
Sent: Friday, February 12, 2016 12:49 PM
To: DPS - DMT
Subject: COMM LOG

FYI

From: Dukette, Jeffery
Sent: Friday, February 12, 2016 12:48 PM
To: Tiersch, Steven
Subject: RE: DataMaster DMT RPC

Officer Tiersch,

Thanks for getting back to me. I'll make a note on the RPC. In the future that would be great if you could send an email to dps.dmt@vermont.gov if there are any errors that occur during an RPC. We will contact you regardless, but being proactive and sending a message while it's fresh in your mind is a good idea.

Have a great weekend,

Jeff

From: Tiersch, Steven
Sent: Friday, February 12, 2016 12:43 PM
To: Dukette, Jeffery
Subject: RE: DataMaster DMT RPC

Good afternoon Jeffery,

I accidentally hit the cancel button during the test. It was an error on my part not the part of the DMT. If you have any further questions please let me know. In the future if there is an error like this would you like me to email you?

Very respectfully,

Steven Tiersch

Officer Steven Tiersch
Patrol Officer
Randolph Police Department
6 Salisbury Street
Randolph VT, 05060
802-728-3737
802-728-2512 (fax)

From: Dukette, Jeffery
Sent: Thursday, February 11, 2016 10:08 AM
To: Tiersch, Steven
Subject: DataMaster DMT RPC

Officer Tiersch,

I am writing to follow up on the RPC you performed on the DataMaster on 2/5/16. I noticed there were two attempts at the RPC, with the first attempt having failed the RF detection test. Can you please let me know why this test failed or why you cancelled the test at this stage?

Thank you,

Jeff Dukette

Forensic Chemist

Vermont Forensic Laboratory

45 State Drive

Waterbury, VT 05671

(802) 241-5275

Jeffery.Dukette@vermont.gov