

DPS - DMT

From: Dukette, Jeffery
Sent: Thursday, March 03, 2016 8:16 AM
To: DPS - DMT
Subject: DMT Comm Log

FYI

From: Dukette, Jeffery
Sent: Wednesday, March 02, 2016 8:43 AM
To: 'Michael Reeve'
Subject: RE: DMT Routine Performance Check

Sounds good. Thanks

From: Michael Reeve [<mailto:MReeve@morristownvt.org>]
Sent: Wednesday, March 02, 2016 8:43 AM
To: Dukette, Jeffery
Subject: Re: DMT Routine Performance Check

Jeff

I saw that blip after I printed the ticket. It did not show during the test. I will complete the test today and email you a color copy this afternoon.

Sent from my iPhone

On Mar 2, 2016, at 08:40, Dukette, Jeffery <Jeffery.Dukette@vermont.gov> wrote:

Cpl. Reeve,

I am following up on the RPC you performed on 3/1/16 for the DMT. It looks like there was a small blip in the alcohol line during the shallow breath of the sample acceptance test. Can you please run a Check-In with a passing sample acceptance test? When you run the Check-In, use "RPC TEST" for the first and last name to avoid any confusion. Please let me know if you have any questions.

Thank you,

Jeff Dukette
Forensic Chemist
Vermont Forensic Laboratory
45 State Drive
Waterbury, VT 05671
(802) 241-5275
Jeffery.Dukette@vermont.gov