

DATAMASTER DMT MAINTENANCE LOG

VSP-Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/21/11	<p>DataMaster Simulator Solution Lot # 12-45-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 12/20/11 SH <i>sun 12/20/11</i></p> <p style="text-align: center; font-size: 1.2em;">Installation <i>sun 12/21/11</i></p>	A. Bobb
2-3-12	<p>DataMaster Simulator Solution Lot # 12-32-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 2-3-12</p> <p style="text-align: center; font-size: 1.2em;">RPC</p>	<i>Fran</i>
02-12-12	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center; font-size: 1.2em;">Diagnostic Test Performed Accuracy - Precision Test</p>	<i>Carroll</i>
5-19-10	<p style="text-align: center; font-size: 1.2em;">ACCURACY & PRECISION TEST</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>FALCON</i>

DATAMASTER DMT MAINTENANCE LOG

VSP-Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/28/12	Accuracy & Precision Test [For RPC or Simulator Solution Change Place Label Here]	Ambroz
5/31/12	Moved DMT to new room, updated software, changed lock Install protocol Lot # 12-45-100	L. Bolz
	TV = 0.100 exp 12/12/12 [For RPC or Simulator Solution Change Place Label Here]	
5-9-12	Accuracy & Precision Test	F. Kellon
8/3/12	Accuracy and Precision test [For RPC or Simulator Solution Change Place Label Here] Lot # 12-45-100 0.100% Passed w/ Avg .096%	Ambroz 244
9/17/12	[For RPC <u>DataMaster Simulator Solution</u> ce Label Here] Lot # 12-73-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013	Ambroz 244
" "	OPENED: 9/17/12 by TDA/244	" "
	Test/Simulator Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/30/12	Routine Performance Check [For RPC or Simulator Solution Change Place Label Here] Done as Required for month of October All systems GO	Ambroz 244
10/20/12	Accuracy & Precision Test [For RPC or Simulator Solution Change Place Label Here]	RAvelin
10/20/12	Calibration test - month of October [For RPC or Simulator Solution Change Place Label Here]	RAvelin
11/28/12	Accuracy & Precision Test [For RPC or Simulator Solution Change Place Label Here] <u>AND</u> Calibration Test for month of November	Ambroz 244

DATAMASTER DMT MAINTENANCE LOG

VSP Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/14/13	DIAGNOSTIC TEST - All OK [For RPC or Simulator Solution Change Place Label Here] changed ink cartridges in printer	S/T Ambroz 244
2/1/13	accuracy/precision test [For RPC or Simulator Solution Change Place Label Here]	S/T Ambroz 244
	[For RPC or Simulator Solution Change Place Label Here]	
2/19/13	test/simulator solution change [For RPC or Simulator Solution Change Place Label Here] EXT STD 1	S/T Ambroz 244
	<p style="text-align: center;">DataMaster Simulator Solution</p> Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 OPENFD: 2/19/13 TDA/244	

DATAMASTER DMT MAINTENANCE LOG

VSP WILLISTON

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/19/13	<p style="text-align: center; font-size: 1.2em;">RPC</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">Ambroz 244 DMT SUPERVISOR</p>
5-12-13	<div style="border: 1px solid black; padding: 5px; margin-bottom: 10px;"> <p>DataMaster Simulator Solution Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013</p> <p>OPENED: 5-12-13</p> </div> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center; font-size: 1.2em;">SIM SOLUTION CHANGE</p>	<p style="text-align: center; font-size: 1.2em;">RAVELIN</p>
5/23/13	<p style="text-align: center; font-size: 1.2em;">Performed APM protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">R. Driscoll DMT TECH</p>
6/12/13	<p style="text-align: center; font-size: 1.2em;">RPC Performed - All Good - PASSED</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">Ambroz/244 DMT SUPERVISOR</p>

DATAMASTER DMT MAINTENANCE LOG

VSP williston
 Department/Agency

100132
 DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/18/13	<p style="text-align: center;">DIAGNOSTIC TEST - All Good</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	S/T Ambroz Supervisor
8/7/13	<p style="text-align: center;">DIAGNOSTIC TEST - All Good - Passed</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>RPC completed - Passed</p>	S/T Ambroz DMT Supervisor
10/31/13	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 10/31/13 Ambroz</p> <p style="text-align: center;">simulator solution change and RPC</p>	S/T Ambroz DMT Supervisor
2/3/14	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 2/3/14 RCO</p> <p style="text-align: center;">Replaced broken simulator fiddle + completed February RPC</p>	R. Dussell DMT TECH.

DATAMASTER DMT MAINTENANCE LOG

VSP Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/19/14	<p>DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: 5/19/14 TDA/204</p>	<p>SGT Ambroz DMT Supervisor</p>
5/19/14	<p>Simulator solution change and RPC Failed 2x due to seal not tight. PASSED after tightening seal</p>	<p>DMT Supervisor</p>
5/27/14	<p>Completed Arm protocol [For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT TECH.</p>
10/18/14	<p>DataMaster Simulator Solution Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 3/12/2014 Expires: 3/12/2015 OPENED: 10/18/14 TDA/204</p>	<p>RPC Performed Simulator solution Changed SGT TODD Ambroz DMT Supervisor</p>
2/14/15	<p>DataMaster Simulator Solution Lot # 14-12-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/10/2014 Expires: 9/10/2015 OPENED: 2/14/15 TDA/204</p>	<p>RPC Performed Simulator solution Changed SGT TODD Ambroz DMT Supervisor</p>
3/2/15	<p>Calibration test performed</p>	<p>SGT TODD Ambroz DMT Supervisor</p>
5/22/16	<p>DIAGNOSTIC TEST 10:50AM [For RPC or Simulator Solution Change Place Label Here]</p>	<p>SGT Ambroz DMT Supervisor</p>

DATAMASTER DMT MAINTENANCE LOG

Williston VSP
Department/Agency

100132
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/4/15	<p>Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT Tech.</p>
6/12/15	<p>RPC Performed / Simulator Solution change</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 14-12-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/10/2014 Expires: 9/10/2015 OPENED: <u>DA/290</u></p>	<p>TODD Ambroz DMT Supervisor</p>
8/3/15	<p>Downloaded data to USB drive, uploaded new software, and installed new comm. PCB. If appears the port is not active so will contact DIRT.</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT Tech.</p>
10/23/15	<p>RPC Performed For RP</p> <p>DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: <u>DA/290 on 10/23/15</u></p>	<p>simulator solution changed</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>TODD Ambroz DMT Supervisor</p>

DATAMASTER DMT MAINTENANCE LOG

VSP Williston

100132

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/25/2016	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 [For RP Place Label Here] 2/16/2016 1/25/2016 OPENED: RPC PERFORMED</p> <p style="text-align: center;">SIMULATOR SOLUTION CHANGE ↳ WORK 3/2/16, WW</p>	<p>Sgt. Wuzam Warner DMT SUPERVISOR</p>
3/2/2016	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 [For RP Place Label Here] SIM SOL. CHANGED, RPC PERFORMED OPENED: 3/2/2016 WARNER, 201</p> <p style="text-align: center;">SIMULATOR SOLUTION CHANGE RPC PERFORMED</p>	<p>Sgt. Wuzam Warner DMT SUPERVISOR</p>
6/29/2016	<p>DataMaster External Standard Solution Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016 [For RPC Place Label Here] OPENED: 6/29/16 WARNER, 201</p> <p style="text-align: center;">SIMULATOR SOLUTION CHANGE R.P.C. PERFORMED</p>	<p>Sgt. Wuzam Warner DMT SUPERVISOR</p>
7/12/16	<p style="text-align: center;">Completed APR Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT Tech</p>

DATAMASTER DMT MAINTENANCE LOG

VSP, WALTON

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	<p>DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 [For RPC e Label Here] OPENED: <u>10/31/16</u> <u>WARNER, 204</u></p>	
10/31/16	<p>RPC + SIMULATOR SOLUTION CHANGE ATTEMPTED MULTIPLE ERROR MESSAGES RECEIVED "SIMULATOR VAPOR MESSAGE TIMED OUT"</p>	
11/1/16	<p>Found leak in simulator jar lid orig. Fixed leak [For RPC or Simulator Solution Change Place Label Here] and recal RPC protocol.</p>	R. Driscoll DMT Tech
2/21/2017	<p>DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 [For RPC e Label Here] OPENED: <u>2/21/2017</u> <u>WARNER, 204</u> RPC PERFORMED, SOLUTION CHANGED</p>	WARNER, 204
2/23/17	<p>DMT communication problems. Ethernet port keeps [For RPC or Simulator Solution Change Place Label Here] getting deactivated. Replaced I/O PCB and will have DIT reactivate the port.</p>	R. Driscoll DMT Tech

DATAMASTER DMT MAINTENANCE LOG

Williston VSP

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/21/17 JAV	<p>DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 3/21/17 RCD</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">Install post simulator update</p>	R. Discoll DMT Tech.
6/30/17 JAV	<p>DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED: 6/30/17 WW, 204</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">SOL. CHANGED, RPC PERFORMED</p>	WARNER 31852
8/3/17 ↑	<p style="text-align: center;">TIME + DATE CHANGED</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">Accuracy + Precision Report DIAGNOSTIC TEST (6/3/17)</p>	WARNER
8/4/17	SHUT DOWN + RE-STARTED PER VFL STAFF	WARNER
9/8/17	<p>Printer failed. Replaced with new printer and [For RPC or Simulator Solution Change Place Label Here]</p>	R. Discoll VFL
10/30/17	<p>DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 10/30/17, WW, 204</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">Solution CHANGED RPC PERFORMED</p>	WARNER

DATAMASTER DMT MAINTENANCE LOG

Williston VSP

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/17	<p style="text-align: center;"><i>Installed version 2.00 PIC chip.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
2/23/18	<p style="text-align: center;"><i>Completed APM protocol.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
2/27/2018	<p style="text-align: center;">DMT Simulator Solution</p> <p>Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018</p> <p>[For RF Place Label Here]</p> <p>OPENED: <i>2/27/2018 Warner 204</i></p> <p style="text-align: center;"><i>RPC</i></p>	<p><i>Warner, 204</i></p>
7/3/18	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19</p> <p>[For F Place Label Here]</p> <p>Opened:</p> <p style="text-align: center;"><i>RPC</i></p>	<p><i>Miller 229</i></p>

DATAMASTER DMT MAINTENANCE LOG

VSP WILLISTON
Department/Agency

1002381
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
16/22/18	<p style="text-align: center;">DMT External Std Solution</p> <p>[For Lot: 18-11-100 EtOH Certified Conc. 0.101 Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/26/18 Exp Date: 06/26/19</p> <p>Opened: 10/22/18</p>	<p>TPR. B. Miller DMT SUPERVISOR</p>
2/18/19	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R Lot: 18-12-100 EtOH Certified Conc. 0.100 Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 7/11/18 Exp Date: 07/11/19</p> <p>Opened: 2/18/19 B. Miller 229</p>	<p>TPR. B. Miller DMT SUPERVISOR</p>
2/19/19	<p style="text-align: center;">Completed APM protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>J. Dukette (UFL)</p>
6/26/19	<p style="text-align: center;">DMT External Std Solution</p> <p>[For Lot: 18-12-100 EtOH Certified Conc. 0.100 Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 7/11/18 Exp Date: 07/11/19</p> <p>Opened: 6/26/19 B. Miller 229</p>	<p>TPR. B. Miller DMT SUPERVISOR</p>

DATAMASTER DMT MAINTENANCE LOG

VSP Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/3/19	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RP Lot: 19-01-100 EtOH Certified Conc. 0.101 Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 1/15/19 Exp Date: 01/15/20</p> <p>Opened: 10/3/19 B. Miller 229</p>	<p>TPR B. MILLER DMT SUPERVISOR</p>
2/5/2020	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC Lot: 19-10-100 EtOH Certified Conc. 0.101 Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/17/19 Exp Date: 06/17/20</p> <p>Opened: 2/5/2020 B. Miller 229</p>	<p>TPR B. MILLER DMT SUPERVISOR</p>
2/14/2020	<p style="color: blue;">Completed APM protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="color: blue;">J Dukette (UFL)</p>
6/4/2020	<p style="text-align: center;">DMT External Std Solution</p> <p>[For F Lot: 19-10-100 EtOH Certified Conc. 0.101 Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/17/19 Exp Date: 06/17/20</p> <p>Opened: 6/4/2020 B. Miller 229</p>	<p>TPR B. MILLER DMT SUPERVISOR</p>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/15/2020	<p><i>Installed DMT software update.</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>S Stratton (VFL)</i></p>
11/21/2020	<div style="border: 1px solid black; padding: 5px; margin: 5px;"> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 20-03-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 1/29/20 Exp Date: 01/29/21</p> <p>Opened: <i>11/21/2020 Szarejko 264</i></p> </div> <p>[For _____ Place Label Here]</p>	<p><i>Tpr. J. Szarejko DMT SUPERUSER</i></p>
2/1/21	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>J Dukette (VFL)</i></p>
2-18-21	<div style="border: 1px solid black; padding: 5px; margin: 5px;"> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 20-14-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 9/22/20 Exp Date: 09/22/22</p> <p>Opened: <i>2-18-21 Underwood</i></p> </div> <p>[For RPC _____ Label Here]</p> <p style="margin-top: 10px;"><i>RPC</i></p>	<p><i>A. Underwood</i></p>

DATAMASTER DMT MAINTENANCE LOG

VSP - Williston

100132

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-19-21	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 20-14-100 EtOH Certified Conc. 0.100 [e Label Here] Range: 0.095 - 0.105 Prep Date: 9/22/20 Exp Date: 09/22/22 Opened: 2-19-21 Underwood</p> <p style="text-align: center;">RPC after DMT supervisor update</p>	A. Underwood Trooper
07/08/2021	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 20-15-100 EtOH Certified Conc. 0.100 [e Label Here] Range: 0.095 - 0.105 Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 07/08/2021 MISS</p> <p style="text-align: center;">RPC</p>	E. Shepard Sgt, Williston Police Dept.
10/27/2021	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 21-09-100 EtOH Certified Conc. 0.099 [e Label Here] Range: 0.094 - 0.104 Prep Date: 6/4/21 Trp Sweet Exp Date: 06/04/23 Opened: 10/27/21</p> <p style="text-align: center;">RPC</p>	B. Sweet Trooper
1/27/22	<p style="text-align: center;">Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	J. Duket (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/20/2022	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 21-09-100 EtOH Certified Conc. 0.099</p> <p>Range: 0.094 - 0.104 <i>Tpr Sweet</i> [For RPC] e Label Here]</p> <p>Prep Date: 6/4/21 Exp Date: 06/04/23</p> <p>Opened: <u>02/20/2022</u></p> <p style="text-align: center; font-size: 1.2em;">RPC</p>	Trooper Brandon Sweet VSP williston
06/19/22	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 22-01-100 EtOH Certified Conc. 0.099</p> <p>Range: 0.094 - 0.104 <i>Tpr Sweet</i> [For RPC] Label Here]</p> <p>Prep Date: 1/4/22 Exp Date: 01/04/24</p> <p>Opened: <u>06/19/2022</u></p> <p style="text-align: center; font-size: 1.2em;">RPC</p>	Trooper Sweet VSP williston
6/21/22	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 22-10-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105 [For RPC] e Label Here]</p> <p>Prep Date: 4/1/22 Exp Date: 04/01/24</p> <p>Opened: <u>6/21/2022 808</u></p> <p><i>Troubleshooting in response to low conc. on failed RPCs. No evidence of leak and 5-way valve worked well. Replaced new solution and promptly ran an ABP test - results expected conc. range. Completed RPC. DMT returned to service.</i></p>	S. Stratton (VFL)
10/17/2022	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 22-10-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105 <i>Trooper Sweet</i> [For RPC] e Label Here]</p> <p>Prep Date: 4/1/22 Exp Date: 04/01/24</p> <p>Opened: <u>10/17/2022</u></p>	Trooper Sweet

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/24/23	<p>Completed APM</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	J Dulcote (VFL)
02/23/2023	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC</p> <p style="font-size: small;">Lot: 22-01-100 EtOH Certified Conc. 0.099</p> <p style="font-size: small;">Range: 0.094 - 0.104</p> <p style="font-size: small;">Prep Date: 1/4/22 Exp Date: 01/04/24</p> <p style="font-size: small;">Opened: BWS 02/28/23</p>	Troy Sweet
3/13/2023	<p>Performed passing sample acceptance via check-in after checking voltages. Did note odor of volatiles from nearby garage.</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	S. Stratton (VFL)
06/14/23	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC</p> <p style="font-size: small;">Lot: 22-14-100 EtOH Certified Conc. 0.098</p> <p style="font-size: small;">Prep Date: 10/5/22 Exp Date: 10/05/24</p> <p style="font-size: small;">Opened: BWS 06/14/23</p>	Troy Sweet

Lee

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/29/23	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC Lot: 22-14-100 EtOH Certified Conc. 0.098 Label Here]</p> <p>Prep Date: 10/5/22 Exp Date: 10/05/24</p> <p>Opened: <i>BWS 10/29/2023</i></p>	<i>Trouper B. Sweet</i>
12/4/23	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC Lot: 23-08-100 EtOH Certified Conc. 0.099 Label Here]</p> <p>Prep Date: 7/19/23 Exp Date: 07/19/25</p> <p>Opened: <i>12/4/23 JSD</i></p> <p><i>Install performed post move to new barracks</i></p>	<i>J Dukette (VFL)</i>
1/23/24	<p style="text-align: center;"><i>Completed APM</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>J Dukette (VFL)</i>
02/29/24	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC Lot: 23-10-100 EtOH Certified Conc. 0.099 Label Here]</p> <p>Prep Date: 11/28/23 Exp Date: 11/28/25</p> <p>Opened: <i>02/29/2024 BWS</i></p>	<i>Trouper B. Sweet</i>

DATAMASTER DMT MAINTENANCE LOG

VSP Williston

Department/Agency

100132

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/18/24	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-08-100 EtOH Certified Conc. 0.099</p> <p>[For RPC or _____ Label Here]</p> <p>Prep Date: 7/19/23 Exp Date: 07/19/25</p> <p>Opened: BWS 06/18/24</p>	Trooper B. Sweet
10/29/24	<p style="text-align: center;">DMT Quality Control Solution</p> <p>Lot: 24-10-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or _____ Label Here]</p> <p>Prep Date: 7/29/24 Exp Date: 07/29/26</p> <p>Opened: BWS 10/29/24</p>	Trooper B. Sweet
1/21/25	APM	A. Bulduc
02/26/25	<p style="text-align: center;">DMT Quality Control Solution</p> <p>Lot: 24-10-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or _____ Label Here]</p> <p>Prep Date: 7/29/24 Exp Date: 07/29/26</p> <p>Opened: BWS 02/26/25</p>	Trooper Sweet
05/21/2025	<p style="text-align: center;">DMT Quality Control Solution</p> <p>Lot: 24-24-100 EtOH Certified Conc. 0.099</p> <p>[For RPC or _____ Label Here]</p> <p>Prep Date: 12/18/24 Exp Date: 12/18/26</p> <p>Opened: BWS 05/21/2025</p>	Trooper Sweet RPC
7/22/25	removed for refurb	 VFL