

DATAMASTER DMT MAINTENANCE LOG

Franklin County S.O. St Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2-3-12	DataMaster Simulator Solution Lot # 12-51-100 EtOH _{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/12/2012 Expires: 01/12/2013 [For RP _____ Place Label Here] OPENED: 2/3/12	Kirk Kimball VDH L Prog. Chief.
2-3-12	Installation per protocol. 2/3/12	
03/30/12	Printed Reports for Franklin [For RPC or Simulator Solution Change Place Label Here] County State's Attorney's Office	JR Lynch
04/27/12	DataMaster Simulator Solution Lot # 12-70-100 EtOH _{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 [For RP _____ Place Label Here] OPENED: 04/27/12 Simulator Solution Change	JR Lynch
5/8/12	Printed Reports for [For RPC or Simulator Solution Change Place Label Here] STATES Attorney	JH May
5/31/12	Software upgrade and lock change	A. Boldt

DATAMASTER DMT MAINTENANCE LOG

FCSO

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/8/12	Printed Reports for [For RPC or Simulator Solution Change Place Label Here] Franklin County States Attorney	J. Green
06/12/12	DataMaster Simulator Solution Lot # 12-51-100 EtOH _{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/12/2012 Expires: 01/12/2013 [For R lace Label Here] OPENED: 06/12/12 RPC	M. Tyel
7/3/12	Printed Reports for SA's office [For RPC or Simulator Solution Change Place Label Here]	J. H. H. H.
8/1/12	Printed Reports for SA's office [For RPC or Simulator Solution Change Place Label Here]	J. H. H. H.

DATAMASTER DMT MAINTENANCE LOG

FCSD

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
08/09/12	DataMaster Simulator Solution Lot # 12-51-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/12/2012 Expires: 01/12/2013 [For RPC or Simulator Solution Change Place Label Here] OPENED: 08/09/12 Simulator Solution Change	<i>AR Lych</i> Sergeant
10/3/12	Printed Records for SA's office [For RPC or Simulator Solution Change Place Label Here]	<i>Jay H. Hines</i>
10/5/12	DataMaster Simulator Solution Lot # 12-70-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: 10/05/12 JHS RPC	<i>J. Hines</i>
11/7/12	Printed Records for SA's office [For RPC or Simulator Solution Change Place Label Here]	<i>Jay H. Hines</i>
12/19/12	Printed Records for S.A.'s office [For RPC or Simulator Solution Change Place Label Here]	<i>AR Lych</i>
1/2/13	DataMaster Simulator Solution Lot # 12-87-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 1/2/13 JHS Simulator Solution Change	<i>J. Hines</i>

DATAMASTER DMT MAINTENANCE LOG

FCSO

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/7/13	Printed Records for SA's office [For RPC or Simulator Solution Change Place Label Here]	J. H. Sweeney
1/11/13	[For RPC or Simulator Solution Change Place Label Here] APM protocol completed.	R. D. Disoll DMT TECH
2/7/13	Printed Records for SA's office	Jay H. Sweeney
2/7/13	DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 2/7/13 JHS [For RPC e Label Here] RPC	Jay H. Sweeney
4/2/13	Printed Records for SA's Feb + March [For RPC or Simulator Solution Change Place Label Here]	Jay H. Sweeney

DATAMASTER DMT MAINTENANCE LOG

FCSO

100135

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
04/22/13	DataMaster Simulator Solution Lot # 12-84-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 9/24/2012 Expires: 9/24/2013 OPENED: 04/22/13 JRL [For F] Place Label Here] Simulator Solution Change	JRL Sergeant
6/6/13	RECORDS TO SA'S Records May [For RPC or Simulator Solution Change Place Label Here]	JRL
06/10/13	DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 06/10/13 JRL [For RF] Place Label Here] R.R.C. Simulator Solution Change	JRL
6/14/13	DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 6/14/13 [For RPC] Place Label Here] RPC	JRL

DATAMASTER DMT MAINTENANCE LOG

FCSD
Department/Agency

100135
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/17/13	DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 [For RP ice Label Here] OPENED: 6/17/13 RCD	R. Driscoll DMT TECH.
	Solution out of range during previous RPC attempt. Changed solution, leak checked simulator, and reran June RPC. In spec RCD 6/17/13	
7/11/13	Records To SA's [For RPC or Simulator Solution Change Place Label Here]	JH Sweeney
9/12/13	Records To SA's	JH Sweeney
10/22/13	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 OPENED: 10/22/13 Chet RPC	Chet
11/10/14	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.

DATAMASTER DMT MAINTENANCE LOG

FCSO

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>2/20/14</i>	DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: <i>2/20/14 JHS</i> <i>RPC</i> [ge Place Label Here]	<i>Jay H Sweeney</i>
<i>6/14/14</i>	Records To SA	<i>Jay H Sweeney</i>
<i>6/13/14</i>	DataMaster Simulator Solution Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: <i>6/13/14 Jay H Sweeney</i> <i>RPC</i> ce Label Here]	<i>Jay H Sweeney</i>
<i>7/1/14</i>	Records To SA	<i>Jay H Sweeney</i>
<i>8/11/14</i>	Records To SA [For RPC or Simulator Solution Change Place Label Here]	<i>Jay H Sweeney</i>
<i>9/2/14</i>	Records To SA	<i>Jay H Sweeney</i>
<i>10/05/14</i>	DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: <i>10/05/14 [Signature]</i> e Label Here]	<i>[Signature]</i>
<i>10/13/14</i>	Records To SA	<i>Jay H Sweeney</i>

DATAMASTER DMT MAINTENANCE LOG

FCSG

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/16/14	DataMaster Simulator Solution Lot # 14-12-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/10/2014 Expires: 9/10/2015 OPENED: 10/16/14 RED [For RPC or Simulator Solution Change Place Label Here] Install post TSI	R. Driscoll DMT TECH.
1/9/15	Completed APM Protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.
1/14/15	Records to LRA	J. Sweeney
1/27/15	Request for October 2014 Records - Records to SA [For RPC or Simulator Solution Change Place Label Here]	J. Sweeney
2/3/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 2/4/15 C. H. [For RPC or Simulator Solution Change Place Label Here]	C. H.

DATAMASTER DMT MAINTENANCE LOG

FCSO

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/11	Set correct time [For RPC or Simulator Solution Change Place Label Here]	N. Vallamouy
4/28	JAN, FEB, MAR Records To Lab	JH Pinner
6/10	<u>DataMaster Simulator Solution</u> Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 [For RPC Preparation Date: 12/9/2014 Expires: 12/9/2015 : Label Here] OPENED: 6/10/15 C. Miles	CH
9/30	2/3/15 RPC Records To SA Brooks	JH Pinner
10/7	<u>DataMaster Simulator Solution</u> Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 [For RPC Preparation Date: 07/28/2015 Expires: 07/28/2016 : Label Here] OPENED: 10/07/15 C. Miles	CH
10/15/15	Downloaded data to USB drive, uploaded new [For RPC or Simulator Solution Change Place Label Here] Software, and installed new I/O PCB. Connected to DPS server.	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

FLSO - ST. Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/15/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
2/11/16	<u>DataMaster Simulator Solution</u> Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: 2/11/16 JHS	RPC J. Sweeney
6/29/16	<u>DataMaster External Standard Solution</u> Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016 OPENED: 6/29/16 JHS	RPC [For RPC or Simulator Solution Change Place Label Here] J. Sweeney
10/9/16	<u>DMT External Standard Solution</u> Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 10/09/2016 NSV	RPC N. Villancourt
1/20/17	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
3/2/17	<u>DMT External Standard Solution</u> Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 3/2/17	RPC J. Sweeney
4/6/17	<u>DMT Simulator Solution</u> Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 4/6/17 ROS Install post simulator update	[For RPC or Simulator Solution Change Place Label Here] R. Driscoll VFL

DATAMASTER DMT MAINTENANCE LOG

FCSO St. Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/28/17	DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED: 06/28/2017 NSV [For RPC RPC Label Here]	N. Veillancourt
10/20/17	DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 OPENED: 10/20/17 Gm [For RPC RPC Label Here]	C. Miles
10/27/17	Installed version 2.08 PTC chip. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
2/16/18	Completed AEM protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL

DATAMASTER DMT MAINTENANCE LOG

FLSO - ST. Albans
Department/Agency

100135
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7-20-18	DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 [For RP or Simulator Solution Change Place Label Here] Miles OPENED: 7-20-18 <i>cm</i> Sgt	
5/4/18	Instrument removed from service due to [For RPC or Simulator Solution Change Place Label Here] Cal. Check errors.	R Dinsell VPL
5/9/18	DMT External Std Solution Lot: 18-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/21/18 Exp Date: 03/21/19 Opened: 5/9/18 <i>RCJ</i> [For RPC or Simulator Solution Change Place Label Here]	R Dinsell VPL
	Installation post TSE	
7/2/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 7/2/18 <i>cm</i> [For RP or Simulator Solution Change Place Label Here]	C. Miles Sgt
	RPC	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/11/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 [For RPC or Simulator Solution Change Place Label Here] Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 10/11/18 C.Miles RPC	C. Miles
2/12/19	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dukette (UFL)
03/15/19	DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 RPC Range: 0.096 - 0.106 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 03/15/19 NSV	N. Vaillancourt
7/7/19	DMT External Std Solution Lot: 19-04-100 EtOH Certified Conc. 0.101 RPC Range: 0.096 - 0.106 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 7/7/19 C.Miles	C. Miles

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/21/19	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 10/21/19	RPC Griswold D/C Reso
12/11/19	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 12/11/19 Solution Change	RPC Griswold SDC Reso
2/3/20	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 2/3/20 RPC passing sample acceptance performed by JSD on 2/11/2020 JSD	RPC Griswold SP/Reso
2/11/2020	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/2/20	DMT External Std Solution Lot: 20-03-100 EtOH Certified Conc. 0.100 <i>RPC</i> [For RPC Range: 0.095 - 0.105 <i>7/2/20</i> <i>SD/C FSD</i> <i>Place Label Here</i>] Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: <i>7/2/20</i>	Griswold SD/C FSD
9/22/20	DMT External Std Solution Lot: 20-03-100 EtOH Certified Conc. 0.100 <i>oct</i> <i>RPC</i> [Range: 0.095 - 0.105 <i>Place Label Here</i>] Prep Date: 1/29/20 Exp Date: 01/29/21 Opened: <i>9/22/20</i>	Griswold SD/C FSD
9/29/2020	DMT removed from service	J Dukette (VFL)
10/9/2020	DMT External Std Solution [For R Lot: 20-14-100 EtOH Certified Conc. 0.100 <i>Place Label Here</i>] Range: 0.095 - 0.105 Prep Date: 9/22/20 Exp Date: 09/22/22 Opened: <i>SD 10/9/2020</i> <i>Replaced o-ring for sensor & cal part due to slow leak. Double o-ring used on jar.</i> <i>install post TSI</i>	J Dukette (VFL)
11/26/2021	DMT removed from service and taken to VFL for repairs. [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)

DATAMASTER DMT MAINTENANCE LOG

FCSD St. Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/12/2021	DMT External Std Solution Lot: 20-15-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 e Label Here] Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: <u>2/12/2021 828</u> <i>Double o-ring used on jar. Performed install and APM protocols.</i>	S. Stratton (VFL)
2/22/2021	<i>Removed DMT for re-cal at the VFL after cal check errors</i> [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
3/4/2021	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 Label Here] Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: <u>828 3/4/2021</u> <i>Completed install - DMT back in service</i>	S. Stratton (VFL)
3/18/21	<i>DMT removed for service</i> [For RPC or Simulator Solution Change Place Label Here]	J Dulcet (VFL)

DATAMASTER DMT MAINTENANCE LOG

FCSO St. Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/14/2021	DMT External Std Solution [For RPC Lot: 21-05-100 EtOH Certified Conc. 0.100 e Label Here] Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: 3/3/21 5/14/2021 Successfully completed installation protocol.	S. Stratton (VFL)
06/22/2021	DMT External Std Solution [For RPC Lot: 20-14-100 EtOH Certified Conc. 0.100 e Label Here] Range: 0.095 - 0.105 Prep Date: 9/22/20 Exp Date: 09/22/22 Opened: 06/22/2021 NSV RPC	N. Vaillancourt
09/04/21	DMT External Std Solution [For RPC Lot: 21-11-100 EtOH Certified Conc. 0.100 e Label Here] Range: 0.095 - 0.105 Date: 7/16/21 Exp Date: 07/16/23 d: 09/04/21 NSV Solution chg	N. Vaillancourt
10/16/21	DMT External Std Solution [For RPC Lot: 21-11-100 EtOH Certified Conc. 0.100 e Label Here] Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: 10/16/21 NSV RPC	N. Vaillancourt

DATAMASTER DMT MAINTENANCE LOG

FCSD St Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/6/22	Manually reassigned the ethernet [For RPC or Simulator Solution Change Place Label Here] connections but was unable to restore connection. Bypassed comm board and worked w/ FCSD to determine its an IT issue. DMT operating properly but no connection	J Durette (VFL)
1/6/22	DMT External Std Solution [For R Lot: 21-12-100 EtOH Certified Conc. 0.099 'lace Label Here] Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 1/6/22	Solution chg C. Miles 099
1/25/22	Completed APM Protocol and [For RPC or Simulator Solution Change Place Label Here] restored connection to internet	J Durette (VFL)
02/19/22	DMT External Std Solution [For RPC Lot: 22-01-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 1/4/22 Exp Date: 01/04/24 Opened: 02/19/2022 NSV	RPC N. Vaillancourt

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/1/22	DMT External Std Solution Lot: 22-01-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 1/4/22 Exp Date: 01/04/24 Opened: Solution Change	[For RP] ice Label Here] ET Miles
05/27/22	DMT External Std Solution Lot: 22-01-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 1/4/22 Exp Date: 01/04/24 Opened: 05/27/2022 NSV June RPC	[For RPC] e Label Here] N. Vaillancourt
7/27/22	DMT External Std Solution Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 7/27/22 com Solution Change	[For RP] ace Label Here] ET Miles
9/30	DMT External Std Solution Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 9/30/22 RPC	[For RP] ace Label Here] Miles

DATAMASTER DMT MAINTENANCE LOG

FLSO St. Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/22/22	Accuracy & Precision TEST [For RPC or Simulator Solution Change Place Label Here]	N. Vaillancourt
10/22/22	DMT External Std Solution Lot: 22-14-100 EtOH Certified Conc. 0.098 <i>Solution Change</i> Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: 10/22/22 <i>NSV</i>	
11/9/23	Completed ALM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Durette (VFL)
01/21/23	DMT External Std Solution Lot: 22-14-100 EtOH Certified Conc. 0.098 <i>RPC</i> [For RP(ce Label Here] Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: 01/21/2023 <i>NSV</i>	N. Vaillancourt
04/23/23	DMT External Std Solution Lot: 22-14-100 EtOH Certified Conc. 0.098 [For Place Label Here] Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: 04/23/23 <i>NSV</i> <i>Solution Change</i> <i>Changed Printer Ink</i>	N. Vaillancourt

DATAMASTER DMT MAINTENANCE LOG

FCSO St. Albans

100135

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>7/7/23</i>	DMT External Std Solution [For R Lot: 22-14-100 EtOH Certified Conc. 0.098 lace Label Here] Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: <i>7/7/23</i> <i>CL</i> <i>RPC</i>	<i>C. Miles</i>
<i>08/22/23</i>	DMT External Std Solution [For RPC Lot: 22-14-100 EtOH Certified Conc. 0.098 Solution Change : Label Here] Prep Date: 10/5/22 Exp Date: 10/05/24 Opened: <i>08/22/23</i> <i>NSV</i>	<i>N. Vaillancourt</i>
<i>10/26/23</i>	DMT External Std Solution [For RPI Lot: 23-08-100 EtOH Certified Conc. 0.099 ce Label Here] Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: <i>10/26/2023</i> <i>NSV</i> <i>RPC</i>	<i>N. Vaillancourt</i>
<i>12/22/23</i>	DMT External Std Solution [For RP Lot: 23-08-100 EtOH Certified Conc. 0.099 ace Label Here] Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: <i>12/22/2023</i> <i>NSV</i>	<i>N. Vaillancourt</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/18/24	<i>Completed APM protocol</i> [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
02/14/24	DMT External Std Solution Lot: 23-10-100 EtOH Certified Conc. 0.099 <i>RPC</i> Prep Date: 11/28/23 Exp Date: 11/28/25 [For RPC or Simulator Solution Change Place Label Here] Opened: <i>02/14/24 NSV</i>	N. Vaillancourt
04/16/24	DMT External Std Solution Lot: 23-10-100 EtOH Certified Conc. 0.099 <i>solution change</i> Prep Date: 11/28/23 Exp Date: 11/28/25 [For RPC or Simulator Solution Change Place Label Here] Opened: <i>04/16/24 NSV</i>	N. Vaillancourt
	DMT Quality Control Solution [For R Lot: 24-07-100 EtOH Certified Conc. 0.100 [Place Label Here] Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: <i>6/7/24 C4m</i> <i>RPC</i>	C. Miles

DATAMASTER DMT MAINTENANCE LOG

FCSO St Albans

Department/Agency

100135

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
08/06/24	DMT External Std Solution Lot: 23-08-100 EtOH Certified Conc. 0.099 [For RPC ce Label Here] Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: 08/06/24 NV Solution Change	N. Vaillancourt
10/13/24	DMT Quality Control Solution Lot: 24-07-100 EtOH Certified Conc. 0.100 [For RP ice Label Here] Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: 10/13/24 NSV RPC	N. Vaillancourt
12/18/24	DMT Quality Control Solution Lot: 24-10-100 EtOH Certified Conc. 0.101 [For RP ce Label Here] Prep Date: 7/29/24 Exp Date: 07/29/26 Opened: 12/18/24 NSV Solution Change	N. Vaillancourt
1/15/25	APM [For RPC or Simulator Solution Change Place Label Here]	J. Boldt

FCSO St. Albans

100135

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/16/25	DMT Quality Control Solution Lot: 24-10-100 EtOH Certified Conc. 0.101 [For RPC ce Label Here] Prep Date: 7/29/24 Exp Date: 07/29/26 Opened: 02/16/25 NSV RPC	N. Vaillancourt
04/21/25	DMT Quality Control Solution Lot: 24-24-100 EtOH Certified Conc. 0.099 [For RP(ce Label Here] Prep Date: 12/18/24 Exp Date: 12/18/26 Opened: 04/21/25 NSV Solution Change	N. Vaillancourt
7/1/25	Removed for retests [For RPC or Simulator Solution Change Place Label Here]	J. Bolden
	[For RPC or Simulator Solution Change Place Label Here]	