

DMT MAINTENANCE LOG

Manistown PD

Department/Agency

100136

DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/2/25	DMT Quality Control Solution [For RPC Lot: 25-11-100 EtOH Certified Conc. 0.102 Label Here] Prep Date: 4/7/25 Exp Date: 04/07/27 Opened: Install	A. Bolduc VFL
10-15-25	DMT Quality Control Solution [For RPC Lot: 25-15-100 EtOH Certified Conc. 0.101 Label Here] LDL 10-15-25 Prep Date: 6/17/25 Exp Date: 06/17/27 Opened: 10-15-25 RPC for October	MPD Senior Patrol Lance D. Lamb Lance D. Lamb
01-18-26	DMT Quality Control Solution [Lot: 25-15-100 EtOH Certified Conc. 0.101 Label Here] LDL 01-18-2026 Prep Date: 6/17/25 Exp Date: 06/17/27 Opened: 01-18-2026 SSC (was out of Range)	MPD Senior Patrol Lance D. Lamb Lance D. Lamb
02-26-26	DMT Quality Control Solution [Lot: 26-01-100 EtOH Certified Conc. 0.101 Label Here] LDL Prep Date: 1/5/26 Exp Date: 01/05/28 Opened: 02-26-2026 RPC for Feb.	MPD Senior Patrol Lance D. Lamb Lance D. Lamb
5/22/26	DMT Quality Control Solution [Lot: 26-01-100 EtOH Certified Conc. 0.101 Label Here] Prep Date: 1/5/26 Exp Date: 01/05/28 Opened: 5/22/26 RPC for June	MPD Jackson Schroder Patrol

6/22/26

removed for annual calibration

B. Shumway
VFL