

DATAMASTER DMT MAINTENANCE LOG

Montpelier PD
Department/Agency

100186
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/10/12	<p>RPC FOR OCT '12</p> <p>[F DataMaster Simulator Solution] </p> <p>Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013</p> <p>OPENED: 10/10/12</p>	<p> SII</p>
11/15/12	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">APM protocol performed</p>	<p>R. Driscoll DMT TECH</p>
2/7/13	<p>DataMaster Simulator Solution</p> <p>Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013</p> <p>OPENED: @ 2/7/13</p> <p>RPC for FEB '13</p>	<p></p>
5/28/13	<p>Response timeout / software reboot - Troubleshooting</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>intermittent fails, Diagnostic check good, power supply voltage good, run check in + AIR TEST with no fails. could have been faulty breath tube or plugged. software reboot sounds like record printer "glitch"</p>	<p>R. Driscoll</p>

DATAMASTER DMT MAINTENANCE LOG

Montpelier P.D.
Department/Agency

10036
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/28/13	<p>DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 5/28/13 C.B.W. (SIC) SOLUTION CHANGE</p>	<p><i>[Signature]</i> SIC</p>
7/15/13	<p>DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 OPENED: 7/15/13 RCD JUNE 2013 RPC</p>	<p>R. Driscoll</p>
10/21/13	<p>DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 OPENED: 10/21/13 C.B.</p>	<p>C. B.W. Patrol Officer SIO</p>
11/13/13	<p>Completed APM protocol. [For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll</p>

DATAMASTER DMT MAINTENANCE LOG

Montpelier Police Dept.
Department/Agency

100136
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/6/14	<p>DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires 05/10/2014</p> <p>OPENED: (3) 3/6/14</p> <p>RPC FEB 2014</p>	<p>[Change Place Label Here]</p> <p><i>[Signature]</i> 810</p>
6/27/14	<p>DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015</p> <p>OPENED: (3) 6/27/14</p>	<p>RPC JUNE '14</p> <p>[Change Place Label Here]</p> <p><i>[Signature]</i> 810</p>
10/29/14	<p>DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015</p> <p>OPENED: (3) RPC OCT 2014</p>	<p>RPC OCT '14</p> <p>[Change Place Label Here]</p> <p><i>[Signature]</i> 810</p>
11/5/14	<p>RESET CLOCK (DAYLIGHT SAVINGS END)</p> <p>(3)</p>	
11/12/14	<p>Completed ARM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>[Signature]</i> R. Driscoll</p>



DATAMASTER DMT MAINTENANCE LOG

Montpelier Police Dept

Department/Agency

100136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/28/15	<p>DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: <u>CB</u></p> <p style="text-align: right;">RPC Feb '15 [Place Label Here]</p>	 810
4/27/15	<p>DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: <u>4/27/15 RCD</u></p> <p style="text-align: right;">[For RPC] [Place Label Here]</p> <p>Reset DMT to regain communication w/ printer. Started solution change for DMT supervisor (replaced) Solution change protocol to be performed by DMT supervisor <u>CB</u></p>	R. Driscoll DMT Tech.
7/15/15	<p>DataMaster Simulator Solution Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: <u>7/15/15 CB</u></p> <p style="text-align: right;">RPC JUNE '15 [Place Label Here]</p>	 810
9/28/15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Downloaded data to USB drive, uploaded new software, and installed new I/O PCB. No Ethernet port available at this time to connect to DM Host server.</p>	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

Montpelier PD

Department/Agency

100136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<p>10/31/16</p>	<p>DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 OPENED: <u>(initials)</u></p> <p>Change Place Label Here]</p> <p>RPC OCT 2016</p> <p>10/31/16</p>	<p><u>J. Ben</u> Patrol Officer SIO</p>
<p>12/1/16</p>	<p>Completed RPM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT Tech.</p>
<p>2/23/17</p>	<p>DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: <u>(initials)</u> 2/23/17</p> <p>Change Place Label Here]</p> <p>Feb. 2017 RPC</p>	<p><u>C. J. Quisnel</u> Officer 2/23/17 <u>(initials)</u></p>
<p>4/3/17</p>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DMT returned to VFL for maintenance</p>	<p>Jeff Duketta (VFL)</p>

DATAMASTER DMT MAINTENANCE LOG

Montpelier P.O.
Department/Agency

100136
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/4/17	<p>DMT Simulator Solution [For RPC or Simulator Solution Change Place Label Here] Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 4/4/17 JSO</p> <p>Install post sim conversion JSO 4/4/17</p>	Jeff Duvigne VFL
7/03/17	<p>DMT Simulator Solution [For RPC or Simulator Solution Change Place Label Here] Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 7/3/17 JSO</p> <p>RPC JUNE '17 JSO 7/3/17</p>	[Signature]
10/20/17	<p>Installed version 2.08 PIC chip and Completed [For RPC or Simulator Solution Change Place Label Here] APM protocol.</p>	R. Driscoll VFL
10/30/17	<p>DMT Simulator Solution [For RPC or Simulator Solution Change Place Label Here] Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 10/30/17 JSO</p> <p>RPC Oct '2017</p>	[Signature] JSO


DATAMASTER DMT MAINTENANCE LOG

Montpelier P.O.

Department/Agency

100136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/28/18	<p>DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 OPENED: <u>2/28/18</u> <u>809</u></p> <p>[For R Place Label Here]</p> <p>RPC FEB '18</p>	 809
6/28/18	<p>DMT External Std Solution Lot: 18-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/21/18 Exp Date: 03/21/19 Opened: <u>6/28/18</u> <u>809</u></p> <p>[For R Place Label Here]</p> <p>RPC JUNE '18</p>	 809
7/31/18	<p style="text-align: center;">Completed APR protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	R. Driscoll VFL
10/28/18	<p>DMT External Std Solution Lot: 18-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/21/18 Exp Date: 03/21/19 Opened: <u>10/28/18</u> <u>809</u></p> <p>[For RPC Place Label Here]</p> <p>RPC OCT '18</p>	 809

DATAMASTER DMT MAINTENANCE LOG

Montpelier PD

100136

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/11/19	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-04-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/19 Exp Date: 02/28/20</p> <p>Opened: 3/11/19 M3</p> <p style="font-size: 1.5em; color: blue;">Feb. 2019 RPC</p>	C. Quinlan Officer M.P.D.
6/28/19	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-04-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/19 Exp Date: 02/28/20</p> <p>Opened: 6/28/19</p> <p style="font-size: 1.5em; color: blue;">RPC JUNE '19</p>	809
7/30/19	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-10-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/17/19 Exp Date: 06/17/20</p> <p>Opened: 10/22/19</p> <p style="font-size: 1.5em; color: blue;">Completed ARM procedure</p>	J Durette (VPL)
10/23/19	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-10-100 EtOH Certified Conc. 0.101</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/17/19 Exp Date: 06/17/20</p> <p>Opened: 10/22/19</p> <p style="font-size: 1.5em; color: blue;">RPC Oct '19</p>	809



DATAMASTER DMT MAINTENANCE LOG

Manspeizer P.O.

Department/Agency

180136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>2/23/20</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RP] Lot: 19-10-100 EtOH Certified Conc. 0.101 [ce Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/17/19 Exp Date: 06/17/20</p> <p>Opened: <i>[Signature]</i> <i>2/23/20</i></p> <p><i>RPC FEB '20</i></p>	 <i>808</i>
<i>7/06/20</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPC] Lot: 20-10-100 EtOH Certified Conc. 0.101 [Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/25/20 Exp Date: 06/25/22</p> <p>Opened: <i>[Signature]</i> <i>7/6/20</i></p> <p><i>RPC BONS '20</i></p>	 <i>808</i>
<i>7/13/2020</i>	<p style="text-align: center;"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>following software update</i></p> <p style="text-align: center;"><i>DMT removed from service due to RF issues ASD 7/13/2020</i></p>	<p><i>J Dulcett</i> <i>(VFL)</i></p>
<i>10/30/20</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RF] Lot: 20-10-100 EtOH Certified Conc. 0.101 [lace Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 6/25/20 Exp Date: 06/25/22</p> <p>Opened: <i>[Signature]</i> <i>10/30/20</i></p> <p><i>RPC OCT '20</i></p>	 <i>808</i>




DATAMASTER DMT MAINTENANCE LOG

Montpelier PD

Department/Agency

100136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/21	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R Lot: 21-04-100 EtOH Certified Conc. 0.100 [Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 2/12/21 Exp Date: 02/12/23</p> <p>Opened: 2/24/21 (CB)</p> <p>RPC FEB '21</p>	 808
6/28/21	<p style="text-align: center;">DMT External Std Solution</p> <p>[Lot: 21-04-100 EtOH Certified Conc. 0.100 [nge Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 2/12/21 Exp Date: 02/12/23</p> <p>Opened: 6/28/21</p> <p>RPC June '21</p>	 808
7/19/2021	<p style="text-align: center;"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>S. Stratton (VFL)</p>
10/17/21	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RP(Lot: 21-09-100 EtOH Certified Conc. 0.099 [ce Label Here]</p> <p>Range: 0.094 - 0.104</p> <p>Prep Date: 6/4/21 Exp Date: 06/04/23</p> <p>Opened: 10/17/21 @ 0220 (CB) 819</p> <p>RPC OCT '21</p>	 819

DATAMASTER DMT MAINTENANCE LOG

MONTPELIER PD

Department/Agency

10013C

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/19/22	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RP Lot: 21-12-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 1/19/22 @ 819]</p> <p style="text-align: center; font-size: 1.2em;">RPC FEB 2022 (w/ 15 Day Window)</p>	@ 819
5/2/22	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R Lot: 21-12-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 10/8/21 Exp Date: 10/08/23 Opened: 5/2/22 @ 816]</p> <p style="text-align: center; font-size: 1.2em;">June RPC 2022 (w/ 15 Day Window)</p>	@ 816
7/8/22	<p style="text-align: center; font-size: 1.5em; color: blue;">Completed APM</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	J Duke (UFL)
1/4/22	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RP Lot: 22-01-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 1/4/22 Exp Date: 01/04/24 Opened: 1/4/22 @ 816]</p> <p style="text-align: center; font-size: 1.2em;">Oct. RPC 22</p>	@ 816

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

1/10/23

100136

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>1/20/23</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RF Lot: 22-06-100 EtOH Certified Conc. 0.100 Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: <i>1/20/23</i> <i>#813</i></p> <p style="text-align: center;"><i>FEB 23</i> <i>RPC</i></p>	<i>#813</i>
<i>5/31/23</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For F Lot: 22-06-100 EtOH Certified Conc. 0.100 Place Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: <i>5/31/23</i> <i>#813</i></p> <p style="text-align: center;"><i>JUNE 23</i> <i>RPC</i></p>	<i>#813</i>
<i>7/6/23</i>	<p style="text-align: center;"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>S Stratton</i> <i>(VFL)</i>
<i>9/24/23</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RF Lot: 22-15-100 EtOH Certified Conc. 0.100 Place Label Here]</p> <p>Prep Date: 10/26/22 Exp Date: 10/26/24</p> <p>Opened: <i>#812</i> <i>9/24/23</i></p> <p style="text-align: center;"><i>OCT 23</i> <i>RPC</i></p>	<i>#812</i>

DATAMASTER DMT MAINTENANCE LOG

Montpelier
 Department/Agency

100136
 DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/20/24	<p align="center">DMT External Std Solution</p> <p>Lot: 22-15-100 EtOH Certified Conc. 0.100 [For RPC] Place Label Here]</p> <p>Prep Date: 10/26/22 Exp Date: 10/26/24</p> <p>Opened: 1/20/24</p> <p align="center">FEB 24 RPC</p>	
7/1/24	<p align="center">DMT Quality Control Solution</p> <p>Lot: 24-07-100 EtOH Certified Conc. 0.100 [For RPC] Place Label Here]</p> <p>Prep Date: 5/29/24 Exp Date: 05/29/26</p> <p>Opened: 7/1/24</p> <p align="center">JUNE '24 RPC</p>	
7/2/24	<p align="center">Completed ARM</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p align="center">J. Duke (VFC)</p>
11-7-24	<p align="center">DMT Quality Control Solution</p> <p>Lot: 24-07-100 EtOH Certified Conc. 0.100 [For RPC] Place Label Here]</p> <p>Prep Date: 5/29/24 Exp Date: 05/29/26</p> <p>Opened: 11-7-24</p> <p align="center">Oct. 24 RPC</p>	

DATAMASTER DMT MAINTENANCE LOG

Montpelier PD

Department/Agency

100136

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<p>2/12/25</p>	<p style="text-align: center;">DMT Quality Control Solution</p> <p>[For RPC Lot: 24-07-100 EtOH Certified Conc. 0.100]</p> <p>Prep Date: 5/29/24 Exp Date: 05/29/26</p> <p>Opened: 2/12/25</p> <p style="font-size: 1.2em; text-align: center;">Feb. '25 RPC</p>	<p><i>[Signature]</i></p>
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	