

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP

Department/Agency

100138

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/25/11	<p><b>DataMaster Simulator Solution</b>                      Lot # 12-13-100 EtOH<sub>aq</sub> Conc: 0.102                      Range: 0.097 - 0.107                      Preparation Date: 09/09/2011 Expires: 09/09/2012</p> <p>[ For RP <span style="float: right;">ice Label Here]</span></p> <p>OPENED: 10/25/11 <span style="float: right;"><del>10/25/11</del></span></p> <p style="font-size: 1.2em; text-align: center;">Install Per Protocol</p>	<p style="font-size: 1.5em;">Steven</p> <p style="font-size: 1.5em;">Hanno's</p> <p style="font-size: 0.8em;">Kirk Kimball Program Chief Organic Chemistry</p>
11/17/11	<p>Installed software update <del>for</del></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p style="font-size: 1.5em;"><del>W. Kimball</del></p> <p style="font-size: 0.8em;">Prog. Chief.</p>
12-06-11	<p><b>DataMaster Simulator Solution</b>                      Lot # 12-13-100 EtOH<sub>aq</sub> Conc: 0.102                      Range: 0.097 - 0.107                      Preparation Date: 09/09/2011 Expires: 09/09/2012</p> <p>Change Place Label Here]</p> <p>OPENED:</p>	<p style="font-size: 1.5em;">Kerry Skochin</p> <p style="font-size: 1.5em;">K7SL</p>
12-06-11	<p style="font-size: 1.2em;">CHANGE TIME</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p style="font-size: 1.5em;">Kerry Skochin</p> <p style="font-size: 1.5em;">K7SL</p>

# DATAMASTER DMT MAINTENANCE LOG

*Perby VSP*

Department/Agency

*100138*

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>01/20/12</i>	<p><i>Check In</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<i>KSKoch</i>
<i>01/20/12</i>	<p>[ For RPC</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 12-45-100 EtOH<sub>aq</sub> Conc: 0.100                      Range: 0.095 - 0.105                      Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>OPENED: <i>01/20/12 KJ</i></p>	<i>KSKoch</i>
<i>1/23/12</i>	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Simulator NOT Leaking, Simulator Temp Adjusted + 0.2°C Diag + A+P Test Run and Passed Witnessed K.S.</i></p>	<i>Stevy Harnois</i>
<i>05/08/12</i>	<p><i>CHANGED TIME</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<i>KSKoch</i>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
05/25/12	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 12-45-100 EtOH<sub>aq</sub> Conc: 0.100                      Range: 0.095 - 0.105                      Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>OPENED: 05/25/12 <i>[Signature]</i></p>	<p><i>K Sll</i> <i>KSKOCHIN</i></p>
<p><i>RD</i> 8/27/100 8/27/12</p>	<p><i>Performed RPM + software update</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i></p>
10/19/12	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 12-51-100 EtOH<sub>aq</sub> Conc: 0.100                      Range: 0.095 - 0.105                      Preparation Date: 01/12/2012 Expires: 01/12/2013</p> <p>OPENED: 10/19/12 <i>[Signature]</i></p>	<p><i>K Sll</i> <i>KSKOCHIN</i></p>
12/10/12	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Set TIME</i></p>	<p><i>K Sll</i> <i>KSKOCHIN</i></p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/04/13	<p><b>DataMaster Simulator Solution</b>                      [ For RPC    Lot # 12-87-100    EtOHaq Conc: 0.101                      Range: 0.096 - 0.106                      Preparation Date: 10/31/2012 Expires: 10/31/2013                      OPENED: 02/04/13 <i>AS</i>    e Label Here]</p>	<p><i>KSKoch</i> <i>KSLL</i></p>
3/17/13	<p>Sim out of range on second test of 13B200792                      [ For RPC or Simulator Solution Change Place Label Here]                      Diagnostic Test Run, ATP Test Run                      Findings:                      - Simulator not completely back into tower                      - 094 Sim concentration on 2nd test                      Time updated for daylight sim</p>	<p><i>B. Shelp</i>                       Back in Range                      after                      solution change</p>
5/21/13	<p><b>DataMaster Simulator Solution</b>                      Lot # 12-73-100    EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 04/27/2012 Expires: 04/27/2013                      OPENED: <i>Shelp 3/17/13 Sim Sol. change</i></p>	<p><i>B. Shelp</i> <i>BS</i></p>
5/22/13	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center; font-size: 2em;">↓</p>	<p><i>KSLL</i> <i>KSKoch</i></p>

**DataMaster Simulator Solution**

Lot # 12-73-100    EtOHaq Conc: 0.098  
 Range: 0.093 - 0.103  
 Preparation Date: 04/27/2012 Expires: 04/27/2013  
 OPENED: *Shelp 3/17/13 Sim Sol. change*

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP  
Department/Agency

100138  
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/16/13	<p style="text-align: center;"><u>Completed annual PM</u></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">R. Driscoll DMT TECH.</p>
10/2/13	<p>[ For RPC <u>DataMaster Simulator Solution</u> Place Label Here]</p> <p>Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014</p> <p>OPENED: <u>10/2/2013 GM RPC</u></p>	<p style="text-align: center;">G. Marsh 10/2/13</p>
11/2/13 DMT rec'd RPC Req. via	<p>[ For RPC <u>DataMaster Simulator Solution</u> Place Label Here]</p> <p>Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014</p> <p>OPENED: <u>11/2/13 BSH RPC</u></p>	<p style="text-align: center;">BSH 11/2/13</p>
2/26/14 RPC	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><u>DataMaster Simulator Solution</u></p> <p>Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014</p> <p>OPENED:</p>	<p style="text-align: center;">G. Marsh 2/26/14</p>

# DATAMASTER DMT MAINTENANCE LOG

Deerby USD

Department/Agency

100138

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/30/14	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 14-05-100 EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 6/19/2014 Expires: 6/19/2015                      OPENED: <u>6/30/14 GM APC</u></p>	<p>GREG MARSH                      6/30/14                      DMT Supervisor</p>
7/16/14	<p style="color: blue;">Completed ARM Protocol</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p style="color: blue;">R. Driscoll                      DMT TECH.</p>
10/28/14	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 14-05-100 EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 6/19/2014 Expires: 6/19/2015                      OPENED: <u>10/28/14 GM APC</u></p>	<p>G. Marsh                      DMT                      Supervisor</p>
2/18/15	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L                      Range: 0.094 - 0.104                      Preparation Date: 12/9/2014 Expires: 12/9/2015                      OPENED:</p>	<p>G. Marsh                      DMT                      Supervisor</p>

# DATAMASTER DMT MAINTENANCE LOG

VSP - Derby Barracks  
 Department/Agency

100138  
 DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/28/15	<p style="text-align: center;"><i>Solution out of Range</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L                      Range: 0.094 - 0.104                      Preparation Date: 12/9/2014 Expires: 12/9/2015</p> <p>OPENED: <i>5/28/15</i> <i>BDS</i> <i>RPC</i></p>	<p><i>Ben Shep</i>  <i>Ser. of Trooper</i>  <i>DMT SIP</i></p>
8/18/15	<p style="text-align: center;"><i>Per Rob Driscoll</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DataMaster Simulator Solution</b>                      Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L                      Range: 0.095 - 0.105                      Preparation Date: 4/28/2015 Expires: 4/28/2016</p> <p>OPENED: <i>8/18/15</i> <i>BDS</i> <i>Solution</i> <i>Change</i></p>	<p><i>B. Shep</i>  <i>SIP</i>  <i>DMT SIP</i></p>
8/19/15	<p><i>Downloaded data to USB drive, uploaded new software, installed new comm. PCB, completed APM, printed breath record history, and verified correct solution lot # and concentration.</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i>  <i>DMT Tech.</i></p>
10/8/15	<p><b>DataMaster Simulator Solution</b>                      Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L                      Range: 0.095 - 0.105                      Preparation Date: 4/28/2015 Expires: 4/28/2016</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>10/8/15</i> <i>GM</i> <i>RPC</i></p>	<p><i>G. Marsh</i>  <i>OCTOBER 14</i></p>

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP  
Department/Agency

100138  
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/4/16	<p style="text-align: center;"><i>Solution out of Range</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><b>DataMaster Simulator Solution</b>                      Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L                      Range: 0.094 - 0.104 g/210L                      Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>OPENED: <u>1/4/16 GM</u></p> <p style="text-align: center;"><i>Solution Change</i></p>	<p><i>GM ARSH</i></p> <p>DMT Supervisor</p>
2/18/16	<p style="text-align: center;"><i>RPC - February</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><b>DataMaster Simulator Solution</b>                      Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L                      Range: 0.094 - 0.104 g/210L                      Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>OPENED: <u>2/18/16 GM</u></p>	<p><i>GM ARSH</i></p> <p>DMT Supervisor</p>
6/20/16	<p style="text-align: center;"><i>RPC - June</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><b>DataMaster External Standard Solution</b>                      Lot # 16-04-100 EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 3/25/2016 Expires: 3/25/2017</p> <p>OPENED: <u>6/20/16 GM</u></p>	<p><i>GM ARSH</i></p> <p>DMT Supervisor</p>
8/30/16	<p style="text-align: center;"><i>Completed APM Protocol</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i></p> <p>DMT Tech</p>

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP  
Department/Agency

100138  
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/18/16	<p><b>DataMaster External Standard Solution</b>                      Lot # 16-04-100 EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 3/25/2016 Expires: 3/25/2017</p> <p>[ For RPC or S... e Label Here]</p> <p>OPENED: _____</p> <p style="text-align: center;">RPC - October</p>	<p>G. Marsh DMT Supervisor</p>
12/21/16	<p><b>DMT External Standard Solution</b>                      Lot # 16-21-100 EtOHaq Conc: 0.098                      Range: 0.093 - 0.103                      Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>[ For RPC or S... bel Here]</p> <p>OPENED: 12/21/16</p> <p style="text-align: center;">Sim. Solution out of range.</p>	<p>G Marsh DMT Supervisor</p>
3/3/17	<p><b>DMT Simulator Solution</b>                      Lot # 17-10-100 EtOHaq Certified Conc: 0.101                      Range: 0.096 - 0.106                      Preparation Date: 02/06/2017 Expires: 02/06/2018</p> <p>[ For RPC or ... abel Here]</p> <p>OPENED: 3/3/17</p> <p style="text-align: center;">RPC - March</p>	<p>G Marsh DMT Supervisor</p>
Nov 21 2017	<p><b>DMT Simulator Solution</b>                      Lot # 17-10-100 EtOHaq Certified Conc: 0.101                      Range: 0.096 - 0.106                      Preparation Date: 02/06/2017 Expires: 02/06/2018</p> <p>[ For RPC or ... abel Here]</p> <p>OPENED: 4/21/17</p> <p style="text-align: center;">Sim. Solution Change DATAMASTER out of range</p>	<p>G Marsh DMT Supervisor</p>

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP  
Department/Agency

100138  
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/28/17	<p><b>DMT Simulator Solution</b>                      Lot # 17-16-100 EtOHaq Certified Conc: 0.101                      Range: 0.096 - 0.106                      Preparation Date: 04/03/2017 Expires: 04/03/2018                      OPENED: <u>4/28/17 RCD</u></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><u>Install post simulator update</u></p>	<p><u>R. Driscoll</u> VFL</p>
4/29/17	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DMT Simulator Solution</b>                      Lot # 17-18-100 EtOHaq Certified Conc: 0.100                      Range: 0.095 - 0.105                      Preparation Date: 05/02/2017 Expires: 05/02/2018                      OPENED: <u>6/19/17 GM</u></p> <p style="text-align: center;"><u>June RPC</u></p>	<p><u>G. Marsh</u> DMT Supervisor</p>
	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p><b>DMT Simulator Solution</b>                      Lot # 17-20-100 EtOHaq Certified Conc: 0.102                      Range: 0.097 - 0.107                      Preparation Date: 07/07/2017 Expires: 07/07/2018                      OPENED: <u>10/24/17 GM</u></p>	<p><u>G. Marsh</u> DMT Supervisor</p>
11/2/17	<p><u>Installed version 2.08 PIC chip.</u></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><u>R. Driscoll</u> VFL</p>

# DATAMASTER DMT MAINTENANCE LOG

Derby VSP

Department/Agency

100138

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/13/18	<p><b>DMT Simulator Solution</b></p> <p>Lot # 17-20-100 EtOHaq Certified Conc: 0.102                      Range: 0.097 - 0.107                      Preparation Date: 07/07/2017 Expires: 07/07/2018</p> <p>OPENED: 2/13/18</p> <p style="text-align: center;">RPC</p>	<p>Greg Marsh DMT Supervisor</p>
4/11/18	<p>Completed APM Protocol.</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll VFL</p>
6/26/18	<p><b>DMT External Std Solution</b></p> <p>Lot: 17-24-100 EtOH Certified Conc. 0.100                      Range: 0.095 - 0.105                      Prep Date: 11/14/17 Exp Date: 11/14/18</p> <p>Opened: 6/26/18 GM</p> <p style="text-align: center;">RPC</p>	<p>G. Marsh DMT Supervisor</p>
10/24/18	<p><b>DMT External Std Solution</b></p> <p>Lot: 18-05-100 EtOH Certified Conc. 0.100                      Range: 0.095 - 0.105                      Prep Date: 3/21/18 Exp Date: 03/21/19</p> <p>Opened: 10/24/18 GM</p>	<p>G. Marsh DMT Supervisor</p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/27/19	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 18-05-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 3/21/18                      Exp Date: 03/21/19</p> <p>Opened:            2/27/19 GM</p>	<p>Greg Marsh DMT SUPERVISOR</p>
4/10/19	<p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">ADM computers</p>	<p>OPATON VFL</p>
6/19/19	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 19-07-100    EtOH Certified Conc.    0.103</p> <p>Range:            0.098 - 0.108</p> <p>Prep Date: 5/31/19                      Exp Date: 05/31/20</p> <p>Opened:            6/19/19 GM</p>	<p>G. Marsh DMT SUPERVISOR</p>
10/22/19	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 19-07-100    EtOH Certified Conc.    0.103</p> <p>Range:            0.098 - 0.108</p> <p>Prep Date: 5/31/19                      Exp Date: 05/31/20</p> <p>Opened:            10/22/19 GM</p>	<p>G Marsh DMT SUPERVISOR</p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
4/10/20	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 19-13-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 9/23/19                      Exp Date: 09/23/20</p> <p>Opened: 1/10/20 <i>GM</i></p> <p style="text-align: center; font-size: 1.2em;">Solution was out of range</p>	<p><i>G. Marsh</i> DMT Supervisor</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 19-13-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 9/23/19                      Exp Date: 09/23/20</p> <p>Opened: 2/21/20 <i>GM</i></p> <p style="text-align: center; font-size: 1.2em;">RPC CHECK 2/21/20</p>	<p><i>G. Marsh</i> DMT Supervisor</p>
3/25/2020	<p>APM protocol completed</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>S. Stratton</i> (VFL)</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 20-04-100    EtOH Certified Conc.    0.101</p> <p>Range:            0.096 - 0.106</p> <p>Prep Date: 2/21/20                      Exp Date: 02/21/21</p> <p>Opened: 6/23/20 <i>GM</i></p> <p style="text-align: center; font-size: 1.2em;">RPC CHECK 6/23/20</p>	<p><i>G. Marsh</i> DMT SUPERVISOR</p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC Lot: 20-04-100 EtOH Certified Conc. 0.101 Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/21/20 Exp Date: 02/21/21</p> <p>Opened: 10/28/20 Gm</p>	<p>RPC G. Marsh 10/28/20</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Label Here]</p> <p>Lot: 20-14-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 9/22/20 Exp Date: 09/22/22</p> <p>Opened: 2/25/21 Gm</p>	<p>RPC G. Marsh 2/25/21</p>
3/22/2021	<p>APM protocol completed. Removed operator use log as they are no longer needed due to networking.</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p>S. Stratton (VFL)</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC Label Here]</p> <p>Lot: 21-05-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/3/21 6/28/21 Exp Date: 03/03/23</p> <p>Opened:</p>	<p>RPC G. Marsh 6/28/21</p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/12/2021	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 21-11-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 7/16/21                      Exp Date: 07/16/23</p> <p>Opened: <u>808 10/12/2021</u></p> <p>Installed DMT after cleaning at the VFL (removed DMT from agency 9/28/21).</p>	S. Stratton (VFL)
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 21-05-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 3/3/21                      Exp Date: 03/03/23</p> <p>Opened: <u>2/1/22 GM</u></p>	G. Marsh RPC CHECK 2/1/22
3/17/2022	<p>Completed APM protocol</p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	S. Stratton (VFL)
4/26/22	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 22-06-100    EtOH Certified Conc.    0.100</p> <p>Range:            0.095 - 0.105</p> <p>Prep Date: 3/7/22                      Exp Date: 03/07/24</p> <p>Opened: <u>4/26/22 GM</u></p> <p>Simulator out of range</p>	G. Marsh Solution Change

# DATAMASTER DMT MAINTENANCE LOG

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/20/22	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC Lot: 22-06-100 EtOH Certified Conc. 0.100 Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: 6/20/22 Gm</p>	<p>RPC Check</p> <p>G. Marsh</p> <p>6/20/22</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC Lot: 22-06-100 EtOH Certified Conc. 0.100 Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: 9/26/22 Gm</p>	<p>RPC Check</p> <p>G. Marsh</p> <p>9/26/22</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Lot: 22-14-100 EtOH Certified Conc. 0.098 Label Here]</p> <p>Prep Date: 10/5/22 Exp Date: 10/05/24</p> <p>Opened: 10/27/22 Gm</p> <p style="text-align: center;">Solution Change, Datamaster</p> <p style="text-align: center;">DUST OF RANGE</p>	<p>10/27/22</p> <p>G. Marsh</p>
	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC Lot: 22-14-100 EtOH Certified Conc. 0.098 Label Here]</p> <p>Prep Date: 10/5/22 Exp Date: 10/05/24</p> <p>Opened: 2/28/23 Gm</p>	<p>2/28/23</p> <p>G. Marsh</p> <p>RPC</p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/16/23	<p><i>Completed APM protocol</i></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>S. Stratton (VFL)</i></p>
<i>6/29/23</i>	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>Lot: 23-06-100    EtOH Certified Conc.    0.100    Label Here]</p> <p>Prep Date: 5/1/23                      Exp Date: 05/01/25</p> <p>Opened: <i>6/29/23 GM</i></p> </div> <p>[ For RPC o</p>	<p><i>RPC 6/29/23 CHECK G. Marsh</i></p>
9/17/2023	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>Lot: 23-08-100    EtOH Certified Conc.    0.099    Label Here]</p> <p>Prep Date: 7/19/23                      Exp Date: 07/19/25</p> <p>Opened: <del>8/28</del> <i>9/17/2023</i></p> </div> <p>[ For RPC o</p> <p><i>Installed post TSI.</i></p>	<p><i>S. Stratton (VFL)</i></p>
	<div style="border: 1px solid black; padding: 5px; margin-bottom: 5px;"> <p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>Lot: 23-06-100    EtOH Certified Conc.    0.100    Label Here]</p> <p>Prep Date: 5/1/23                      Exp Date: 05/01/25</p> <p>Opened: <i>10/23/23 GM</i></p> </div> <p>[ For RPC o</p>	<p><i>RPC CHECK 10/23/23 G. Marsh</i></p>

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency \_\_\_\_\_

DataMaster DMT Serial Number \_\_\_\_\_

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[ For RPC or Simulator Solution Change Place Label Here]	
10/26/23	Removed from Service for DMT Refurb. for maintenance, not refurb 808 11/28/2023	AP 10/26/23
	<b>DMT External Std Solution</b>	
11/28/2023	Lot: 23-08-100 EtOH Certified Conc. 0.099 [ For RPC ] Prep Date: 7/19/23 Exp Date: 07/19/25 Label Here] Opened: 808 11/28/2023	S. Stratton (VFL)
	Installed post TSI	
	<b>DMT External Std Solution</b>	
2/21/24	Lot: 23-08-100 EtOH Certified Conc. 0.099 [ For RPC ] Label Here] Prep Date: 7/19/23 Exp Date: 07/19/25 Opened: 2/21/24 GM	RPC CHECK C. MARSH
3/14/24	Completed APM protocol [ For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)

# DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/10/24	<p style="text-align: center;"><b>DMT External Std Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 23-08-100    EtOH Certified Conc.    0.099</p> <p>Prep Date: 7/19/23    Exp Date: 07/19/25</p> <p>Opened: 6/10/24 Gm</p>	<p style="text-align: center;">Gregory Marsh RPC Check</p>
	<p style="text-align: center;"><b>DMT Quality Control Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 24-10-100    EtOH Certified Conc.    0.101</p> <p>Prep Date: 7/29/24    Exp Date: 07/29/26</p> <p>Opened: 9/20/24 Gm</p>	<p style="text-align: center;">Gregory Marsh RPC Check</p>
	<p style="text-align: center;"><b>DMT Quality Control Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 24-10-100    EtOH Certified Conc.    0.101</p> <p>Prep Date: 7/29/24    Exp Date: 07/29/26</p> <p>Opened: 11/30/24 Gm</p>	<p style="text-align: center;">Gregory Marsh Solution Change</p>
	<p style="text-align: center;"><b>DMT Quality Control Solution</b></p> <p>[ For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 24-17-100    EtOH Certified Conc.    0.100</p> <p>Prep Date: 10/21/24    Exp Date: 10/21/26</p> <p>Opened: 11/15/25 Gm</p>	<p style="text-align: center;">Gregory Marsh Solution Change</p>

# DATAMASTER DMT MAINTENANCE LOG

VSP Derby  
Department/Agency

100138  
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/20/25	[ For RPC <b>DMT Quality Control Solution</b> e Label Here] Lot: 24-17-100      EtOH Certified Conc.      0.100 Prep Date: 10/21/24      Exp Date: 10/21/26 Opened: 2/20/25 OM	G. Marsh 2/20/25 RPC Check
3/11/25	APM	A. Bolduc VFL
6/25/25	[ For RPC <b>DMT Quality Control Solution</b> e Label Here] Lot: 24-17-100      EtOH Certified Conc.      0.100 Prep Date: 10/21/24      Exp Date: 10/21/26 Opened: 6/25/25 OM	G. Marsh RPC CHECK
8/8/25	[ For RPC or Simulator Solution Change Place Label Here] Removed for rehub	A. Bolduc
	[ For RPC or Simulator Solution Change Place Label Here]	