

DATAMASTER DMT MAINTENANCE LOG

Bennington Pol

Department/Agency

100139

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/11/11	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [For RPC or Simulator Solution Change Place Label Here] OPENED: 10/11/11</p>	Steven Harris
	<p>Install Per Protocol 10/11/11</p>	
11/10/11	<p>Replace Ink (Black) in Printer [For RPC or Simulator Solution Change Place Label Here] adjust Time -</p>	Laurzon
12/14/11	<p>Test [For RPC or Simulator Solution Change Place Label Here] "Interference" Reported Prior to test.</p>	Laurzon
12/19/11	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [For RPC or Simulator S OPENED: Sim Sol. Change</p>	Laurzon

DATAMASTER DMT MAINTENANCE LOG

100139

Bennington Rd

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/19/2012	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Updated Software</p>	Steven Harris
2/6/12	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>RPC</p> <p>DataMaster Simulator Solution Lot # 12-45-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>OPENED:</p>	Lawson
4/1/12	<p>adjust time - 1 hr. ahead (DST)</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	Lawson
4/27/12	<p>Correct Date & adjust time</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	Lawson

DATAMASTER DMT MAINTENANCE LOG

Bennington P.D.

Department/Agency

100139

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>6/11/12</i>	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013</p> <p>OPENED:</p>	<i>Lanzon</i>
<i>6/8/12</i>	<p style="text-align: center;"><i>Changed lock, software</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>update & APM protocol</i></p>	<i>A. B. R.</i>
<i>10/6/12</i>	<p style="text-align: center;">DataMaster Simulator Solution Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>RPC</i></p> <p>OPENED: <i>10-6-12</i> <i>adjust Time 6m is slow</i></p>	<i>Lanzon</i>
<i>11/4/12</i>	<p style="text-align: center;"><i>adjust Time - 1 hr back</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>Lanzon</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/2/13	<p>DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED: 2/2/13</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>RPC</p>	J. Stemp
3/10	<p>Adjust Time (DST)</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>1 hr. + 8 min Slow</p>	Lawson
5/31/13	<p>DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 5/31/13 RCD</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>performed sim solution change as it was out of range, and ran an APPL protocol</p>	R. Driscoll DMT TECH.
6/2/13	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>(See 5/31/13)</p> <p>RPC</p>	Lawson

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/23/13	<p>DataMaster Simulator Solution Lot # 13-03-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/03/2013 Expires: 04/03/2014</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>Solution Change</i></p>	<i>Lawson</i>
10/13/13	<p>DataMaster Simulator Solution Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>RPC</i></p>	<i>Stemp</i>
11/3/13	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Adjust Clock</i> <i>End (DST) 4:20 - 3:29 AM</i></p>	
2/7/14	<p>DataMaster Simulator Solution Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>RPC Failed</i> <i>Flow Rate too high</i> <i>out of service</i></p>	<i>Lawson</i>

DATAMASTER DMT MAINTENANCE LOG

Bennington PD
Department/Agency

100139
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/10/14	<p>DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014</p> <p>[For RPC e Label Here]</p> <p>OPENED: 2/10/14 RCD</p> <p>Flow rate too high during RPC. Replaced "T" with adjustable flow rate valve. Also found broken fitting so replaced. Set volume to 1.5L. Completed install.</p>	R. Driscoll
3/9/14	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Reset Clock PST 0632 - 0737</p>	Lauzon
5/9/14	<p>Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	R. Driscoll DMT TECH
6/18/14	<p>DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014</p> <p>[For RPC or e Label Here]</p> <p>OPENED:</p> <p>RPC adjust Clock 34 mins</p>	Lauzon

DATAMASTER DMT MAINTENANCE LOG

Bennington P.O.
Department/Agency

100139
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/5/14	<p>DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: 10/5/14</p> <p>[For RPC Label Here]</p> <p style="font-size: 1.5em; margin-left: 20px;">RPC</p>	Lanzon
04/3/14	<p style="font-size: 1.2em; text-align: center;">Adjust Clock - Back 1hr</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="font-size: 1.2em; text-align: center;">Check in Test</p>	Lanzon
01/17/15	<p>DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: 01/17/15</p> <p>[For RP se Label Here]</p> <p style="font-size: 1.5em; margin-left: 20px;">RPC</p>	Stamp
3/9/15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="font-size: 1.2em; text-align: center;">Adjust Time</p> <p style="font-size: 1.2em; text-align: center;">14:24 → 15:30</p>	Lanzon

DATAMASTER DMT MAINTENANCE LOG

Bennington PD

Department/Agency

100139

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/12/15	<p style="text-align: center;"><u>Completed APT Protocol</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">R. Driscoll DMT Tech.</p>
5/16/15	<p style="text-align: center;">DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: <u>5/16/15</u></p> <p>[For RPC Label Here]</p>	<p style="text-align: center;">Lanzon</p>
8/2/15	<p style="text-align: center;"><u>Diagnostic Test</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><u>Sim Solution Change</u></p> <p style="text-align: center;">DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: <u>8/2/15</u></p>	<p style="text-align: center;">Lanzon</p>
8/24/15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><u>Downloaded data to usb drive, uploaded new software, and installed new comm. PCB. No Ethernet port available at this time to connect to DPS network.</u></p>	<p style="text-align: center;">R. Driscoll DMT Tech.</p>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/16/15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>OPENED:</p> <p style="text-align: center; font-size: 1.5em;">RPC</p>	<p>Lawson</p>
12/21/15	<p>Accuracy Test</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Simulator Solution Change</p> <p>DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>OPENED:</p>	<p>Lawson</p>
02/15/16	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster External Standard Solution Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016</p> <p>OPENED:</p> <p style="text-align: center; font-size: 1.5em;">RPC</p>	<p>Stemp</p>
5/12/16	<p>ARM protocol completed</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll</p>

DATAMASTER DMT MAINTENANCE LOG

BENNINGTON PD.

100139

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/22/16	<p>DataMaster External Standard Solution Lot # 15-24-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 11/12/2015 Expires: 11/12/2016</p> <p>[For RPC or Label Here]</p> <p>OPENED: 5/22/16 <i>cc</i></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">RPC</p>	Lanzon
9/30/16	<p>DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017</p> <p>[For F Label Here]</p> <p>OPENED: 9/30/16 <i>ROD</i></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">Replaced simulator paddle and completed October RPC.</p>	R. Dixoll DMT Tech
2/12/17	<p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>[For RPC or Label Here]</p> <p>OPENED: 2/12/17 <i>cc</i></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">RPC</p>	Lanzon
5/10/17	<p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>[For RPC or Label Here]</p> <p>OPENED: 5/10/17 <i>ROD</i></p> <p style="font-size: 1.5em; font-family: cursive; text-align: center;">Install post simulator update.</p>	R. Dixoll VFL

DATAMASTER DMT MAINTENANCE LOG

Bennington PD

Department/Agency

100139

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>6/30/17</i>	<p>DMT Simulator Solution Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018 OPENED: <i>6/30/17</i> <i>(a)</i></p> <p>[For RPC] [Label Here]</p> <p><i>RPC</i></p> <p><i>Solution Change Protocol - entered .105 conc. instead of .100</i></p>	<i>LANZON</i>
<i>10/4/17</i>	<p>DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: <i>10-4-17</i> <i>CL</i></p> <p>[For RPC] [Label Here]</p> <p><i>RPC</i></p>	<i>LANZON</i>
<i>11/7/17</i> <i>STV</i>	<p><i>Installed version 2.00 PIC chip</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>R Driscoll</i> <i>VFC</i>
<i>11/8/17</i>	<p><i>Showed Internet Connection window</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Unplugged Internet - Restart.</i></p> <p><i>Ran Diagnostic Test</i></p>	<i>Lanzon</i>

DATAMASTER DMT MAINTENANCE LOG

Bennington PD
Department/Agency

100139
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/29/17	<p style="text-align: center;"><i>Installed new printer</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>P</i></p>	<p style="text-align: center;"><i>R. Discall</i> VFL</p>
2/11/18	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or DMT Simulator Solution label Here]</p> <p>Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018</p> <p>OPENED: _____</p>	<p style="text-align: center;"><i>Lanzon</i></p>
5/2/18	<p style="text-align: center;"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;"><i>R. Discall</i> VFL</p>
7/1/18	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18/ Exp Date: 02/28/19 Opened: <i>7/1/18</i></p>	<p style="text-align: center;"><i>Lanzon</i></p>

DATAMASTER DMT MAINTENANCE LOG

Bonn PD
Department/Agency

100139
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/3/18	<p align="center"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p align="center">DMT External Std Solution</p> <p>Lot: 18-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/18 Exp Date: 02/28/19</p> <p>Opened:</p> </div>	Lawson
2/25/19	<p align="center"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p align="center">DMT External Std Solution</p> <p>Lot: 18-16-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 10/29/18 Exp Date: 10/29/19</p> <p>Opened:</p> </div>	Lawson
4/30/19	<p align="center"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	J Durette (UFL)
6/16/19	<p align="center"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p align="center">DMT External Std Solution</p> <p>Lot: 18-16-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 10/29/18 Exp Date: 10/29/19</p> <p>Opened:</p> </div>	Stemp

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/10/19	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-06-100 EtOH Certified Conc. 0.098 Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened:</p>	<i>Stemp</i>
2/3/2020	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-06-100 EtOH Certified Conc. 0.098 Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened:</p>	<i>Lauran</i>
3/4/2020	<p style="text-align: center;"><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>J Dukette (VFL)</i>
6/17/20	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: <i>6/17/20</i></p>	<i>Lauran</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/18/20	<p style="text-align: center;"><i>Re-enter Correct Sim Lot #</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">6/17 - #19-06-100</p> <p style="text-align: center;">6/18 #19-16-100</p>	<i>Lanzon</i>
7/14/20	<p style="text-align: center;"><i>Software Update</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>[Signature]</i>
10/19/20	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 20-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/21/20 Exp Date: 02/21/21</p> <p>Opened: 10/19/20 <i>CEL</i></p> </div>	<i>Lanzon</i>
2/4/21	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin-top: 10px;"> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 20-14-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 9/22/20 Exp Date: 09/22/22</p> <p>Opened: 2/4/21 <i>JS</i></p> </div>	<i>Stemp</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/28/22	<p><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: 80%;"> <p style="text-align: center; margin: 0;">DMT External Std Solution</p> <p>Lot: 22-04-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 1/26/22 Exp Date: 01/26/24</p> <p>Opened: <u>6/30/22</u></p> </div>	S. Stratton (VFL)
6/30/22	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	Stemp
10/24/22	<p style="text-align: center;"><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: 80%;"> <p style="text-align: center; margin: 0;">DMT External Std Solution</p> <p>Lot: 22-06-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: <u>10/24/22</u></p> </div>	Stemp
2/24/23	<p><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	S. Stratton (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/22/24	<p><i>Completed APM protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>S. Stratton</i> <i>(VFL)</i></p>
6/23/24	<p><i>RPC</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 10px auto; width: fit-content;"> <p style="text-align: center; margin: 0;">DMT External Std Solution</p> <p style="margin: 0;">Lot: 23-07-100 EtOH Certified Conc. 0.100</p> <p style="margin: 0;">Prep Date: 6/20/23 Exp Date: 06/20/25</p> <p style="margin: 0;">Opened: <i>6/23/24</i> <i>JS</i></p> </div>	<p><i>Stemp</i></p>
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	