

DMT MAINTENANCE LOG

VSP Derby
Department/Agency

100140
DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/8/25	DMT Quality Control Solution [For RPC Lot: 25-15-100 EtOH Certified Conc. 0.101 e Label Here] Prep Date: 6/17/25 Exp Date: 06/17/27 Opened: Install	A. Bolduc VFL
10/23/25	DMT Quality Control Solution [For RPC Lot: 25-15-100 EtOH Certified Conc. 0.101 e Label Here] Prep Date: 6/17/25 Exp Date: 06/17/27 Opened: 10/23/25 Gm	G. Marsh RPC CHECK
12/30/25	DMT Quality Control Solution [For RPC or Lot: 25-26-100 EtOH Certified Conc. 0.100 ibel Here] Prep Date: 10/2/25 Exp Date: 10/02/27 Opened: 12/30/25 Gm	G. MARSH Sim Solution CHANGE (OUT OF RANGE)
2/27/26	DMT Quality Control Solution [For RPC o Lot: 25-26-100 EtOH Certified Conc. 0.100 Label Here] Prep Date: 10/2/25 Exp Date: 10/02/27 Opened: 2/27/26 Gm	G. MARSH RPC CHECK
5/15/26	DMT Quality Control Solution [For RPC Lot: 26-05-100 EtOH Certified Conc. 0.100 Label Here] Prep Date: 2/17/26 Exp Date: 02/17/28 Opened: 5/15/26 Gm	G. Marsh Sim Solution Change - OUT OF RANGE

DMT MAINTENANCE LOG

VSP Dorby
Department/Agency

100140
DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/25/26	DMT Quality Control Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 26-05-100 EtOH Certified Conc. 0.100 [bel Here] Prep Date: 2/17/26 Exp Date: 02/17/28 Opened: 6/25/26 Gm	G. Mmeslh RPC CHECK 6/25/26
7/28/26	removed for annual calibration [For RPC or Simulator Solution Change Place Label Here]	B. Johnson VFL
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	