

SOLUTION CHANGE

DataMaster DMT: 100140
Location: Bradford VSP
Date: 01/07/2017
Performed by: SEAN P BRENNAN




Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 16-23-100
Range = 0.096 - 0.106
Average = 0.098 g/210L
Std Dev = 0.0004

RED
119117

Simulator Temperature: 33.9°C

Performed by



Date

01/07/2017
