

SOLUTION CHANGE

DataMaster DMT: 100140
Location: Bradford VSP
Date: 04/07/2017
Performed by: SEAN P BRENNAN



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 17-10-100
Range = 0.096 - 0.106
Average = 0.098 g/210L
Std Dev = 0.0004

*RD
4/10/17*

Simulator Temperature: 33.8°C

Performed by

SP Brennan

Date

04/07/2017