

Simulator Temperature Check

Simulator Serial Number: MP3510

DMT Serial Number / Simulator Location: 100140 / VSP Bradford

NIST Traceable Thermometer Serial Number: 160294076

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	34.00	34.01	0.01	Y

* 34.0 +/- 0.2 degree Celsius acceptable range


Was temperature adjustment performed? Y / (N)

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By: 

Reviewed by: 

Date: 7/12/19

Date: 7/19/19