

SOLUTION CHANGE

DataMaster DMT: 100140
Location: BRADFORD VSP
Date: 12/22/2019
Performed by: SEAN P BRENNAN



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 19-13-100
Range = 0.095 - 0.105
Average = 0.099 g/210L
Std Dev = 0.0005

SP
12/23/2019

Simulator Temperature: 34.0°C

Performed by _____

Date 12/22/2019