

SOLUTION CHANGE

DataMaster DMT: 100140
Location: DERBY VSP
Date: 12/30/2025
Performed by: GREGORY R MARSH



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 25-26-100
Range = 0.095 - 0.105
Average = 0.100 g/210L
Std Dev = 0.0003

Brittany Smg 12/31/2025

Simulator Temperature: 34.0°C

Performed by

G. Marsh

Date

12/30/2025