

DATAMASTER DMT MAINTENANCE LOG

Randolph Pcl

Department/Agency

100143

DataMaster DMT Serial Number

P 1
10/27/11 - 1/31/12

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/27/11	DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [For RPC or Simulator Solution Change Place Label Here] OPENED: 10/27/11 Install Per Protocol	Steven Harnois Wt. L. Kelly Prog. ch. ref. org. chem.
11/8/11	[For RPC or Simulator Solution Change Place Label Here] Updated Software and Corrected Time	Steven Harnois
2/10/12	DataMaster Simulator Solution Lot # 12-32-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 [For RPC or Simulator Solution Change Place Label Here] OPENED: 2/10/12 Test R & L	Simpson Tom
3/11/12	[For RPC or Simulator Solution Change Place Label Here] Adj. time manually for DST	Simpson Tom

PC/21-11/5/12

DATAMASTER DMT MAINTENANCE LOG

P 2

RANDOLPH P. P

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/5/12	DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012 [For RPC] OPENED: 5/5/12 Solution Change	SIMPSON TOM
6/29/12	DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [For RPC] OPENED: 6/29/12 (TAS) RPC	SIMPSON TOM
8/15/12	[For RPC or Simulator Solution Change Place Label Here] Reformed APM	8/15/12
10/21/12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [For RPC] OPENED: 10/21/12 (TAS) RPC	SIMPSON TOM

DATAMASTER DMT MAINTENANCE LOG

P 3

RANDOLPH P.D.

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/16/12	Adj clock for EST [For RPC or Simulator Solution Change Place Label Here]	TOM S. M314
2/15/13	<u>DataMaster Simulator Solution</u> Lot # 12-72-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013 [For RPC e Label Here] OPENED: Test R.P.C.	TOM SIMPSON
6/30/13	<u>DataMaster Simulator Solution</u> Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 [For RPC e Label Here] OPENED: Test RPC	Tom Simpson
6/30	[For RPC or Simulator Solution Change Place Label Here] Adj time for DST	Tom Simpson

DATAMASTER DMT MAINTENANCE LOG

Randolph PD

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/8/13	<u>Completed APM protocol</u> [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.
11/1/13	DataMaster Simulator Solution Lot # 13-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/10/2013 Expires: 05/10/2014 OPENED: 11/1/13 RPC	Tom Simpson
11/8/12	Adj time for EST [For RPC or Simulator Solution Change Place Label Here]	Tom Simpson
2/21/14	DataMaster Simulator Solution Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014 OPENED: 2/21/14 <u>TAS</u> RPC	TOM SIMPSON

DATAMASTER DMT MAINTENANCE LOG

RANDOLPH P.D

100143

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/21/14	Adj clock time [For RPC or Simulator Solution Change Place Label Here]	TOM SIMPSON
3/28/14	Adj clock again for DST [For RPC or Simulator Solution Change Place Label Here]	TOM SIMPSON
6/20/14	DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: 6/20/14 1420 HRS [For RPC e Label Here]	TOM SIMPSON
8/2/14	reset clock to computer time [For RPC or Simulator Solution Change Place Label Here]	TOM SIMPSON

DATAMASTER DMT MAINTENANCE LOG

Randolph PD

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/13/14	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
10/12/14	DataMaster Simulator Solution Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015 OPENED: 10/12/14 [For RPC or Simulator Solution Change Place Label Here] Solution Change	Tom Simpson
10/18/14	[For RPC or Simulator Solution Change Place Label Here]	Tom Simpson
11/6/14	Set clock for EST [For RPC or Simulator Solution Change Place Label Here]	Tom Simpson

DATAMASTER DMT MAINTENANCE LOG

RANDOLPH P.D.

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/20/15	DataMaster Simulator Solution Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015 [For RP ice Label Here] OPENED: 2/20/15 <i>TAS</i> TEST RPC	Tom Simpson
3/8/15	[For RPC or Simulator Solution Change Place Label Here] Adjust clock for DST	Tom Simpson
6/29/15	DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 [For RPC ice Label Here] OPENED: Test RPC	Tom Simpson
7/23/15	DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 [For F Place Label Here] OPENED: 7/23/15 ROS Installation post ISI	R. Disroll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

Randolph PD

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/9/15	DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 OPENED: 10/09/15 [For RP] ace Label Here] Test / RPC	Steven Tiersch Officer DMT Supervisor
2/5/16	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 2/5/16 [For RP] ace Label Here] Test / RPC	Steven Tiersch Officer DMT Supervisor
6/24/16	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 6/24/16 [For RP] ace Label Here] Test / RPC	Steven Tiersch Officer DMT Supervisor
8/9/16	Completed APH Protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/21/16	DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 10/21/16 [For RPC ()] [Place Label Here] TEST / RPC	Steven Tiersch
		DMT Supervisor
02/25/17	DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED: 2/25/17 [For RPC ()] [Place Label Here] Test / RPC	Steven Tiersch
		DMT Supervisor
3/22/17	[For RPC or Simulator Solution Change Place Label Here] DMT returned to UFL for maintenance JSD 3/22/17	Jeff Dukette UFL
3/23/17	DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED: 3/23/17 JSD [For RPC ()] [Place Label Here] Install post sim conversion JSD 3/23/17	Jeff Dukette UFL

DATAMASTER DMT MAINTENANCE LOG

PLANDOLPH PD

Department/Agency

1000894

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/29/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 6/29/17 MFC [For RP ace Label Here] RPC ALL PASSED	MATTHEW CHAN PATROLMAN DMT SUP.
10/23/17	Installed version 2.00 PIC chip and completed AEM [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
10/29/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: MFC 10/29/17 [For RP ce Label Here] RPC - ALL PASSED	MATTHEW CHAN DMT SUP
2/28/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 2/28/18 MFC [For RPC or Label Here] RPC - ALL PASSED	MATTHEW CHAN DMT SUP.

DATAMASTER DMT MAINTENANCE LOG

Randolph PD

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/25/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Label Here] Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 6/25/18 (NFC) PASSED ALL TESTS	MATTHEW CHAN PASSEMAN
7/24/18	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
10/29/18 RPC	DMT External Std Solution Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Place Label Here] Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: 10/29/18 WBS	W. By Senior Deputy
2/11/2019	DMT External Std Solution Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Place Label Here] Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: 2/11/2019 (W)	Wayne Briggs Senior Deputy W. B.

DATAMASTER DMT MAINTENANCE LOG

OCSD *Randolph*

Department/Agency

100143

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-28-19	DMT External Std Solution Lot: 19-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 1/15/19 Exp Date: 01/15/20 Opened: 6-28-19 <i>Random OC17</i> Passed all tests	<i>JEAN MIGUEL</i> <i>BARBER</i> <i>SERGEANT OCSD</i>
7/15/19 <i>Moore CS</i> <i>JTV</i>	DMT removed for service [For RPC or Simulator Solution Change Place Label Here]	<i>J Dukette</i> <i>(VFL)</i>
7/30/19	DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 8/30 7/30/19 Install post maintenance	<i>J Dukette</i> <i>(VFL)</i>
10-31-19	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 10-31-19 <i>Random OC17</i> Failed RPC / RF detection 2nd attempt passed after talking to VEE	<i>JEAN-MIGUEL</i> <i>BARBER</i> <i>SERGEANT OC17</i>

DATAMASTER DMT MAINTENANCE LOG

ORANGE County Sheriff's Office Randolph
Department/Agency

160143
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02-29-20	DMT External Std Solution Lot: 19-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/17/19 Exp Date: 06/17/20 Opened: 02-29-20 [For RPC or Simulator Solution Change Place Label Here]	JMBREDAU SERGEANT
7/1/2020	DMT External Std Solution Lot: 20-05-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 4/27/20 Exp Date: 04/27/21 Opened: 7/1/2020 [For RPC or Simulator Solution Change Place Label Here]	W. B. Rogers Senior Det
7/19/2020	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
10/28/2020	DMT External Std Solution Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: 10/28/2020 [For RPC or Simulator Solution Change Place Label Here]	CPL Wayne Bayless


DATAMASTER DMT MAINTENANCE LOG

Orange County Sheriff

Department/Agency

10043

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02-19-21	DMT External Std Solution Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: 02-19-21 [For RPC Label Here]	 S. Stratton Sergeant
05-31-21	DIV I External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 05-31-21 [For R lace Label Here]	 S. Stratton Sergeant
7/8/2021	While performing the APM, found the sim jar had a crack in it. [For RPC or Simulator Solution Change Place Label Here] Removed the instrument from service until a new sim can be obtained.	S. Stratton (VFL)
7/19/2021	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 828 7/19/2021 [For RPC ce Label Here] New sim jar (MP3482). Completed SSC protocol and APM protocol. Instrument returned to service and ready for use.	S. Stratton (VFL)

DATAMASTER DMT MAINTENANCE LOG

ORANGE County Sheriff Dept (Rancho Ph VT)
Department/Agency

100143
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-31-21	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: <u>10-31-2021</u>	<i>Jean-Miguel Barteau</i> SERGEANT
2/28/22	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: <u>2/28/22</u>	<i>Wynne Bays</i> (General)
6/4		
7/3/22	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: <u>7/3/22</u>	<i>Jean-Miguel Barteau</i> SERGEANT
7/7/22	<i>Completed APM protocol</i> [For RPC or Simulator Solution Change Place Label Here]	<i>S. Stratton</i> (VFL)

DATAMASTER DMT MAINTENANCE LOG

Orange CA SO
Department/Agency

100143
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/28/2022	DMT External Std Solution [For R Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: 10/24/22 Place Label Here]	Wayne B. Joff CPL
11/29/23	DMT External Std Solution [For Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 11-29-23 e Place Label Here]	JM BARTON SERGEANT
6/27/23	DMT External Std Solution [For RPC Lot: 23-06-100 EtOH Certified Conc. 0.100 Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 6/27/23 e Label Here] Completed install to change name to Randolph PD and completed APM. Printer networking	J Duke (UFL)
10-31-23	DMT External Std Solution [For RP Lot: 23-06-100 EtOH Certified Conc. 0.100 Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 10-31-23 ace Label Here]	JM BARTON M312
11/31/24	DMT removed for refurbishment	J Duke (UFL)