

ROUTINE PERFORMANCE CHECK REPORT

DataMaster DMT: 100144
Location: SPRINGFIELD PD
Calibration Date: 04/14/2017
Certification Date: 04/14/2017
Installation Date: 04/18/2017
RPC Date: 10/06/2018
Supervisor Name: GREGORY J MOLGANO



Diagnostic Results

VERSIONS
DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube = 47.8°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.66 V
Cooler Voltage = 1.66 V
Bias Voltage = 80 V
Chopper Freq = 541 Hz

PUMP INFO

Flow Rate = 5.608 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.1717	0.1738
MIN(V)	0.1701	0.1724

FILTER INFO

Filter	Value	Zero
Filter 1	0.173	true
Filter 2	0.787	true
Filter 3	1.670	true

CALIBRATION CHECK

Xq = 0.082 0.96%

Routine Performance Check Passed

Accuracy and Precision Check


Concentration	= 0.101 g/210L
Lot #	= 18-11-100
Range	= 0.096 - 0.106
Average	= 0.102 g/210L
Std Dev	= 0.0000

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by

Gregory J. Molgano

Date

10/06/2018

Reviewed by

[Signature]

Date

10/8/18

[Signature]
10/8/18

DMT Serial Number #100144

Page 1 of 1

10/06/2018 2:23 AM