

SOLUTION CHANGE

DataMaster DMT: 100144
Location: SPRINGFIELD PD
Date: 01/01/2023
Performed by: RYAN P CLARK



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 22-14-100
Range = 0.093 - 0.103
Average = 0.099 g/210L
Std Dev = 0.0004

JSD 1/3/23

Simulator Temperature: 34.0°C

Performed by

Ryan Clark

Date

01/01/2023