

## SOLUTION CHANGE

DataMaster DMT: 100144  
Location: SPRINGFIELD PD  
Date: 09/09/2023  
Performed by: RYAN P CLARK



### Accuracy and Precision Check

Concentration = 0.100 g/210L  
Lot # = 23-01-100  
Range = 0.095 - 0.105  
Average = 0.101 g/210L  
Std Dev = 0.0000

SD 9/11/23

Simulator Temperature: 34.0°C

Performed by \_\_\_\_\_

Date 09/09/2023