

DataMaster Operator's Logbook Instructions

- 1) Pages should not be removed from this logbook.
- 2) At the top of each page, record the DataMaster agency and serial number.
- 3) Record EVERY test performed on the DataMaster.
- 4) Record the date, DMT operator's name, subject's name and the Simulator Vapor reading (s).
- 5) If not processing a subject, record test type under subject name (e.g. RPC or solution change).
- 6) This logbook is an agency record for your DataMaster DMT. When completed, file with other DataMaster DMT agency records.

241001

DATA MASTER DMT OPERATOR USE LOG

Brighton Pcd

Department/Agency

100145

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
9/9/11	Steven e Harnois	Install	.101	
9/17/11	Jeffrey Noyes	Cedamans, Michael	.099	
10/31/11	Jeff Noyes	RPC	.101	
11/20/11	Stephen Gilbert	Todd Scizza	.101	
12/23/11	Stephen Gilbert	Luke Dagle	.101	
1/6/12	Stephen Gilbert	Sheila Libby	.101	
3/8/12	Jeff Noyes	RPC	.100	
3/23/12	Stephen Gilbert	Brian Maxwell	.100	
4/14/12	Stephen Gilbert	Curt Widora	.100	
6/26/12	Trisha Conti	APM	.095	
11/30/12	Jeff Noyes	RPC	.100	
8/11/12	Stephen Gilbert	Kevin Guckin	.100	
8/24/12	Stephen Gilbert	James Peck	.100	
10/20/12	Stephen Gilbert	Eric Scott	.100	
11/02/12	Jeff Noyes	RPC	.099	
2/1/13	Jeff Noyes	RPC	.098	
5/3/13	Stephen Gilbert	Herbert Roby	.101	
05/11/13	AMY BROSARZ	STARE LZBBY	.096	
6/21/13	R. Driscoll Test / RPC PC	TEST / RPC	.099	—
6/21/13	R. Driscoll	TEST / APM	.099	—

DATAMASTER DMT OPERATOR USE LOG

Brighton PD

Department/Agency

100145

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/3	Stephen Gilbert	Michelle Verge	.099	
8/16	Stephen Gilbert	Gregory Wilcox	.099	
8/31	Stephen Gilbert	Jeremy Murphy	.097	
9/20	Stephen Gilbert	Ross Hackett	.096	
10/24	Noyes, Jeff	RPC	.098	
10/24	Noyes, Jeff			
2/22	Gilbert, Stephen	Alan Valmieris	.095	
2/27/14	Noyes, Jeff	RPC	.097	
3/15/14	Munson, Debra	Dawn Monfette	.097	.097
6/19/14	R. Driscoll	TEST / Admin	.094	5/15/16
6/19/14	J NOYES	RPC	.099	—
6/15/14	Gilbert, Stephen	Gazelle, Stacey	.098	
7/25/14	Gilbert, Stephen	Perkins, David	.097	.097
10/24/14	Noyes, Jeff	RPC	.100	—
2/28/15	Gilbert Stephen	Tedrauth, Tyler	.095	
2/28/15	Noyes, Jeff	RPC / Test	.095	
3/21/15	Gilbert Stephen	Stadon, Walter	.095	
04/06/15	Munson, Debra	Guillemette, Alain	.094	

DATAMASTER DMT OPERATOR USE LOG

Brighton PD
Department/Agency

100145
DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/3/13	Stephen Gilbert	Michelle Verge	.099	
8/16/13	Stephen Gilbert Gregory Wilcox	Gregory Wilcox	.099	
8/31/13	Stephen Gilbert	Jeremy Murphy	.097	
08/20/13	Stephen Gilbert	Rose Hackett	.096	
10/24	Noyes, Jeff	RPC/test	.098	
02/22/14	Gilbert, Stephen	Alain Vallieres	.095	
02/27/14	Noyes, Jeff	RPC/test	.097	
03/15/14	Munson, Debra	Dawn Montfette	.097	.097
06/19/14	Driscoll, R	Test /APM	.094	
06/19/14	Noyes, Jeff	RPC/test	.099	
06/15/14	Gilbert, Stephen	Gazaille, Stacey	.096	
7/25/14	Gilbert, Stephen	David Perkins	.097	.097
10/24/14	Noyes, Jeff	RPC/test	.100	
02/28/15 11/26/14	Noyes, Jeff	Stephen Abell	.099	.099
02/28/15	Gilbert, Stephen	Tyler Tebrouk	.095	
02/28/15	Noyes, Jeff	RPC/test	.099	
03/21/15	Gilbert, Stephen	Stanley, Walter	.095	
04/6/15	Munson, Debra	Alain Guillemette	.096	
03/14/15	Noyes, Jeff	Brian Colletti	.095	

DATAMASTER DMT OPERATOR USE LOG

Brighton

Department/Agency

100145

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
5/8/15	Stephen Gilbert	Lee, Lancia	.095	
06/26/15	Noyes, Jeff	RPC/Test	.099	
10/24/15	Noyes, Jeff	RPC/Test	.099	
11/5/15	R. Driscoll	TEST / APM	.097	—
03/7/16	R. Driscoll	TEST / RPC	.098	—
5/29/16	Noyes, Jeff	Worth, Rebecca	.099	—
06/30/16	Noyes, Jeff	RPC/Test		
08/03/16	Noyes, Jeff	Herold, Thomas	.097	✓
10/27/16	Noyes, Jeff	RPC/Test	.099	✓
10/28/16	R. Driscoll	TEST / APM	.098	—
12/17/16	Noyes, Jeff	Taylor, Chad	.099	
04/08/17	Noyes, Jeff	Colcombe, Craig	.099	—
03/16/17	Noyes, Jeff	RPC/Test	.099	
4/28/17	R. Driscoll	TEST / INSTALL	0.100	—
6/23/17	Noyes, Jeff	Merrill, Hardy	0.101	—
6/28/17	Noyes, Jeff	RPC/Test	0.101	—
11/17/17	Noyes, Jeff	RPC/Test	0.100	—
11/18/17	Stetson, Lynn			
1/11/18	Noyes, Jeff	Stetson, Lynn	.100	—
2/27/18	Noyes, Jeff	RPC	.100	✓

BRIGHTON

100145

[illegible]

DATAMASTER DMT OPERATOR USE LOG

Department/Agency

DataMaster DMT Serial Number

[illegible]