

DATAMASTER DMT OPERATOR USE LOG

Brighton PD

Department/Agency

100145

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
8/3/13	Stephen Gilbert	Michelle Verge	.099	
8/10/13	Stephen Gilbert Gregory Wilcox	Gregory Wilcox	.099	
8/31/13	Stephen Gilbert	Jeremy Murphy	.097	
08/20/13	Stephen Gilbert	Rose Hackett	.096	
10/24	Noyes, Jeff	RPC/test	.098	
02/22/14	Gilbert, Stephen	Alain Vallieres	.095	
02/27/14	Noyes, Jeff	RPC/test	.097	
03/15/14	Munson, Debra	Dawn Monfette	.097	.097
06/19/14	Driscoll, R	Test /APM	.094	
06/19/14	Noyes, Jeff	RPL/test	.099	
06/15/14	Gilbert, Stephen	Grazzillo, Stacey	.096	
7/25/14	Gilbert, Stephen	David Perkins	.097	.097
10/24/14	Noyes, Jeff	RPL/test	.100	
08/25/15 11/26/14	Noyes, Jeff	Stephen Abell	.099	.099
02/28/15	Gilbert, Stephen	Tyler Tetreault	.095	
02/28/15	Noyes, Jeff	RDC/test	.099	
03/21/15	Gilbert, Stephen	Stanley Walter	.095	
04/6/15	Munson, Debra	Alain Guillemette	.096	
02/14/15	Noyes, Jeff	Brian Colletti	.095	

DATAMASTER DMT MAINTENANCE LOG

Brighton PD
Department/Agency100145
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/19/14	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
0/19/14	<u>DataMaster Simulator Solution</u> [For RP Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: RPC	J NOYES
10/24/14	<u>DataMaster Simulator Solution</u> [For F Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: RPC	J Noyes
2/25/15	<u>DataMaster Simulator Solution</u> [For RPC Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 OPENED:	J Noyes

Breath Record History

Page 1
04/18/2015

Test Date	Test Time	Agency	Operator Name	Alcohol Val. 1	Alcohol Val. 2	Sim. Conc. 1	Sim. Conc. 2	Status Message
04/06/2015	00:15:19	DEPT WSP	DEPT. MUNSON	0.167		0.096		RESPONSE TIMEOUT
03/21/2015	20:40:12	BRIGHTON PD	STEPHEN M GILBERT			0.097		RESPONSE TIMEOUT
03/14/2015	18:47:13	BRIGHTON Police	JEFFREY M NOYES	REFUSED		0.095		
02/28/2015	00:37:44	BRIGHTON PD	STEPHEN M GILBERT	0.046		0.096		
11/26/2014	16:17:32	BRIGHTON PD	JEFFREY M NOYES	0.128		0.097		
11/26/2014	18:06:36	BRIGHTON PD	JEFFREY M NOYES	0.130		0.096	0.097	RESPONSE TIMEOUT