

DMT Upgrade Checklist

DMT Serial # 100145

	Initials	Date
Property Identification Tag # <u>1000902</u>	JSD	1/15/25
Replace Serial # sticker	JSD	1/15/25
Affix QR code sticker	JSD	1/15/25
Install current software (43393A-B)	JSD	1/15/25
Connect printer and ensure it prints color	JSD	1/21/25
Perform Adjustment Protocol	JSD	1/21/25
Perform Calibration Protocol	JSD	1/21/25
Run MCS Solution with calibration	JSD	1/21/25
Complete Verification Testing	JSD	1/21/25 ²³ JSD 1/25/25
Linearity 1	JSD	1/21/25
Linearity 2	JSD	1/22/25
Linearity 3	JSD	1/23/25
Interference solutions (Acetone, MeOH, IPA)	JSD	1/21/25
Reviews completed and DMT ready for installation	JSD	1/31/25
Install DMT	JSD	1/31/25
Location: <u>USP Sr Johnsbury</u>	JSD	1/31/25
12V500 s/n: <u>MP3508</u>	JSD	1/31/25
New log book	JSD	1/31/25
Troubleshooting Guide	JSD	1/31/25
Fill out DMT electronic log	JSD	1/31/25