

SOLUTION CHANGE

DataMaster DMT: 100145

Location: LAMOILLE COUNTY SO

Date: 5/25/2026

Performed by: BRAD E CARRIERE



Accuracy and Precision Check

Concentration = 0.101 g/210L

Lot # = 25-27-100

Range = 0.096 - 0.106

Average = 0.100 g/210L

Std Dev = 0.0003

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "Brad E. Carriere", is written over a horizontal line.

Date

5/25/2026