

DataMaster DMT Monthly Update Check-list

Department/Agency: VSP-Middlesex

DM serial #: 100146

Supervisor: Tpr. C. Maurice

Date: 05/19/16

* April 2016 *

Provided

None this month

* Operators Logbook	☒	☐
* Maintenance Logbook	☐	☒
* Breath Test History Report	☐	☒
* Status History Report	☐	☒
* Log On History Report	☐	☐
Calibration Report	☐	☐
Certification Report	☐	☐
Installation Report	☐	☐
* Annual Preventative Maintenance	☐	☒
* Routine Performance Check	☐	☒
* Simulator Solution Change	☐	☒
Other Reports	☐	☐

List other report type and date:

DATA MASTER DMT OPERATOR USE LOG

VSP Middlesex
Department/Agency

100146

DataMaster DMT Serial Number

Date	DMT Operator's Name	Subject's Name	Simulator Vapor 1	Simulator Vapor 2
1.8.16	C. Brown	Spencer Jacobs	.098	.097
1/12/16	S. PRACK	Kira Clements	.099	.098
3/8/16	R. Driscoll	TEST / INSTALL	.101	—
03/08/16	J. VOORIS	Thomas Long	.101	
3.10.16	C. Brown	Bridgette A. Lindsay	.101	.101
03/10/16	C. Brown	KRISTINA BISBEE	.099	
3/12/16	Maurice	Richard Tachette	.100	
3.26.16	C. Brown	Paul Reynolds	.100	
4/2/16	Maurice	Lauren Kithredge	.100	
04/07/16	PHELPS	DARREN JONES	.099	.098
04/11/16	PHELPS	AMANDA ST. CYR	.099	
04/11/16	PHELPS	AMANDA ST. CYR	.099	.098
4/12/16	MAZZILLI	Jeff Fitzbinner	.099	
04/13/16	PHELPS	MICHAEL MORSE	.098	.098
04/13/16	VOORIS	ESPERANZA ASH	.099	
4/22/16	C. Brown	Dennis Diego	.098	
5/1/16	Boyslawski	Jordan Stark		
05/13/16	PARLIN	EDWARD KILEY	.098	.098
5/19/16	VOORIS	Daniel Heyward	.098	