

SOLUTION CHANGE

DataMaster DMT: 100146
Location: MIDDLESEX VSP
Date: 11/21/2019
Performed by: CRISTA MAURICE



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 19-10-100
Range = 0.096 - 0.106
Average = 0.099 g/210L
Std Dev = 0.0004

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "Crista Maurice", is written over a horizontal line.

Date

11/21/2019