

12/8/11 - 1/10/25

DATAMASTER DMT MAINTENANCE LOG

Berlin Pol

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/8/11	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [Range Place Label Here] OPENED: 12/8/11 Install Per Procedure	Steven Harnois <i>Ed 1/10/11 VDHL Prog. Chert.</i>
2/17/12	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [Range Place Label Here] OPENED: RPC	Chad Bassett
4/7/12	DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [Range Place Label Here] OPENED: Solution Change	Chad Bassett
6/7/12	DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [Range Place Label Here] OPENED: MARK MONTAITH	MARK MONTAITH Chad Bassett

DATAMASTER DMT MAINTENANCE LOG

Bellingham

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/19/12	<u>DataMaster Simulator Solution</u> Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: 10/19/12 [For RPC or Simulator Solution Change Place Label Here]	CTB RPC
11/15/12	[For RPC or Simulator Solution Change Place Label Here] APM protocol performed	R. Driscoll DMT TECH.
2-2-13	<u>DataMaster Simulator Solution</u> Lot # 12-90-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/18/2012 Expires: 12/18/2013 OPENED: RPC	MW MW
2-2-13	<u>DataMaster Simulator Solution</u> Lot # 12-90-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/18/2012 Expires: 12/18/2013 OPENED: Simulator Solution Change	MW MW



DATAMASTER DMT MAINTENANCE LOG

Bellin R.D

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-27-13	<u>DataMaster Simulator Solution</u> Lot # 12-84-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 9/24/2012 Expires: 9/24/2013 OPENED: _____ [For RPC or Simulator Solution Change Place Label Here]	mm
6-27-13	RPC	mm
10/5/13	[For RPC or Simulator Solution Change Place Label Here] <u>DataMaster Simulator Solution</u> Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 OPENED: 10 10/5/13	
10/5/13		
11/13/13	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Dinscull DMT TECH.
12/29/13	[For RPC or Simulator Solution Change Place Label Here] <u>DataMaster Simulator Solution</u> Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 OPENED: 12/29/13 Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Berlin P.D

Department/Agency

100148


DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/5/14	<u>DataMaster Simulator Solution</u> [For RPC Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014 OPENED: 3/5/14 CTB	CPL. Bassett
6-11-14	<u>DataMaster Simulator Solution</u> [For RPC or Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 6-11-14 Adjusted clock +7min	Sgt Montlitt
6-29-14	<u>DataMaster Simulator Solution</u> [For RPC or Lot # 14-01-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 1/9/2014 Expires: 1/9/2015 OPENED: 8-29-14 Solution Change only	Sgt Montlitt
10/25/14	<u>DataMaster Simulator Solution</u> [For RF Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: 10/25/14 OS	RPC OS

DATAMASTER DMT MAINTENANCE LOG

Berlin PD
Department/Agency

100148
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/12/14	Completed APT Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.
2/7/15	DataMaster Simulator Solution Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: <u>CS 2/7/15</u> [For RPC or Simulator Solution Change Place Label Here] RPC	
3.9.15	[For RPC or Simulator Solution Change Place Label Here] Changed Clock + 9 min	MM
6/13/15	DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 OPENED: <u>6/13/15 CS</u> [For RPC or Simulator Solution Change Place Label Here] RPC	

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/2/15	Downloaded data to USB drive, uploaded new [For RPC or Simulator Solution Change Place Label Here] software, installed new camera, PCB and completed APM protocol. Connected to DPS network	R. Driscoll DMT TECH
9.26.15	DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016 OPENED: 9.26.15 [For] Place Label Here	M. [Signature] Test / RPC
11/27/15	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 11/24/15 [For RP] ce Label Here Solution Change	[Signature] SET
2/14/15	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 2/14/16 [For R] Place Label Here RPC	[Signature]

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100146

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-7-16	<u>DataMaster External Standard Solution</u> Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 06-07-16 <i>mu</i> [For RF Place Label Here]	<i>Mark Martin</i>
9/21/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
10/9/16	RPC [For RF Place Label Here] <u>DataMaster External Standard Solution</u> Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 OPENED: 10/9/16 <i>OTB</i>	<i>C. Fasset</i>
10/10/16	Simulator jar was not sealed, fixed leak and completed RPC protocol.	R. Driscoll DMT Tech.
2-27-17	<u>DataMaster External Standard Solution</u> Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017 OPENED: 2-27-17 <i>ma</i> [For RPC Place Label Here]	<i>Mark Martin</i>

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/20/17	[For RPC or Simulator Solution Change Place Label Here] DMT returned to UFL for routine maintenance JSD 3/20/17	Jeff Dukette UFL
3/21/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 [For RPC e Label Here] OPENED: 3/21/17 JSD Install, post simulator update JSD 3/21/17	Jeff Dukette UFL
6/6/17 R.P.C.	DMT Simulator Solution Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 [For RPC e Label Here] OPENED: 6-6-17	Mantente
9/26/17 R.P.C.	DMT Simulator Solution Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 [nge Place Label Here] OPENED: 9-26-17	Buckette

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/20/17	Installed version 2.08 PTC chip and completed APM protocol. [For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 OPENED: 1-14-18	R. Driscoll VFL
01-14-18 R.P.C.		Monteith
2/23/18	RPC [For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 OPENED: 2/23/18	
2/27/18	DMT External Std Solution Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18 Opened: 2/27/18 [For RPC or Simulator Solution Change Place Label Here] leak checked simulator and replaced solution. performed solution change protocol.	R. Driscoll VFL
6/3/18	DMT External Std Solution Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: 6/3/18 [For RPC or Simulator Solution Change Place Label Here] RPC	

DATAMASTER DMT MAINTENANCE LOG

Berlin, CT

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>7/31/18</i>	<i>Completed APM Protocol</i> [For RPC or Simulator Solution Change Place Label Here]	<i>R. Driscoll</i> <i>VFL</i>
<i>10/7/18</i>	DMT External Std Solution [For F Lot: 18-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/18 Exp Date: 02/28/19 Opened: <i>10/7/18 SGT</i> <i>RPC</i>	<i>Steven</i> <i>Tiersch</i> <i>DMT</i> <i>Supervisor / Officer</i>
<i>12/23/18</i>	DMT External Std Solution [For RP Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: <i>12/23/18 SGT</i> <i>Solution Change</i>	<i>Steven</i> <i>Tiersch</i> <i>DMT Supervisor</i> <i>Officer</i>
<i>2/2/19</i>	DMT External Std Solution [For RP Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: <i>2/2/19 BAC</i> <i>KPC</i>	<i>BAC</i> <i>CAVARETTA</i> <i>DMT Supervisor</i> <i>Officer</i>

DATAMASTER DMT MAINTENANCE LOG

Bear PD

Department/Agency

100178

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/1/19	DMT External Std Solution [For RPC Lot: 18-19-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 6/1/19 BAC ILPC	<i>B. Cavanaugh</i> DMT Supervisor Officer
7/30/19	<i>Completed ARM protocol</i> [For RPC or Simulator Solution Change Place Label Here]	<i>J. Dukette</i> (VFL)
10/1/19	DMT External Std Solution [For RP Lot: 18-19-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 10/1/19 BAC ILPC	<i>B. Cavanaugh</i> DMT Supervisor Officer
2/4/20	DMT External Std Solution [For RPC Lot: 19-06-100 EtOH Certified Conc. 0.098 Place Label Here] Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened: 2/4/20 SGT RPC	<i>Steven Tiersch</i> DMT Supervisor Officer

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

100148

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/10/20	[For DMT External Std Solution Place Label Here] Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 6/10/20	RPC
7/14/2020	Updated software. Completed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	J. Dukette (UIFL)
7.27.20	Test	Test
7/27/20 7/27/20	DMT External Std Solution [For RF Lot: 19-16-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 7/27/20 7/27/20 solution change	Steven Tiersch
7/31/20	DMT External Std Solution [For RPC Lot: 20-11-100 EtOH Certified Conc. 0.100 Place Label Here] Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 7/31/20 Solution Change	Steven Tiersch

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

10018ST48

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/28/20	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 10/28/20 RPC	Steven Tiersch Officer Data Master sup
1/20/2021	DMT External Std Solution Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: 1/20/2021 Solution Change	Steven Tiersch Officer
2/17/21	DMT External Std Solution Lot: 20-05-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 4/27/20 Exp Date: 04/27/21 Opened: 2/17/21 RPC	Steven Tiersch Officer Data Master Sup
5-11-21	DMT External Std Solution Lot: 21-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: 5-11-21 Solution Change	Mark Mantel

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/12/21	DMT removed for maintenance [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)
5/27/21	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 e Label Here] Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: JSD 5/27/21 Install post TSI	J Dukette (VFL)
7/17/21	DMT removed for service [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)
9/10/21	DMT External Std Solution [For RP Lot: 21-11-100 EtOH Certified Conc. 0.100 ace Label Here] Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: JSD 9/10/21 Install post TSI Completed APM protocol	J Dukette (VFL)


DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/24/21	DMT External Std Solution Lot: 21-04-100 EtOH Certified Conc. 0.100 [Label Here] Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: 10/24/21 ST RPC	Steven Tiersch Data master SUP
1/5/22	DMT External Std Solution [For RPC Lot: 21-06-100 EtOH Certified Conc. 0.100 Label Here] Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened:	Solution Change 
2/24/22	DMT External Std Solution [For RPC Lot: 21-11-100 EtOH Certified Conc. 0.100 Label Here] Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: 2/24/22 ST RPC	Steven Tiersch Datamaster SUP
2/27/22	DMT External Std Solution [For Lot: 21-06-100 EtOH Certified Conc. 0.100 Place Label Here] Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 2/27/22 	Solution Change 

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/20/22	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 22-10-100 EtOH Certified Conc. 0.100 Label Here] Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: 5/20/22 Install post TSI	J. Dukette (VFL)
7/18/22	Completed ARM [For RPC or Simulator Solution Change Place Label Here]	J. Dukette (VFL)
10/28/22	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 22-12-100 EtOH Certified Conc. 0.101 Label Here] Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 10/28/22 (CW) RPC	L. Willey DATAMASTER SUP
2/1/23	DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 21-06-100 EtOH Certified Conc. 0.100 Label Here] Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 2/1/23 (CW) Feb month RPC	L. Willey DATAMASTER SUPERVISOR

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/29/23	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 1/25/23 Exp Date: 01/25/25 Opened: 3/29/23 LW SSC	OFC Levi Willey DATAMASTER SUPERVISOR
5/23/23	DMT External Std Solution Lot: 23-01-100 EtOH Certified Conc. 0.100 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 1/25/23 Exp Date: 01/25/25 Opened: 5/23/23 LW RPC (pass on 5/24)	OFC Levi Willey Datamaster Supervisor
7/6/23	Adjusted tubing. Completed APM. Recalibrated touch screen. [For RPC or Simulator Solution Change Place Label Here]	S Stratton (VFL)
7-31-23	DMT External Std Solution Lot: 23-07-100 EtOH Certified Conc. 0.100 [For RPC or Simulator Solution Change Place Label Here] Prep Date: 6/20/23 Exp Date: 06/20/25 Opened: 7-31-23 Solution Change	[Signature]

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/16/23	DMT External Std Solution	Tiersch
	Lot: 23-07-100 EtOH Certified Conc. 0.100 [For RPC Prep Date: 6/20/23 Exp Date: 06/20/25 ce Label Here] Opened: 9/16/23 ST	
	RPC	
11/4/23	DMT External Std Solution	Tiersch
	Lot: 23-06-100 EtOH Certified Conc. 0.100 [For RPC Prep Date: 5/1/23 Exp Date: 05/01/25 e Label Here] Opened: 11/4/23 ST	
	Solution Change	
12/8/23	DMT External Std Solution	Wiley
	Lot: 23-06-100 EtOH Certified Conc. 0.100 [For RPC Prep Date: 5/1/23 Exp Date: 05/01/25 e Label Here] Opened: 12/8/23 LW	
	SSC	
11/12/24	DMT External Std Solution	Tiersch
	Lot: 23-10-100 EtOH Certified Conc. 0.099 [For RPC Prep Date: 11/28/23 Exp Date: 11/28/25 ice Label Here] Opened: 11/12/24 SGT	
	Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

100148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/13/24	Report of filter wheel errors. Opened top and manually unscrewed wheel. Ran 2 filter wheel tests w/ no errors. Passing diagnostic performed. A sim solution change protocol needs to be completed prior to return to service	J Outkate (VFL)
1/13/24	Solution Protocol as the STAP came to fix the problem with the OFFICER Tiersch already changed solution on 1/12/24. Passed all tests & back in operation - emailed the STAP	
2/4/24	[For RPC or Simulator Solution Change Place Label Here] STEVEN RPC	DMT SUP
	DMT External Std Solution Lot: 23-10-100 EtOH Certified Conc. 0.099 2/4/24 SGT Prep Date: 11/28/23 Exp Date: 11/28/25	Steven Tiersch MPB
4/14/24	[For RP DMT External Std Solution Lot: 23-10-100 EtOH Certified Conc. 0.099 Prep Date: 11/28/23 Exp Date: 11/28/25 Opened: 4/14/24 SGT Solution change	S Tiersch DMT SUP

Berlin PD
Department/Agency

100148
DataMaster DMT Serial Number

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Berlin PD

Department/Agency

160148

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/16/24	DMT Quality Control Solution Lot: 24-07-100 EtOH Certified Conc. 0.100 [For RP] [Place Label Here] Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: 11/16/24 [Signature]	C. Bassett
11/21/24	DMT Quality Control Solution Lot: 24-02-100 EtOH Certified Conc. 0.099 [For RP] [Place Label Here] Prep Date: 4/9/24 Exp Date: 04/09/26 Opened: 11/21/24 ST Solution change	S. Tiersch
12/18/24	DMT Quality Control Solution Lot: 24-07-100 EtOH Certified Conc. 0.100 [For RP] [Place Label Here] Prep Date: 5/29/24 Exp Date: 05/29/26 Opened: 12/18/24 LW SSC	L. Willey DMT Supervisor
12/20/24	DMT Quality Control Solution Lot: 24-17-100 EtOH Certified Conc. 0.100 [For RP] [Place Label Here] Prep Date: 10/21/24 Exp Date: 10/21/26 Opened: 12/20/24 LW SSC	L. Willey DMT Supervisor

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/10/25	DMT Quality Control Solution [For RPC (Lot: 24-24-100 EtOH Certified Conc. 0.099 Label Here] Prep Date: 12/18/24 Exp Date: 12/18/26 Opened: 11/10/25 JSO Interference in sim. DMT remained OOS JSO 11/10/25	J Duket (WFO)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	