

DATAMASTER DMT MAINTENANCE LOG

Bristol Pd

100150

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/22/12	<p>DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013</p> <p>[F] _____ ge Place Label Here]</p> <p>OPENED: <u>2/22/12</u></p> <p align="center">Installation</p>	<p align="center"><i>Steven Harnois</i></p> <p align="center"><i>Sub</i></p>
10-30-12	<p>DataMaster Simulator Solution Lot # 12-72-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013</p> <p>[For RPC] _____ ce Label Here]</p> <p>OPENED: _____</p> <p align="center">RPC</p>	<p align="center"><i>[Signature]</i></p>
1/25/13	<p>DataMaster Simulator Solution Lot # 12-72-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013</p> <p>[For RPC] _____ Label Here]</p> <p>OPENED: <u>1/25/13</u> <u>RPC</u></p> <p>Amount preventative maintenance fee</p> <p align="center">APM completed, solution out of range so performed solution change / RPC (RPC is due in Feb. so completed)</p>	<p align="center"><i>R. Drson DMT TECH.</i></p>
6-16-13	<p>DataMaster Simulator Solution Lot # 12-89-100 EtOH_{aq} Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013</p> <p>[For RPC or] _____ abel Here]</p> <p>OPENED: _____</p> <p align="center">RPC TEST COMPLETED NEW BASE RADIDS SET OFF RF - WARNING SW ATTACHED X RADID</p>	<p align="center"><i>[Signature]</i></p>

DATAMASTER DMT MAINTENANCE LOG

Bristol PD

Department/Agency

100150

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/30/13	<p>DataMaster Simulator Solution Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 [Label Here]</p> <p>OPENED: 9/30/13 RCD</p> <p>Completed site evaluation for new PD & performed Install protocol.</p>	<p>R. Drison DMT TECH.</p>
10-20-13	<p>DataMaster Simulator Solution Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 [Label Here]</p> <p>OPENED:</p> <p>INSTALLED NEW PROCESSING COUNTER & SOLUTION CHANGE</p>	<p><i>[Signature]</i></p>
9-31-14	<p>DataMaster Simulator Solution Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014 [Label Here]</p> <p>OPENED:</p> <p style="font-size: 2em; text-align: center;">RPC</p>	<p><i>[Signature]</i></p>
4-18-14	<p>DataMaster Simulator Solution Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014 [Label Here]</p> <p>OPENED: 4-18-14</p> <p>SOLUTION CHANGE SIM OUT OF RANGE</p>	<p><i>[Signature]</i></p>




DATAMASTER DMT MAINTENANCE LOG

Bristol P.D.

Department/Agency

100150

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7-15-14	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 3/12/2014 Expires: 3/12/2015</p> <p>OPENED: _____</p> <p style="font-size: 2em; text-align: center;">RPC</p>	
10/7/14	<p style="text-align: center;">Completed APM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll</p>
11-24-14	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 14-12-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/10/2014 Expires: 9/10/2015</p> <p>OPENED: <u>11-24-14</u></p> <p style="font-size: 2em; text-align: center;">RPC</p>	
2-16-15	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015</p> <p>OPENED: <u>2-16-15</u></p> <p style="font-size: 2em; text-align: center;">RPC</p>	


DATAMASTER DMT MAINTENANCE LOG

BELSTON P.D.

100150

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7-4-15	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015</p> <p>OPENED: 7-4-15</p> <p style="text-align: center; font-size: 2em;">RPC</p>	
	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016</p> <p>7/13/15 JCS</p> <p style="text-align: center;">Simulator Solution Change</p>	
8/3/15	<p>Downloaded data to usb drive, uploaded new software,</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>installed new comm PCB, and connected to DES network</p> <p>Tried to run RPC and having filter wheel errors. Taking the instrument back to the lab for repairs. RCD 8/3/15</p>	<p>R. Driscoll DMT Tech</p>
8/4/15	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>OPENED: 8/4/15 RCD</p> <p style="text-align: center;">Installation post TSI</p>	<p>R. Driscoll DMT Tech</p>

DATAMASTER DMT MAINTENANCE LOG

Bristol PD

Department/Agency

160150

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-20-15	<p>DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016</p> <p>[For RPC or S _____] [Label Here]</p> <p>OPENED: _____</p> <p style="font-size: 1.5em; color: blue;">RPC COMPLETED</p>	
3-5-16	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016</p> <p>[For RPC or _____] [Label Here]</p> <p>OPENED: _____</p> <p style="font-size: 1.5em; color: blue;">RPC COMPLETED</p>	
6-30-16	<p>DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016</p> <p>[For RPC or _____] [Label Here]</p> <p>OPENED: _____</p> <p style="font-size: 1.5em; color: blue;">RPC FAILED 12 ATTEMPTS</p>	
7/5/16	<p>DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017</p> <p>[For RPC or _____] [Label Here]</p> <p>OPENED: 7/5/16 RED</p> <p>replaced leaking simulator with Guth s/n G 8970, matched temps using NIST Thermometer 111814367. Completed RPC protocol.</p>	<p>R. Driscoll DMT Tech.</p>

DATAMASTER DMT MAINTENANCE LOG

Bristol PD

Department/Agency

100150

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/11/16	<p style="text-align: center;"><u>Completed APM Protocol</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R. Driscoll DMT Tech.</p>
10-8-16	<p style="text-align: center;"><u>DataMaster External Standard Solution</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017</p> <p>OPENED: 10-8-16 KEG</p> <p style="font-size: 2em; text-align: center;">RPC COMPLETED</p>	<p style="text-align: right;"><i>[Signature]</i></p>
2-19-17	<p style="text-align: center;"><u>DMT External Standard Solution</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>OPENED: 2-19-17 KEG</p> <p style="font-size: 1.5em; text-align: center;">RPC COMPLETED</p> <p style="font-size: 1.2em; text-align: center;">INSTALLED 2 PRINTER CARTRIDGES IN PRINTER</p>	<p style="text-align: right;"><i>[Signature]</i></p>
3/13/17	<p style="text-align: center;"><u>DMT Simulator Solution</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018</p> <p>OPENED: 3/13/17 R00</p> <p style="text-align: center;"><u>Install post simulator replacement</u></p>	<p>R. Driscoll DMT Tech.</p>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7-4-17	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 5px 0;"> <p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>OPENED: 7-4-17 KEG</p> </div> <p style="font-size: 2em; font-weight: bold; margin-top: 10px;">RPC</p>	<p style="font-size: 1.5em; font-weight: bold;">Steve [Signature]</p>
10/19/17	<p>Installed version 2.09 PIC chip and completed APM protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="font-size: 1.2em;">R. Drisoll VFL</p>
11/27/17	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 5px 0;"> <p>DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018</p> <p>OPENED: 10/27/17</p> </div> <p style="font-size: 2em; font-weight: bold; margin-top: 10px;">RPC Completed</p>	<p style="font-size: 1.2em;">11/27/17</p>
	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <div style="border: 1px solid black; padding: 5px; margin: 5px 0;"> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18</p> <p>Opened: STS 2/22/18</p> </div> <p style="font-size: 1.5em; font-weight: bold; margin-top: 10px;">RPC completed</p>	

DATAMASTER DMT MAINTENANCE LOG

Bristol PD

Department/Agency

100150

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/24/18	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 17-24-100 EtOH Certified Conc. 0.100</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 11/14/17 Exp Date: 11/14/18</p> <p>Opened: <u>STS 6/24/18</u></p>	<p>RPC Completed by Szarejko VSP</p>
6/25/18	<p><u>Re-do RPC</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>J. Szarejko VSP</p>
7/10/18	<p><u>Completed APM protocol.</u></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p>R Driscoll VFL</p>
8/13/18	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 18-16-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 10/29/18 Exp Date: 10/29/19</p> <p>Opened:</p>	<p>J. Szarejko VSP</p>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 18-16-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 10/29/18 Exp Date: 10/29/19</p> <p>Opened: <i>STS 2/28/19</i></p>	<p><i>RPC completed by Szarejko VSP</i></p>
<i>6/18/19</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/19 Exp Date: 02/28/20</p> <p style="text-align: center;"><i>6/18/19 MC</i></p> <p style="text-align: center;"><i>- RPL TEST -</i></p>	<p><i>RPC OFC MATT COLLINS BRISTOL PD</i></p>
<i>7/18/19</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>Completed APM protocol</i></p>	<p><i>J Dukette (UFL)</i></p>
<i>10/12/19</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>M Collins 10/12/19</i></p> <p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/19 Exp Date: 02/28/20</p> <p style="text-align: center;"><i>10/12/19</i></p> <p style="text-align: center;"><i>OCT RPL</i></p>	<p><i>M. Collins</i></p>

DATAMASTER DMT MAINTENANCE LOG

BRISTOL PD

100150

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/21	OCT RPC [For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 21-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: <u>10-31-21</u>	
3/31/22	Feb RPC [For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: <u>3-31-22 M108</u>	
6/21/22	Completed APM [For RPC or Simulator Solution Change Place Label Here]	J Durette (VFL)
8/11/22	RPC DMT Supervisor Training + RPC	 J Boldue VFL

DATAMASTER DMT MAINTENANCE LOG

Bristol Police Department

100150

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>10/26/22</i>	<p align="center">DMT External Std Solution</p> <p>[For RPC] Lot: 22-12-100 EtOH Certified Conc. 0.101 [ce Label Here]</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 5/31/22 Exp Date: 05/31/24</p> <p>Opened: <i>10/26/22 FIS</i></p> <p><i>RPC</i> <i>October Routine Performance Check</i> <i>↳ 2 tests. First one was completed im properly. FIS</i></p>	<i>Francis Sutek</i> <i>DMT supervisor</i>
<i>03/05/23</i>	<p align="center">DMT External Std Solution</p> <p>[For RP] Lot: 22-06-100 EtOH Certified Conc. 0.100 [ce Label Here]</p> <p>Range: 0.095 - 0.105</p> <p>Prep Date: 3/7/22 Exp Date: 03/07/24</p> <p>Opened: 03/05/23 <i>6/2</i></p> <p><i>RPC</i> <i>February Routine Performance Check</i></p>	<i>Francis Sutek</i>
<i>6/19/23</i>	<p align="center">DMT External Std Solution</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>DMT Removed for service</i></p>	<i>J Dukette</i> <i>(VFC)</i>
<i>7/11/23</i>	<p align="center">DMT External Std Solution</p> <p>Lot: 23-07-100 EtOH Certified Conc. 0.100</p> <p>Prep Date: 6/20/23 Exp Date: 06/20/25</p> <p>Opened: <i>install and APM</i></p>	<i>S. Bolla</i>
<i>9/1/23</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Fixed touch screen</i></p>	<i>S. Bolla</i>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/22/2023	<p style="text-align: center;">DMT External Std Solution</p> <p>[For RPI Lot: 23-02-100 EtOH Certified Conc. 0.101 Prep Date: 2/24/23 Exp Date: 02/24/25 Opened: _____]</p>	<p style="text-align: center;">DMT Supervisor</p>
10/25/23	<p>October RPC</p>	
10/25/23	<p style="text-align: center;">Removed for rehub</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p style="text-align: center;">A. Bolivar</p>
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	