

DATAMASTER DMT MAINTENANCE LOG

Newport Pd

Department/Agency

100151

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>10/25/11</i>	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>10/25/11</i> <i>KKK 10-25-11</i></p> <p style="text-align: center; font-size: 1.2em;"><i>Install Per Protocol</i></p>	<p><i>Steven Harnois</i></p> <p><i>Whiz Kill</i> Prog. Chief Organic Chemstrs.</p>
<i>11/17/11</i>	<p><i>Software update Installed</i> <i>KKK</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>Whiz Kill</i> Prog. Chief VDTL</p>
<i>12/06/2011</i>	<p>DataMaster Simulator Solution Lot # 12-13-100 EtOH_{aq} Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 09/09/2011 Expires: 09/09/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>12-6-2011</i></p>	<p style="font-size: 1.5em; text-align: center;"><i>Solution Change</i></p> <p><i>C. Moulton (NCS)</i></p>
<p style="font-size: 2em; opacity: 0.5;">X</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: right;"><i>Sent to S.D.S. 1-19-12</i></p>		
<i>1-19-12</i>	<p>DataMaster Simulator Solution Lot # 12-45-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>1-19-12</i></p>	<p><i>C. Moulton</i></p> <p><i>Feb RPC & Solution Change</i></p>

DATAMASTER DMT MAINTENANCE LOG

Newport Pd

100151

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>1/23/2012</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>Report of Vapor Recovery Found No Leak in The Simulator, Found Sim Temp 0.2°C Low Adjusted To 34.0°C</i></p> <p style="text-align: right;"><i>Rein Diag</i></p> <p><i>Officer Marcoux Witnessed and ATP</i></p>	<p><i>Steven Hawkins</i></p>
<i>4-16-12</i>	<p>DataMaster Simulator Solution Lot # 12-45-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 12/12/2011 Expires: 12/12/2012</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: <i>4-16-12</i></p>	<p><i>[Signature]</i> M1387</p>
<i>6-30-12</i> <i>Sent to SA - 6-30-12</i>	<p>DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013</p> <p>OPENED: <i>6-30-12</i></p>	<p><i>Chad Mout</i> M1387</p>
<i>8/27/12</i>	<p><i>APM Completed + Software updated</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i></p>
<i>10-31-12</i> <i>com</i> <i>Sent to SA 10-31-12</i>	<p>DataMaster Simulator Solution Lot # 12-72-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013</p> <p>OPENED:</p>	<p><i>[Signature]</i> M1386</p>
<i>2-28-13</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>DataMaster Simulator Solution Lot # 12-72-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/24/2012 Expires: 04/24/2013</p> <p>OPENED: <i>C. Moutan 2-28-13</i></p>	<p><i>RPC</i></p> <p><i>C. Moutan</i></p>

DATAMASTER DMT MAINTENANCE LOG

Newport City Police
Department/Agency

#100151
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<p>6/28/13 S.A 6-28-13</p>	<p>DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 [For RPC or Simulator Solution Change Place Label Here] OPENED: C. Moulton 6-28-13</p>	<p>C. Moulton</p>
<p>8/1/13</p>	<p>[For RPC or Simulator Solution Change Place Label Here] Completed APM Protocol</p>	<p>R. Driscoll DMT TECH.</p>
<p>10/29/13 A 10-26-13</p>	<p>October RPC DataMaster Simulator Solution Lot # 12-87-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/31/2012 Expires: 10/31/2013 OPENED: 10-29-13 [Signature]</p>	<p></p>
<p>In house 01/01/2014</p>	<p>[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 01/01/2014 [Signature]</p>	<p>R. Moulton 1/3/14</p>
<p>2-28-14</p>	<p>Feb RPC DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 2-28-14 [Signature]</p>	<p>C. Moulton 2/13/14 [Signature]</p>
<p>6-20-14</p>	<p>[For RPC or Simulator Solution Change Place Label Here] June RPC DataMaster Simulator Solution Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: 6-20-14 [Signature]</p>	<p>[Signature] Moulton</p>

DATAMASTER DMT MAINTENANCE LOG

Newport PD

Department/Agency

100151

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>8/7/14</i>	<p style="text-align: center;"><i>Completed APM Protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>Also adjusted clock ahead 13 minutes to correct time</i></p>	<i>R. Driscoll DMT TECH.</i>
<i>9-20-14</i>	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015</p> <p>[For RPC _____ e Label Here]</p> <p>OPENED: <i>9-20-14</i></p>	<i>Math</i>
<i>2-15-15</i>	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015</p> <p>[For RPC _____ ce Label Here]</p> <p>OPENED: <i>2-15-15 Math</i></p>	<i>Math</i>
<i>6-30-15</i>	<p style="text-align: center;">DataMaster Simulator Solution</p> <p>Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016</p> <p>[For RP _____ lace Label Here]</p> <p>OPENED: <i>6-30-15 Math</i></p>	<i>Math</i>

Sent to S.A. 9-20-14

Sent to S.A. 2-15-15

6-30-15

DATAMASTER DMT MAINTENANCE LOG

Newport PD

Department/Agency

100151

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-2-15	<p>DataMaster Simulator Solution Lot # 15-01-100 EtOHaq Conc: 0.098 g/210L Range: 0.093 - 0.103 g/210L Preparation Date: 1/21/2015 Expires: 1/21/2016</p> <p>[For RPC _____ e Label Here]</p> <p>OPENED: <u>C. Moulton 10-2-15</u></p>	<p><i>E. Moulton</i> at RPC</p>
11/5/15	<p>Completed APM protocol, Downloaded data [For RPC or Simulator Solution Change Place Label Here]</p> <p>to USB drive, uploaded new software, and installed new f/w PCB. No port available at this time. Will email DMT supervisor.</p>	<p><i>R. Driscoll</i> DMT Tech.</p>
2/16/16	<p>DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016</p> <p>[For RP _____ ce Label Here]</p> <p>OPENED: <u>2/16/16 RD</u></p> <p>Installation into new processing room. Site evaluation completed as well.</p>	<p><i>R. Driscoll</i> DMT Tech.</p>
4/21/16	<p>DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016</p> <p>[For RP _____ ace Label Here]</p> <p>OPENED: <u>4/21/16 Moulton</u></p> <p>Solution Change 4</p>	<p><i>E. Moulton</i></p>

DATAMASTER DMT MAINTENANCE LOG

Wauwatosa Police
Department/Agency

100151
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/1/16	<p>DataMaster External Standard Solution Lot # 16-05-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 4/19/2016 Expires: 4/19/2017</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 7/1/16 </p>	Sgt Moulton
10-28-16	<p>DMT External Standard Solution Lot # 16-21-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 9/13/2016 Expires: 9/13/2017</p> <p>OPENED: 10-28-16 </p>	Sgt Moulton
1/20/17	<p style="text-align: center;">Completed RPM Protocol</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	P. J. Dasek II DMT Tech.
2-28-17	<p>DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017</p> <p>RPC</p> <p>OPENED: 3-1-17 2-28-17 </p>	Sgt Moulton
4/26/17	<p>DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: 4/26/17 RO </p> <p style="text-align: center;">Install post simulator update simulator</p>	R. J. Dasek II VFL
6-28-17	<p>DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>OPENED: Moulton 6-28-17 </p>	
11-1-17	<p>DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018</p> <p>OPENED: Moulton 11/1/17 </p>	


DATAMASTER DMT MAINTENANCE LOG

Newport PD

Department/Agency

100151

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/2/17	<p><i>Installed version 2.00 RPC chip</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
<i>2-28-18</i>	<p>DMT Simulator Solution</p> <p>Lot # 17-19-100 EtOHaq Certified Conc: 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Preparation Date: 05/30/2017 Expires: 05/30/2018</p> <hr/> <p>OPENED: <i>Sgt Moulton 2-28-18</i></p>	
4/11/18	<p><i>Completed APM Protocol</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<p><i>R. Driscoll</i> VFL</p>
<i>6/29/18</i>	<p>DMT External Std Solution</p> <p>Lot: 18-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/18 Exp Date: 02/28/19</p> <hr/> <p>Opened: <i>Moulton 6-29-18</i></p> <p><i>True RPC</i></p>	
<i>10/30/18</i>	<p>DMT External Std Solution</p> <p>[For RPC or Simulator Solution Change Place Label Here]</p> <p>Lot: 18-04-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 2/28/18 Exp Date: 02/28/19</p> <hr/> <p>Opened: <i>Moulton 10/30/18</i></p>	
<i>2/15/19</i>	<p>DMT External Std Solution</p> <p>Lot: 18-19-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 12/19/18 Exp Date: 12/19/19</p> <hr/> <p>Opened: <i>Moulton 2/15/19</i></p>	
4/10/19	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p><i>APM COMPLETE</i></p>	<p><i>D. Driscoll</i> VFL</p>
<i>6/29/19</i>	<p>DMT External Std Solution</p> <p>Lot: 18-19-100 EtOH Certified Conc. 0.101</p> <p>Range: 0.096 - 0.106</p> <p>Prep Date: 12/19/18 Exp Date: 12/19/19</p> <hr/> <p>Opened: <i>Moulton 6/29/19</i></p>	

DATAMASTER DMT MAINTENANCE LOG

Nassau PD
Department/Agency

100151
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/31/19	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R] Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: <u> </u></p>	<p style="text-align: center;">[Place Label Here]</p> <p style="text-align: right;"><i>Sgt Moulton</i></p>
02/02/20	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 19-13-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: <u>Moulton 2/2/20</u></p>	<p style="text-align: right;"><i>Sgt Moulton</i></p>
6/30/20	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R] Lot: 19-13-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 9/23/19 Exp Date: 09/23/20 Opened: <u>6-30-2000 Moulton</u></p>	<p style="text-align: center;">[Place Label Here]</p> <p style="text-align: right;"><i>[Signature]</i></p>
10/13/20	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 20-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/21/20 Exp Date: 02/21/21 Opened: <u> </u></p>	<p style="text-align: right;"><i>[Signature]</i></p>
11/28/21	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R] Lot: 20-15-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: <u>1/28/21 SSI</u></p> <p style="text-align: center;"><i>Install post-TSI</i></p>	<p style="text-align: right;"><i>J. Dukette (VFL)</i></p>
4/19/2021	<p style="text-align: center;">DMT External Std Solution</p> <p>[For R] Lot: 21-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: <u>SSI 4/19/2021</u></p> <p style="text-align: center;"><i>Performed install and APM protocols.</i></p>	<p style="text-align: right;"><i>S. Stratton (VFL)</i></p>

DATAMASTER DMT MAINTENANCE LOG

Newport PD
Department/Agency

100151
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>6-30-2021</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 21-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: <i>6-30-2021</i> <i>month</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>9-30-2021</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 21-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: <i>month</i> <i>9/30/2021</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>1-14-2022</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 21-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: <i>month</i> <i>1/14/2022</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>2-21-2022</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 21-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: <i>month</i> <i>2-21-22</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>3/17/2022</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Completed <i>R 3/17/2022</i> <i>APM</i> protocol [For RPC or Simulator Solution Change Place Label Here]</p>	<i>S. Stratton (VFL)</i>
<i>6-30-22</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: <i>month</i> <i>6-30-22</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>10-26-22</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: <i>month</i> <i>10-26-22</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	
<i>2-28-23</i>	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-01-100 EtOH Certified Conc. 0.100 Prep Date: 1/25/23 Exp Date: 01/25/25 Opened: <i>2-28-23</i> <i>month</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	

DATAMASTER DMT MAINTENANCE LOG

Newport PD

100151

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/16/23	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
3/25/23	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-02-100 EtOH Certified Conc. 0.101</p> <p>Prep Date: 2/24/23 Exp Date: 02/24/25</p> <p>Opened: <u>Mark</u> <u>02-25-23</u></p> <p style="color: blue; font-size: 1.2em;">Solution Change</p>	
6-13-23	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-01-100 EtOH Certified Conc. 0.100</p> <p>Prep Date: 1/25/23 Exp Date: 01/25/25</p> <p>Opened: <u>6-13-23</u> <u>Mark</u></p> <p style="color: blue; font-size: 1.2em;">June RPC Change Place Label Here]</p>	
10-27-23	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-02-100 EtOH Certified Conc. 0.101</p> <p>Prep Date: 2/24/23 Exp Date: 02/24/25</p> <p>Opened: <u>10-27-23</u> <u>Mark</u></p> <p style="color: blue; font-size: 1.2em;">October RPC</p>	
1-4-24	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-08-100 EtOH Certified Conc. 0.099</p> <p>Prep Date: 7/19/23 Exp Date: 07/19/25</p> <p>Opened: <u>1-4-24</u> <u>Mark</u></p> <p style="color: blue; font-size: 1.2em;">Solution Change Change Place Label Here]</p>	
3-1-24	<p style="text-align: center;">DMT External Std Solution</p> <p>Lot: 23-08-100 EtOH Certified Conc. 0.099</p> <p>Prep Date: 7/19/23 Exp Date: 07/19/25</p> <p>Opened: <u>Mark</u> <u>3-1-24</u></p> <p style="color: blue; font-size: 1.2em;">Feb RPC</p>	
3/14/24	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
6-25-24	<p style="text-align: center;">DMT Quality Control Solution</p> <p>Lot: 24-02-100 EtOH Certified Conc. 0.099</p> <p>Prep Date: 4/9/24 Exp Date: 04/09/26</p> <p>Opened: <u>Mark</u></p>	

DATAMASTER DMT MAINTENANCE LOG

Newport PD

100151

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>9-26-24</i>	<p>DMT Quality Control Solution</p> <p>Lot: 24-02-100 EtOH Certified Conc. 0.099</p> <p>Prep Date: 4/9/24 Exp Date: 04/09/26</p> <p>Opened: <i>Moultan 9-26-24</i></p> <p style="text-align: right;"><i>October RPC</i> [For RPC or Simulator Solution Change Place Label Here]</p>	<i>[Signature]</i> <i>miller</i> <i>Nov 14</i>
<i>1-18-25</i>	<p>DMT Quality Control Solution</p> <p>Lot: 24-10-100 EtOH Certified Conc. 0.101</p> <p>Prep Date: 7/29/24 Exp Date: 07/29/26</p> <p>Opened: <i>Moultan 1-18-25</i></p> <p style="text-align: right;"><i>Feb RPC</i></p>	<i>[Signature]</i> <i>miller</i>
<i>3/16/25</i>	<p>[For RPC or Simulator Solution Change Place Label Here]</p> <p style="text-align: center;"><i>APM</i></p>	<i>A. Boldue</i> <i>VFL</i>
<i>5-2-25</i>	<p>DMT Quality Control Solution</p> <p>Lot: 24-10-100 EtOH Certified Conc. 0.101</p> <p>Prep Date: 7/29/24 Exp Date: /26</p> <p>Opened: <i>Moultan 5-2-25</i></p> <p style="text-align: right;"><i>Solution Change</i></p>	<i>[Signature]</i>
<i>6-27-25</i>	<p>DMT Quality Control Solution</p> <p>Lot: 25-11-100 EtOH Certified Conc. 0.102</p> <p>Prep Date: 4/7/25 Exp Date: 04/07/27</p> <p>Opened: <i>Moultan 6/27/25</i></p> <p>[For RPC or Simulator Solution Change Place Label Here]</p>	<i>[Signature]</i>
<i>8/1/25</i>	<p>removed for rehub</p>	<i>A. Boldue</i>
	<p>[For RPC or Simulator Solution Change Place Label Here]</p>	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	