

SOLUTION CHANGE

DataMaster DMT: 100151
Location: NEWPORT PD
Date: 04/21/2016
Performed by: CHARLES D MOULTON



Accuracy and Precision Check

Concentration = 0.097 g/210L
Lot # = 15-23-100
Range = 0.092 - 0.102
Average = 0.099 g/210L
Std Dev = 0.0005

RCD
4/22/16

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "RCD", is written over a horizontal line.

Date

04/21/2016