

SOLUTION CHANGE

DataMaster DMT: 100151
Location: NEWPORT PD
Date: 01/14/2022
Performed by: CHARLES D MOULTON



Accuracy and Precision Check

Concentration = 0.100 g/210L
Lot # = 21-11-100
Range = 0.095 - 0.105
Average = 0.097 g/210L
Std Dev = 0.0004

828 1/18/2022

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "CD Moulton", is written over a horizontal line.

Date

01/14/2022