

SOLUTION CHANGE

DataMaster DMT: 100151
Location: NEWPORT PD
Date: 01/04/2024
Performed by: CHARLES D MOULTON



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 23-08-100
Range = 0.094 - 0.104
Average = 0.098 g/210L
Std Dev = 0.0000

SDS 01/05/2024

Simulator Temperature: 34.0°C

Performed by *CDM*

Date 01/04/2024

SOLUTION CHANGE

DataMaster DMT: 100151
Location: NEWPORT PD
Date: 01/04/2024
Performed by: CHARLES D MOULTON



Accuracy and Precision Check

Concentration
Lot #
Range
Average
Std Dev
RF Detected

Simulator Temperature:

Solution Change Failed

Performed by

A handwritten signature in black ink, appearing to read "C. Moulton", is written over a horizontal line.

Date

01/04/2024