

SOLUTION CHANGE

DataMaster DMT: 100151

Location: BARRE CITY PD

Date: 12/14/2025

Performed by: MICHAEL R REALE



Accuracy and Precision Check

Concentration = 0.100 g/210L

Lot # = 25-21-100

Range = 0.095 - 0.105

Average = 0.099 g/210L

Std Dev = 0.0002

Simulator Temperature: 34.0°C

Brittany Sny 12/15/2025

Performed by

MB Hoy

Date

12/14/2025