

SOLUTION CHANGE

DataMaster DMT: 100151
Location: NEWPORT PD
Date: 05/02/2025
Performed by: CHARLES D MOULTON



Accuracy and Precision Check

Concentration = 0.101 g/210L
Lot # = 24-10-100
Range = 0.096 - 0.106
Average = 0.098 g/210L
Std Dev = 0.0004

JD 5/5/25

Simulator Temperature: 34.0°C

Performed by 

Date 05/02/2025