

SOLUTION CHANGE

DataMaster DMT: 100151

Location: BARRE CITY PD

Date: 5/28/2026

Performed by: MICHAEL R REALE



Accuracy and Precision Check

Concentration = 0.100 g/210L

Lot # = 26-08-100

Range = 0.095 - 0.105

Average = 0.100 g/210L

Std Dev = 0.0003

Simulator Temperature: 34.0°C

Performed by

A handwritten signature in black ink, appearing to read "MR", is written over a horizontal line.

Date

5/28/2026