

BAC DATAMASTER CHECK-UP and MAINTENANCE LOG

Woodsrock

Department/Agency

100152

DataMaster Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY
	[Place Simulator Solution Label Here] Old logbook went missing. Last entry was 7/18/22 APM. Only one RPC was completed before instrument was removed for construction.	
10/30/21	October [Place Simulator Solution Label Here] RPC	Jacob Holmes (DMT Supervisor)
	DMT External Std Solution Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 10/30/21 JH	Transcribed by J Dukette (VFL)
5/24/22	DMT External Std Solution Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: 850 5/24/22 Install post construction	J Dukette (VFL)
6/18/23	DMT External Std Solution Lot: 22-15-100 EtOH Certified Conc. 0.100 Prep Date: 10/26/22 Exp Date: 10/26/24 Opened: KM 6/18/23 June RPC	KIRSTEN MURPHY DMT Supervisor
7/3/23	[Place Simulator Solution Label Here]	J Dukette (VFL)

ALC 803 Rev. 2 06/07

photocopied and added as Everything for current datamaster
1
in one book after construction and installation in
current location

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/7/22	APM completed. New Maintenance Log now in use. [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
10/3/2022	DMT External Std Solution [For R Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 JH Exp Date: 04/01/24 Opened: RPC + Solution Change	J. Holmes
2/14/23	DMT External Std Solution [For R Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: JH 2/14/23 RPC + Solution Change	J. Holmes
[For RPC or Simulator Solution Change Place Label Here]		

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/02/23	DMT External Std Solution [For RPC Lot: 22-15-100 EtOH Certified Conc. 0.100 Label Here] Prep Date: 10/26/22 Exp Date: 10/26/24 Opened: KM 11/02/2023 RPC OCTOBER	KIRSTEN MURPHY DMT SUPERVISOR
11/03/23	[For RPC or Simulator Solution Change Place Label Here] Sarah Stratton from State Lab Request RPC test - redo to make sure everything is right - yesterday process started with low sim temp	KIRSTEN MURPHY DMT SUPERVISOR
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

