

DATAMASTER DMT MAINTENANCE LOG

Woodstock Pol

Department/Agency

100152

~~103709~~ 100152 9-22-11

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/22/11	<u>DataMaster Ext. Std. Solution</u> Lot # 11-41-100 EtOH _{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 [range Place Label Here] OPENED: 9/22/11 <u>KKK 9-22-11</u> <u>Instal Per Protocol</u>	Steven Harnois <u>W. L. Hill</u> Prog. Chief.
11/7/11	<u>Software Update Installed 1/2 KKK</u> [For RPC or Simulator Solution Change Place Label Here]	<u>W. L. Hill</u> Prog. Chief. Org Chem. VDH L
02/16/2012	[For RPC or Simulator Solution Change Place Label Here] <u>See "OLD" Book</u> <u>For Service Performed</u>	Steven Harnois
2/29/12	[For RPC or Simulator Solution Change Place Label Here] <u>Removed Instrument</u> <u>RFI Set Point Not</u> <u>Holding During Power</u> <u>On/Off</u>	Steven Harnois

DATAMASTER DMT MAINTENANCE LOG

Woodstock Pol
Department/Agency

100152
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/8/12	<u>DataMaster Simulator Solution</u> Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [Range Place Label Here] OPENED: 3/8/12 Installation of Instrument After Service at DPS Lab	Steven Hanson
6/2/12	<u>DataMaster Simulator Solution</u> Lot # 12-73-100 EtOH_{aq} Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013 [Range Place Label Here] OPENED: 6-2-2012 Routine Performance Check	William Hanson
8/13/12	See old Maint. log ²⁰³ 8/17/12 [For RPC or Simulator Solution Change Place Label Here] Sim Solution Change <u>DataMaster Simulator Solution</u> Lot # 12-73-100 EtOH_{aq} Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 04/27/2012 Expires: 04/27/2013 OPENED: 8-13-12	W. Hansen
8/17/12	Sim Solution Protocol and APM completed [For RPC or Simulator Solution Change Place Label Here]	R. Discoll

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-10-2012	<u>DataMaster Simulator Solution</u> Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 OPENED:	William Hansen (Supv)
02-04-2013	<u>DataMaster Simulator Solution</u> Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: 2/4/13	W. Hansen
2/7/13	RPC Failing for DMT Supervisor. (Twice for Sample acceptance, once for A+P check fail). Checked Simulator over, and reran RPC. Instrument working to spec.	R. Driscoll
	<u>DataMaster Simulator Solution</u> Lot # 12-90-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/18/2012 Expires: 12/18/2013 OPENED: 6-22-13	W. Hansen
6-22-13	Passed	W. Hansen
7/18/13	Completed APM [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH.

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9-12-2013	[For] <u>DataMaster Simulator Solution</u> Place Label Here] Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 OPENED: Solution Change	W. Hanson
10/09/13	<u>DataMaster Simulator Solution</u> [For RPC] Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 OPENED: 10/09/13 RPC SOLUTION CHANGE	J. OTRANTO
02/09/14	<u>DataMaster Simulator Solution</u> [For RPC] Lot # 13-13-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/24/2013 Expires: 06/24/2014 OPENED: 02/09/14 RPC SOLUTION CHANGE	J. OTRANTO
05/30/14	<u>DataMaster Simulator Solution</u> [For RPC] Lot # 13-21-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: 05/30/14 RPC SOLUTION CHANGE	J. OTRANTO

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD
Department/Agency

100152
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/21/14	<u>Completed: APM Protocol</u> [For RPC or Simulator Solution Change Place Label Here] <u>PER STATE MANDATE</u>	R. Driscoll DMT TECH.
10/8/14	<u>DataMaster Simulator Solution</u> [For RPC Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 3/12/2014 Expires: 3/12/2015 OPENED: <u>10/08/14</u> <u>RPC</u> <u>SOLUTION CHANGES</u>	J. OTTAUTO
10/29/14	<u>DataMaster Simulator Solution</u> [For RPC Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: <u>10/29/14 KED</u> <u>Installation per TSI</u>	R. Driscoll DMT TECH.
02/02/15	[For RPC or Simulator Solution Change Place Label Here] <u>DataMaster Simulator Solution</u> Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: <u>02/02/15</u> <u>RPC +</u> <u>SOLUTION</u> <u>CHANGES</u>	J. OTTAUTO DMT ADMIN

DATAMASTER DMT MAINTENANCE LOG

WOODSTOCK PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/14/15	[For RPC or Simulator Solution Change Place Label Here] RPC + SOLUTION CHANGE DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 06/14/15	J. OTTAVIO
8/18/15	Downloaded data to USB drive, uploaded new [For RPC or Simulator Solution Change Place Label Here] software, and installed communication PCB.	R. Driscoll DMT Tech.
10/25/15	RPC & SOLUTION CHANGE [For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 15-07-100 EtOHaq Conc: 0.100g/210L Range: 0.095 - 0.105 Preparation Date: 5/11/2015 Expires: 5/11/2016 OPENED: 10/25/15	J. OTTAVIO D. M. SUPERVISOR
11/18/15	Completed APM protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.

DATAMASTER DMT MAINTENANCE LOG

WOODSTOCK PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/22/16	[For RPC or Simulator Solution Change Place Label Here] RPC (FEB) DataMaster Simulator Solution Lot # 15-07-100 EtOHaq Conc: 0.100g/210L Range: 0.095 - 0.105 Preparation Date: 5/11/2015 Expires: 5/11/2016 OPENED: 2/22/16	J. OTTAVIO DM SUPERVISOR
05/15/16	[For RPC or Simulator Solution Change Place Label Here] RPC (EXPIRED SOLUTION) DataMaster External Standard Solution Lot # 16-03-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 2/2/2016 Expires: 2/2/2017 OPENED: 05/15/2016	
10/16/16	[For RPC or Simulator Solution Change Place Label Here] RPC (OCT) DataMaster External Standard Solution Lot # 16-03-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 2/2/2016 Expires: 2/2/2017 OPENED: 10/16/16	J. OTTAVIO DM SUPERVISOR
10/27/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DM Tech

DATAMASTER DMT MAINTENANCE LOG

WOODSTOCK PD
Department/Agency

100152
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/27/17	[For RPC or Simulator Solution Change Place Label Here] FEB 2017 DMT External Standard Solution Lot # 16-23-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/19/2016 Expires: 10/19/2017 OPENED: 02/27/17	J. OTTAVANTO DM SUPERVISOR
4/11/17	DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: 4/11/17 RED [For F Place Label Here] Install post simulator update	R. Driscoll VFL
6/13/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 6/13/17 [For RP ce Label Here] RPC & SOLUTION CHANGES	J. OTTAVANTO
10/16/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 10/16/17 [For RPC ce Label Here] RPC & SOLUTION CHANGES	J. OTTAVANTO

DATAMASTER DMT MAINTENANCE LOG

Woodstock PD

Department/Agency

100152

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/23/17	<u>Installed 2.00 PIC chip and Completed APM</u> [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
2/17/18	DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 [Fo e Place Label Here] OPENED: <u>2/17/18 JH</u> <u>RPC & Solution Change</u>	J. Holmes
6-29-18	DMT Simulator Solution Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 [Fo e Place Label Here] OPENED: <u>6-29-18 JH</u> <u>RPC & Solution Change</u>	J. Holmes
7/19/18	<u>Complete APM Protocol</u> [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-26-18	DMT External Std Solution [For RPC Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: 10-26-18 RPC & Solution Change	J. Holmes
2-27-19	DMT External Std Solution [For Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: 2-27-19 RPC & Solution Change	J. Holmes
6/24/19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 6/24/19 RPC & Solution Change	J. Holmes
7/15/19	Completed APM procedure [For RPC or Simulator Solution Change Place Label Here]	J Oukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number - _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/24/19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 10/24/19	J. Holmes
	Change Place Label Here]	
	RPC & Solution Change	
1/16/20	DMT External Std Solution Lot: 19-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/17/19 Exp Date: 06/17/20 Opened: 1/16/20	J. Holmes
	Change Place Label Here]	
	Put out of service - Changed Solution Contacted J. Dukette via phone/email RPC Check 1/26/2020 - Solution had changed on 1/16/20 (Back in Service) 1/26/20	
4/29/2020	DMT External Std Solution Lot: 20-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/21/20 Exp Date: 02/21/21 Opened: 4/29/2020	J. Holmes
	Change Place Label Here]	
	DMT External Std Solution Solution Change - Sample DUI Good	
6/29/2020	DMT External Std Solution [For Lot: 20-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/21/20 Exp Date: 02/21/21 Opened: 6/29/2020 JH	J. Holmes
	e Place Label Here]	
	RPC & Solution Change	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/9/2020	<i>APM protocol completed</i> [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
7/15/2020	<i>Installed DMT software update</i> [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
9/30/2020	<i>RPC & Solution Change</i> [DMT External Std Solution inge Place Label Here] Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: <i>9-30-2020 JH</i>	J. Holmes
12/21/2020	[DMT External Std Solution je Place Label Here] Lot: 20-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/25/20 Exp Date: 06/25/22 Opened: <i>12-21-2020 JH</i> <i>Started on 12/21/2020 ended on 12/22/2020</i>	J. Holmes

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/22/2021	DMT External Std Solution Lot: 20-15-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 [Place Label Here] Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 2/22/2021 JH RPC & Solution Change	J. Holmes
5-30-21	DMT External Std Solution Lot: 20-15-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 [Place Label Here] Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 5/30/21 JH RPC & Solution Change	J. Holmes
7/8/2021	Completed APM protocol. Removed operator log as no longer needed. Adjusted RP detection. [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
	[For RPC or Simulator Solution Change Place Label Here]	