

DATAMASTER DMT MAINTENANCE LOG

Brattboro Rd

100153

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/2/11	DataMaster Ext. Std. Solution Lot # 11-41-100 EtOH _{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 06/29/2011 Expires: 06/29/2012 OPENED: 9/2/11 RPK Change Place Label Here]	Steven Harnois RPK/CLP Prog. Chief. VIDHL
	Instal Per Procedure	
	[For RPC or Simulator Solution Change Place Label Here]	
9/28/11	updated Software	A. Bole Bole 11234
10/28/11	DataMaster Ext. Std. Solution Lot # 11-40-100 EtOH _{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/15/2011 Expires: 06/15/2012 10/28/11 RPK [For Place Label Here]	Kirkpatrick P0809
	TEST/SIMULATOR SOLUTION CHANGE TEST/RPC (Routine Performance Check)	
11/2/11	Check in confirmation test [For RPC or Simulator Solution Change Place Label Here]	Kirkpatrick P0809
11/7/11	Software update installed by KLV	KLV 11/7/11 Prog. Chief

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/14/11	[For RPC or Simulator Solution Change Place Label Here] Corrected time for PST/EST.	Mh EJ Sgt. Mark Carignan
2/15/12	DataMaster Ext. Std. Solution Lot # 11-40-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 06/15/2011 Expires: 06/15/2012 [For RP ice Label Here] Routine Performance Check (RPC) Sent to SA 2/15/12 MFC	Mh EJ Sgt. Mark Carignan
2/18/12	Screen Frozen - Turned off for 1 min [For RPC or Simulator Solution Change Place Label Here] and turned back on - Successful Restart	P. Witherbee Sgt. Witherbee
4/25/12	[For RPC or Simulator Solution Change Place Label Here] Corrected time for DST/EST Sent to SA 5/31/12 MFC	Mh EJ Sgt. Mark Carignan

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/3/12	DataMaster Simulator Solution Lot # 12-32-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 OPENED: 6/3/12 <u>MSC</u> [For RPC e Label Here]	<i>MSC</i> Sgt. Mark Carignan
Sent to SA 6/3/12		
8/10/12	[For RPC or Simulator Solution Change Place Label Here] APM Performed	<i>R. Driscoll</i>
Sent to SA 8/10/12 <u>MSC</u>		
9/30/12	DataMaster Simulator Solution Lot # 12-70-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: 9/30/12 <u>MSC</u> [For RPC o .apel Here] DMT reported "Standard out of Range" at 1:40am on 9/30/12. Solution change performed.	<i>MSC</i> Sgt. Mark Carignan
<u>MSC</u> Sent to SA 9/30/12		
10/25/12	DataMaster Simulator Solution Lot # 12-70-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: 10/25/12 <u>MSC</u> [For RPC e Label Here] RPC done <u>MSC</u>	<i>MSC</i> Sgt. Mark Carignan
<u>MSC</u> Sent to SA 10/25/12		

DATAMASTER DMT MAINTENANCE LOG

Brentleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/10/13 ↑ Sent to SA 1/10/13	[For RPC or Simulator Solution Change Place Label Here] Time corrected for EST.	Mh S Sgt. Mark Corrigan
2/3/13 ↑ Sent to SA 2/3/13 / Resent 3/11/13	[For RPC (DataMaster Simulator Solution Lot # 12-80-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/5/2012 Expires: 7/5/2013 Label Here] OPENED: 2/3/13 RPC completed	Mh S Sgt. Mark Corrigan
3/14/13 ↑ Sent to SA 3/14/13	[For RPC or Simulator Solution Change Place Label Here] Time corrected for DST	Mh S Sgt. Mark Corrigan
6/17/13 ↑ Sent to SA 6/17/13	[For RPC (DataMaster Simulator Solution Lot # 12-84-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 9/24/2012 Expires: 9/24/2013 Label Here] OPENED: 6/17/13 RPC completed	Mh S Sgt. Mark Corrigan

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/18/13	Completed Annual PM [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
But to SA 7/18/13	MFC	
	OPENED: DataMaster Simulator Solution Lot # 12-90-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 12/18/2012 Expires: 12/18/2013 [For RPC or Simulator Solution Change Place Label Here] 10/11/13 MFC RPC/Sim Change attempted - touch screen failure. VFL Rob Driscoll Advised. Machine placed OOS - 10/11/13 MFC	10/11/13 MFC
10/11/13	[For RPC or Simulator Solution Change Place Label Here] DMT Touch screen unresponsive. Found faulty ribbon cable, so replaced with spare. DataMaster needs RPC completed will inform DMT Supervisor.	R. Driscoll DMT TECH.
	[For RPC or Simulator Solution Change Place Label Here] RPC completed 10/11/13 Both printer inks replaced	
Sent to SA 10/11/13		

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/10/13	[For RPC or Simulator Solution Change Place Label Here] Time changed for Standard time.	Carignan MHC
12/21/13	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-15-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 08/9/2013 Expires: 08/9/2014 OPENED: 12/21/13 <u>MHC</u> Expired Simulator solution changed.	Carignan MHC
↑ Sent to SA 12/31/13 <u>MHC</u>	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: <u>MHC</u> 2/10/14	Carignan MHC
↑ Sent to SA 2/10/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: <u>MHC</u> 2/10/14	Carignan MHC
↑ Sent to SA 5/29/14	[For RPC or Simulator Solution Change Place Label Here] Pump error encountered by officer. VFL Rob Driscoll Advised and said machine is okay. 5/29/14	Carignan MHC

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	DataMaster Simulator Solution Lot # 13-19-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/25/2013 Expires: 10/25/2014 OPENED: <u>6/20/14</u> <u>MR</u> [For RPC or Simulator Solution Change Place Label Here]	Sgt. Mark Carignan
↑ 6/20/14 Sent to SA		
	APM completed by Rob Driscoll 7/21/14 [For RPC or Simulator Solution Change Place Label Here]	
	DataMaster Simulator Solution Lot # 14-04-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 5/27/2014 Expires: 5/27/2015 OPENED: <u>10/31/14</u> <u>CZ</u> <u>RP</u> [For RPC or Simulator Solution Change Place Label Here]	LT Robert Perkins
10/31/14 SENT TO SA	RPC - ROBERT PERKINS 10/31/14	
11/4/14	[For RPC or Simulator Solution Change Place Label Here] CORRECTED TIME FOR DST/EST	LT Robert Perkins

DATAMASTER DMT MAINTENANCE LOG

Brattleboro Police

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/23/15	Screen frozen - turned power off [For RPC or Simulator Solution Change Place Label Here] and restarted	P. Withube M1239
2-1-15	DataMaster Simulator Solution [For RPC Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 2-1-15 AB M1219 RPC - Adam Belville 2-1-15	SGT. Adam Belville AB Belville M1219
Sent to SA 2-2-15		
3-11-15	Adjusted time for Daylight Savings [For RPC or Simulator Solution Change Place Label Here]	SGT Adam Belville AB Belville M1219
6-2-15	DataMaster Simulator Solution [For RPC Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 6-1-15 AB M1219 RPC - Adam Belville 6-2-15	SGT Adam Belville AB Belville M1219

DATAMASTER DMT MAINTENANCE LOG

Bathelme PD

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/22/15	DataMaster Simulator Solution Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 OPENED: 7/22/15 RCD [For RPC e Label Here]	R. Driscoll DMT Tech.
	Installation post TSI	
10-02-15	DataMaster Simulator Solution Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016 OPENED: 10-02-2015 AB M119 [For RPC e Label Here]	Sgt. Adam Belville
Sent to SA 10-02-15	RPC-Adam Belville 10-02-15	AB M119
10/23/15	DataMaster Simulator Solution Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: 10/23/15 RCD [For RPC e Label Here]	R. Driscoll DMT Tech.
	Install post TSI dated 10/22/15	
11-1-15	Adjusted for Daylight Savings [For RPC or Simulator Solution Change Place Label Here]	Sgt. Adam Belville
		AB M119

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
<i>2-1-16</i>	<i>Re-started DataMaster & Printer</i>	<i>Sgt. Adam Belville Mdlg</i>
	[For RPC or Simulator Solution Change Place Label Here]	
	<i>DataMaster wasn't communicating with</i>	
	<i>Printer</i>	
	<i>2-1-16 @ 21:44 hrs</i>	
<i>2-1-16</i>	[For RPC or Simulator Solution Change Place Label Here]	<i>Sgt. Adam Belville Mdlg</i>
	<i>RPC</i>	
	DataMaster Simulator Solution	
	Lot # 15-06-100 EtOHaq Conc: 0.100 g/210L Range: 0.095 - 0.105 Preparation Date: 4/28/2015 Expires: 4/28/2016	
	OPENED: <i>2-1-16 AB Mdlg</i>	
<i>5-19-16</i>	<i>Re-started DataMaster.</i>	<i>Sgt. Adam Belville Mdlg</i>
	[For RPC or Simulator Solution Change Place Label Here]	
	<i>DataMaster wasn't communicating</i>	
	<i>with the printer.</i>	
	<i>5-19-16 @ 22:01 hrs</i>	
<i>5-20-16</i>	<i>Re-started DataMaster</i>	<i>Sgt. Adam Belville Mdlg</i>
	[For RPC or Simulator Solution Change Place Label Here]	
	<i>"out of range"</i>	
	<i>Timer restarted for DUI then remained</i>	
	<i>"busy" for an extended period of</i>	
	<i>time.</i>	
	<i>5-20-16 @ 03:30 hrs</i>	

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6-3-16	[For RPC or Simulator Solution Change Place Label Here] RPC DataMaster External Standard Solution Lot # 15-25-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/30/2015 Expires: 11/30/2016 OPENED: 6-3-16 AB mjd19	SGT. Adam Belville mjd19
8/2/16	Completed APN protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
8-12-16	Ran Accuracy : Precision Report [For RPC or Simulator Solution Change Place Label Here] DMT stated Solution out of Range During DUT Processing	SGT. Adam Belville mjd19
8/22/16	DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 8/22/16 RCD [For RPC or S el Here] changed solution	R. Driscoll DMT Tech

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD

100153

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10-01-16	[For RPC or Simulator Solution Change Place Label Here] RPC	Sgt. Adam Belville M1219
12/16/16	OUT OF SERVICE "SIMULATOR OUT OF RANGE". [For RPC or Simulator Solution Change Place Label Here]	LT Robert Perkins M1235
12/26/16	[For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: 12/26/16 Ref R. SIMULATOR SOLUTION CHANGE	LT Robert Perkins M1235
12/27/16	CONDUCTED 2ND SOLUTION CHANGE PROTOCOL [For RPC or Simulator Solution Change Place Label Here] NOT INCORRECTLY ENTERED 0.095 FOR STANDARD. CONDUCTED PROTOCOL WITH 0.100 FOR STANDARD.	LT Robert Perkins

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
02/01/17	RPC [For RPC or Simulator Solution Change Place Label Here] DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: 02/01/17	SGT. Adam Belville M1219
3/16/17	RPC [For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED: 3/16/17	LT R. PERKINS M1235
5/4/17	[For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-16-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 04/03/2017 Expires: 04/03/2018 OPENED: 5/4/17 INSTALL AFTER DMT UPGRADE	D. PATLAK VFL
7/3/17	RPC [For RPC or Simulator Solution Change Place Label Here] DMT Simulator Solution Lot # 17-07-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 01/18/2017 Expires: 01/18/2018 OPENED:	LT R. PERKINS M1235

DATAMASTER DMT MAINTENANCE LOG

Brattleboro PD

Department/Agency

100153

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/26/17	DMT Simulator Solution Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 OPENED: 9/26/17 RCD [For RPC or Simulator Solution Change Place Label Here] Install in new building. Replaced PIC chip as well	R. Driscoll VFL
10/31/17	DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 10/31/17 PAW [For RPC or Simulator Solution Change Place Label Here]	P. Wetherbee M1239
11/11/17 11/21/17	RPC 2ND RPC COMPLETED DUE TO FAILED RP/SAMPLE ACCEPT TEST. OPERATOR ERROR - M1235	PAW M1235
2/18/18	DMT Simulator Solution Lot # 17-19-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 05/30/2017 Expires: 05/30/2018 OPENED: 2/18/18 TP [For RPC or Simulator Solution Change Place Label Here]	PAW M1235
5/1/18	Completed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/28/18	<u>RPC COMPLETED</u> [For RPC or Simulator Solution Change Place Label Here] <u>DMT Simulator Solution</u> Lot # 17-20-100 EtOHaq Certified Conc: 0.102 Range: 0.097 - 0.107 Preparation Date: 07/07/2017 Expires: 07/07/2018 OPENED: 6/28/18 T2P	T2P R. LT M1235
9/26/18	<u>DIAGNOSTIC CHECK COMPLETE (PASS)</u> <u>Accuracy & Precision Check Complete (Pass)</u> [For RPC or Simulator Solution Change Place Label Here] * REPORTED SIM TEMP OUT OF RANGE ON 9/25/18.	T2P R. LT M1235
10/3/18	[For RPC <u>DMT External Std Solution</u> ce Label Here] Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: 10/3/18 T2P * PERFORMED SULK BACK TEST INCORRECTLY. RESTART RPC FROM ACCURACY & PRECISION TEST. (T2P)	T2P R. LT M1235
01/13/19	<u>DMT External Std Solution</u> [For RPC ce Label Here] Lot: 18-11-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/26/18 Exp Date: 06/26/19 Opened: _____ Changed sim solution Reported Sim temperature out of range on 01/13/19	[Signature] LT. M1268





DATAMASTER DMT MAINTENANCE LOG

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/15/19	[For RPC or Simulator Solution Change Place Label Here] TRPC COMPLETED DUE TO SOLUTION CHANGE ON 1/13/19 (SEE PREVIOUS PAGE)	TRPC MIZST LT
03/12/19	[For RPC or Simulator Solution Change Place Label Here] RPC / Solution Change completed DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 03/12/19 AP	[Signature] Fenock MIZ68
3/19/19	[For RPC or Simulator Solution Change Place Label Here] APM completed	[Signature] D. PARM VPL
06/30/19	DMT External Std Solution Lot: 18-19-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 12/19/18 Exp Date: 12/19/19 Opened: 06/30/19 AP [For RPC or Simulator Solution Change Place Label Here] RPC Completed w/ Solution Change -suck back error generated	[Signature] Petlock MIZ68

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/7/19	DMT External Std Solution [For RPC Lot: 19-04-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: RPC / SOLUTION CHANGE COMPLETED	 M1235
10/12/19	SIM TEMP OUT OF RANGE ERROR. / PRINT [For RPC or Simulator Solution Change Place Label Here] ERROR. RESET POWER, OP V OK CALLED NDH TO REPORT.	 M1235
1/7/20	SIM SOLUTION CHANGE (SIM OUT OF RANGE) [For --- Place Label Here] DMT External Std Solution Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened:	 M1235 1/7/20
2/3/20	DMT External Std Solution [For RF Lot: 19-16-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 2/3/20 TAP	 M1235 2/3/20

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/10/2020	APM protocol completed [For RPC or Simulator Solution Change Place Label Here]	S Stratton (VFL) SOS
4/27/20	ACCURACY & PRECISION V. LAST TEST INDICATED SIM OUT OF RANGE ON 4/25/20 AVG 0.096	RLR M1235
5/15/20	DMT External Std Solution [For RPC Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 5/13/20 (RP)	SIM CHANGE SIM OUT OF RANGE SOLUBLE TIMES RLR M1235 5/13/20
6/1/20	DMT External Std Solution [For R Lot: 19-16-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 11/14/19 Exp Date: 11/14/20 Opened: 6/1/20 (RP)	place Label Here] RLR M1235
8/21/20	DMT External Std Solution [For RF Lot: 20-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/21/20 Exp Date: 02/21/21 Opened: 8/21/20 (RP)	SIM OUT OF RANGE place Label Here] RLR M1235

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/25/21	DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 [Range Place Label Here] Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 10/14/20 <i>70 600</i> <i>RPC w/ Solution change</i>	<i>[Signature]</i> PERLOCK MIZG8
5/25/21	SIM OUT OF RANGE 1/10/21 SIM CHANGE NEEDED 094 [For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 1/10/21 <i>PERKINS</i>	<i>[Signature]</i> PERKINS MIZ35 1/10/21
1/21/21	FEB 2021 RPC [For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 2/5/21 <i>PERKINS MIZ35</i>	<i>[Signature]</i> PERKINS MIZ35 2/5/21
3/5/21	Completed ARM protocol [For RPC or Simulator Solution Change Place Label Here]	<i>[Signature]</i> DUKELLE (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
5/4/21	<u>SIM OUT OF RANGE / SIM SOLUTION CHANGE</u> [For RF DMT External Std Solution Place Label Here] Lot: 21-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: 5/4/21 <u>Thel R. M1235</u>	<u>Thel R.</u> <u>LT/M1235</u>
6/3/21	<u>E JUNE RPC</u> [For RF DMT External Std Solution Place Label Here] Lot: 21-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 2/12/21 Exp Date: 02/12/23 Opened: 6/3/21 <u>Thel R. M1235</u>	<u>Thel R.</u> <u>LT/M1235</u>
8/26/21	<u>SIM SOLUTION REPORTED OUT OF RANGE (8/25)</u> [For RPC or Simulator Solution Change Place Label Here] <u>Accuracy & Precision Check Completed</u> <u>SIM OUT OF RANGE / DMT OUT OF SERVICE.</u>	<u>Thel R.</u> <u>LT/M1235</u>
9/2/21	<u>SIM SOLUTION CHANGE</u> [For RPC o DMT External Std Solution Label Here] Lot: 21-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: 9/2/21 <u>Thel R. M1235</u>	<u>Thel R.</u> <u>LT/M1235</u>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/21/21	October RPC DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 21-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/16/21 Exp Date: 07/16/23 Opened: <i>MC</i> 10/21/21	<i>MC</i> SGT M1221
02/26/22	Feb. RPC [For RPC or Simulator Solution Change Place Label Here] (no label) Prep date: 07/16/21 Exp date: 07/16/23	<i>AN</i> M1268/CT
3/2/22	Re-did testing portion of RPC to correct sample acceptance and lot number. [For RPC or Simulator Solution Change Place Label Here] Completed APM protocol	S. Stratton (VFL)
06/10/22	JUNE RPC [For R DMT External Std Solution [For RPC or Simulator Solution Change Place Label Here] Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: <i>AN</i> 06/06/22	<i>AN</i> M1268

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/09/22	Opened: _____ [For RP _____ DMT External Std Solution _____ Place Label Here] Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: _____ OCT RPC Ⓢ REPAIR FAILED TO PRINT	Ⓢ M1268
02/28/23	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 22-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/7/22 Exp Date: 03/07/24 Opened: _____	Ⓢ M1268
3/1/23	Completed APM [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)
6/2/23	DMT External Std Solution [For RP _____ Lot: 23-01-100 EtOH Certified Conc. 0.100 _____ Place Label Here] Prep Date: 1/25/23 Exp Date: 01/25/25 Opened: 6/2/23 MB Simulator Solution change Ⓢ RPC	MB M1230a ET

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/30/23	[For DMT External Std Solution Change Place Label Here] Lot: 23-07-100 EtOH Certified Conc. 0.100 Prep Date: 6/20/23 Exp Date: 06/20/25 Opened: SSC	MB M1230a
10/13/23	[For DMT External Std Solution ge Place Label Here] Lot: 23-07-100 EtOH Certified Conc. 0.100 Prep Date: 6/20/23 Exp Date: 06/20/25 Opened: RPC	MB M1230a
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	