

12/8/11 - 10/11/23 100154 Washington CSO Department/Agency DataMaster DMT Serial Number DATAMASTER DMT MAINTENANCE LOG

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/8/11	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [Place Label Here] OPENED: 12/8/11 <i>KZK</i> Install Per Procedure	Steven Harris <i>KZK</i> VDHL Prog. Chief
2/16/12	DataMaster Simulator Solution Lot # 12-38-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [Place Label Here] OPENED: 2-16-12	Bruce McClure
2/21/12	RPC .099	M. Poulton
5/30/12	DataMaster Simulator Solution Lot # 12-58-100 EtOH_{aq} Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [Change Place Label Here] OPENED: 5.30.2012	Bart Meyer
10/22/12	DataMaster Simulator Solution Lot # 12-64-100 EtOH_{aq} Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/17/2012 Expires: 02/17/2013 [Change Place Label Here] OPENED: BSM 10.22.2012	Bart Meyer

85/11/01-11/8/11

DATAMASTER DMT MAINTENANCE LOG

Washington County SD
Department/Agency100154
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/15/12	[For RPC or Simulator Solution Change Place Label Here] APM protocol performed	R. Driscoll DMT TECH.
2/5/13	[For <u>DataMaster Simulator Solution</u> Lot # 12-70-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 04/17/2012 Expires: 04/17/2013 OPENED: M. P. P. 2/5/13 0205 RPC .099	M. P. P.
6/14/13	[For <u>DataMaster Simulator Solution</u> Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: M. P. P. 6/14/13 0255 RPC .099	M. P. P.
10/24/13	[For <u>DataMaster Simulator Solution</u> Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: 10/24/13 M.P.P. RPC .099	M. P. P.

DATAMASTER DMT MAINTENANCE LOG

Washington County SD
Department/Agency

100154
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/13/13	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT TECH
11/21/13	DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 11/21/13 M.P. [For RPC or Simulator Solution Change Place Label Here]	M. Paulin
	Solution Change	
1/28/14	DataMaster Simulator Solution Lot # 13-16-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: 1/28/14 M.P. [For RPC or Simulator Solution Change Place Label Here]	M. Paulin
	R.P.C	
4/17/14	DataMaster Simulator Solution Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 3/12/2014 Expires: 3/12/2015 OPENED: 4/17/14 [For RPC or Simulator Solution Change Place Label Here]	M. Paulin
	Solution change	

DATAMASTER DMT MAINTENANCE LOG

⑨

Washington County SD

Department/Agency

100154

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/11/14	DataMaster Simulator Solution Lot # 14-02-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 3/12/2014 Expires: 3/12/2015 OPENED: <u>6/11/14 M.P.</u> <u>RPC .100</u>	Place Label Here] <u>M. Paulin</u>
10/2/14	DataMaster Simulator Solution Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: <u>10/2/14 M.P.</u>	Place Label Here] <u>M. Paulin</u>
11/12/14	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	<u>R. Driscoll</u> <u>DAT TECH.</u>
12/24/14	[For RPC or Simulator Solution Change Place Label Here] DataMaster Simulator Solution Lot # 14-10-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 7/14/2014 Expires: 7/14/2015 OPENED: <u>12/24/14 M.P.</u>	<u>M. Paulin</u> <u>.099</u>

DATAMASTER DMT MAINTENANCE LOG

WACSD

Department/Agency

100154

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/6/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 2/6/15 M.P. [For RPC or Simulator Solution Change Place Label Here] Label Here] 096 RPC	M. Poulin
5/11/15	DataMaster Simulator Solution Lot # 14-15-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 Preparation Date: 12/9/2014 Expires: 12/9/2015 OPENED: 5/21/15 M.P. [For RPC or Simulator Solution Change Place Label Here] Label Here] 099 Solution change	M. Poulin
6/23/15	DataMaster Simulator Solution Lot # 15-07-100 EtOHaq Conc: 0.100g/210L Range: 0.095 - 0.105 Preparation Date: 5/11/2015 Expires: 5/11/2016 OPENED: 6/23/15 M.P. [For RPC or Simulator Solution Change Place Label Here] Label Here] 100 RPC	M. Poulin
9/20/15	Downloaded data to usb drive, uploaded new software, [For RPC or Simulator Solution Change Place Label Here] and installed new i/o PCB. Detector voltage > 0.600 v after upgrade so taking back to VFL for detector optimization.	R. Discoll DMT Tech

DATAMASTER DMT MAINTENANCE LOG

Washington CSO

(Department/Agency)

100154

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/29/15	DataMaster Simulator Solution Lot # 15-12-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 07/28/2015 Expires: 07/28/2016 OPENED: 9/29/15 RCD [For RPC of _____ Label Here] Installation post test dated 9/22/15	R. Discoll DMT Tech
10/29/15	DataMaster Simulator Solution Lot # 15-07-100 EtOHaq Conc: 0.100g/210L Range: 0.095 - 0.105 Preparation Date: 5/11/2015 Expires: 5/11/2016 OPENED: 10/29/15 [For _____ Place Label Here]	M. Poulin
1/1/16	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 1/1/16 [For _____ Place Label Here] Solution change .099	M. Poulin
2/11/16	DataMaster External Standard Solution Lot # 15-23-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 10/29/2015 Expires: 10/29/2016 OPENED: 2/11/16 [For RPC _____ Place Label Here] RPC .099	M. Poulin

DATAMASTER DMT MAINTENANCE LOG

W. CSD

Department/Agency

100157

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/15/16	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 OPENED: 6/15/16 [For RPC or Simulator Solution Change Place Label Here]	M. Polin
8/25/16	DataMaster External Standard Solution Lot # 16-04-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 3/25/2016 Expires: 3/25/2017 OPENED: 8/25/16 [For RPC or Simulator Solution Change Place Label Here]	M. Polin
9/21/16	Completed AOM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
10/28/16	DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 10/28/16 [For RPC or Simulator Solution Change Place Label Here]	M. Polin

DATAMASTER DMT MAINTENANCE LOG

Waco
Department/Agency

100154
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
1/15/17	DMT External Standard Solution Lot # 16-15-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 7/20/2016 Expires: 7/20/2017 OPENED: 1/15/17 [For RPC e Label Here]	M. Paulin
3/28/17	[For RPC or Simulator Solution Change Place Label Here] DMT returned to VFL for maintenance JSD 3/28/17	Jeff Durette VFL
3/31/17	DMT External Standard Solution Lot # 16-35-100 EtOHaq Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 11/14/2016 Expires: 11/14/2017 OPENED: 3/31/17 JSD [For RPC Label Here] Install post sim conversion JSD 3/31/17	Jeff Durette VFL
6/13/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 6/13/17 M.P. [For RPC se Label Here] RPC	M. Paulin

DATAMASTER DMT MAINTENANCE LOG

Waco
Department/Agency

100154 129 061
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/17/17	DMT Simulator Solution Lot # 17-13-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/28/2017 Expires: 02/28/2018 OPENED: 10/17/17 MR. [For RPC or Simulator Solution Change Place Label Here] .102	MR. Paulin 8/18/17
10/20/17	Installed version 2.08 PIC chip and completed APM protocol. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL TES RCO 10/20/17
2/21/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 2/21/18 [For RPC or External Std Solution Change Place Label Here] RPC .101	MR. Paulin
6/4/18	DMT External Std Solution Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 6/4/18 MR. [For RPC or External Std Solution Change Place Label Here] RPC .01	MR. Paulin

DATAMASTER DMT MAINTENANCE LOG

Waco CSD
Department/Agency

100154
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/31/18	<u>Completed APM Protocol</u> [For RPC or Simulator Solution Change Place Label Here]	<u>R. Driscoll</u> VFL
10/16/18	<u>RPC</u> [For RP DMT External Std Solution Place Label Here] Lot: 17-24-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/14/17 Exp Date: 11/14/18 Opened: <u>10/16/18</u> .100	<u>M. Paulin</u>
2/13/19	<u>RPC</u> [For RPC DMT External Std Solution Place Label Here] Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: <u>2/13/19 M.P.</u> .101	<u>M. Paulin</u>
6/16/19	<u>RPC</u> [For RPC DMT External Std Solution Place Label Here] Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: <u>6/16/19 M.P.</u> .101	<u>M. Paulin</u>

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/30/19	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFL)
10/5/19	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 10/5/19 RPC	M Poolin
2/11/2020	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 2/11/2020 RPC	M Poolin
4/2/2020	DMT External Std Solution Lot: 19-07-100 EtOH Certified Conc. 0.103 Range: 0.098 - 0.108 Prep Date: 5/31/19 Exp Date: 05/31/20 Opened: 4/2/2020 RPC	M Poolin
6/2/2020	DMT External Std Solution Lot: 19-10-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 6/17/19 Exp Date: 06/17/20 Opened: 6/2/2020 M.P. .101	M Poolin
7/13/2020	Updated software and completed [For RPC or Simulator Solution Change Place Label Here] APM protocol. Removed operator log book	J Dukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/5/2020	DMT External Std Solution Lot: 20-05-100 EtOH Certified Conc. 0.101 [For RPC Range: 0.096 - 0.106 ce Label Here] Prep Date: 4/27/20 Exp Date: 04/27/21 Opened: 10/5/2020 WP .100	WP
2/29/21	DMT External Std Solution Lot: 20-05-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 4/27/20 Exp Date: 04/27/21 Opened: 2/29/21 WP	WP
6/4/21	DMT External Std Solution Lot: 21-05-100 EtOH Certified Conc. 0.100 [For RPC Range: 0.095 - 0.105 ce Label Here] Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: 6/4/21 .100	WP
7/9/2021	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
10/12/2021	DMT External Std Solution Lot: 21-05-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: 10-12-2021 .100	B. Meyer
2-8-2022	DMT External Std Solution [For RPC Lot: 21-05-100 EtOH Certified Conc. 0.100 ce Label Here] Range: 0.095 - 0.105 Prep Date: 3/3/21 Exp Date: 03/03/23 Opened: 2-8-2022 .100	B Meyer

DATAMASTER DMT MAINTENANCE LOG

Waco

Department/Agency

100154

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/30/2022	DMT External Std Solution Lot: 22-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/26/22 Exp Date: 01/26/24 Opened: 6-30-2022 [For RPC e Label Here]	B Meyer
7/18/22	Completed APM [For RPC or Simulator Solution Change Place Label Here]	J Dulake (UFL)
10/11/22	DMT External Std Solution Lot: 22-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/26/22 Exp Date: 01/26/24 Opened: 10/11/2022 [For RP ce Label Here] Oct RPC	E Grant
01/23/23	DMT External Std Solution Lot: 22-04-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 1/26/22 Exp Date: 01/26/24 Opened: 01/23/23 [For RF ace Label Here] Feb RPC	E Grant

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/05/2023	DMT External Std Solution [For RPC Lot: 22-04-100 EtOH Certified Conc. 0.100 e Label Here] Range: 0.095 - 0.105 Prep Date: 1/26/22 Exp Date: 01/26/24 Opened: 06/05/2023 June RPC	E Grant
6/16/23	DMT External Std Solution [For RPC Lot: 23-06-100 EtOH Certified Conc. 0.100 e Label Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 6/16/23 Install post TSI	J Dukette (VFL)
7/6/23	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
9/18/23	DMT External Std Solution [For R Lot: 22-12-100 EtOH Certified Conc. 0.101 Place Label Here] Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 9/18/23 Install completed post flood relocation	J Dukette (VFL)

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____

DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/11/2023	DMT External Std Solution Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: 10/11/2023 <i>Emel</i> [For RPC or Simulator Solution Change Place Label Here]	<i>E. Grant,</i> <i>Deputy</i> <i>October RPC</i>
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

[illegible]