DATAMASTER DMT MAINTENANCE LOG

Rutland VSP
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
12/01/11	**DataMaster Simulator Solution** Lot # 12-32-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 [Change Place Label Here] OPENED: 12/01/11 Install Per Protocol	Steven Harnais 12-1-11 Prog. Chief Org Chem VDH L.
1/4/11	**DataMaster Simulator Solution** Lot # 12-38-100 $EtOH_{aq}$ Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 11/17/2011 Expires: 11/17/2012 [For RPC ... ce Label Here] OPENED: Solution change	Blake Cushing
2/10/12	**DataMaster Simulator Solution** Lot # 12-32-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 10/26/2011 Expires: 10/26/2012 [For RPC ... e Label Here] OPENED: RPC	Blake Cushing
4/15	**DataMaster Simulator Solution** Lot # 12-58-100 $EtOH_{aq}$ Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 02/09/2012 Expires: 02/09/2013 [For RPC ... ce Label Here] OPENED: Solution Change	Blake Cushing

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP- Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/15/12	Update Software & Lock change [For RPC or Simulator Solution Change Place Label Here]	A. Bor
6/19/12	**DataMaster Simulator Solution** Lot # 12-77-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/08/2012 Expires: 05/08/2013 [For RF ... ace Label Here] OPENED: RPC	Blake Cushing
8/15/12	[For RPC or Simulator Solution Change Place Label Here] APM Performed	R. Driscoll
10/23/12	**DataMaster Simulator Solution** Lot # 12-77-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 05/08/2012 Expires: 05/08/2013 [For R ... Place Label Here] OPENED: RPC	Blake Cushing

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP- Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
2/24/13	[For RPC **DataMaster Simulator Solution** Lot # 12-88-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 11/9/2012 Expires: 11/9/2013 OPENED: e Label Here] RPC	Blake Cushing
3/14/13	[For RPC or Simulator Solution Change Place Label Here] Printer not communicating with DataMaster. Tried resetting printer + rebooting DMT with no luck. Tried new USB cable with same results. Replaced printer with HP 6100. Working again.	R. Driscoll DMT TECH.
6/25/13	[For RP(**DataMaster Simulator Solution** Lot # 12-89-100 EtOHaq Conc: 0.097 Range: 0.092 - 0.102 Preparation Date: 11/21/2012 Expires: 11/21/2013 OPENED: ce Label Here] RPC	Blake Cushing
7/12/13	[For RPC or Simulator Solution Change Place Label Here] Performed APM protocol	R. Driscoll DMT TECH.

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP-Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
10/28/13	**DataMaster Simulator Solution** Lot # 13-03-100 EtOHaq Conc: 0.098 [For RPC Range: 0.093 - 0.103 ce Label Here] Preparation Date: 04/03/2013 Expires: 04/03/2014 OPENED: RPC	Blake Cushing
1/6/14	[For RPC or Simulator Solution Change Place Label Here] Change Time from 9:53 to 10:05	Blake Cushing
2/8/14	**DataMaster Simulator Solution** Lot # 13-16-100 EtOHaq Conc: 0.099 [For RPC Range: 0.094 - 0.104 e Label Here] Preparation Date: 9/26/2013 Expires: 09/26/2014 OPENED: RPC	Blake Cushing
6/30/14	**DataMaster Simulator Solution** Lot # 13-21-100 EtOHaq Conc: 0.099 [For RPC Range: 0.094 - 0.104 e Label Here] Preparation Date: 12/20/2013 Expires: 12/20/2014 OPENED: RPC	Blake Cushing

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP - Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
7/10/14	Completed AAM Protocol. Noticed incorrect [For RPC or Simulator Solution Change Place Label Here] Lot number entered into DMT during June RPC. Concentration of .099 is correct. DMT ~~supervisor~~ TECH. RD 7/10/14 will complete solution change protocol to input correct lot #. RD 7/10/14	R. Driscoll DMT TECH. DMT TECH.
10/28/14	[For RPC DataMaster Simulator Solution Lot # 14-05-100 EtOHaq Conc: 0.098 Range: 0.093 - 0.103 Preparation Date: 6/19/2014 Expires: 6/19/2015 OPENED: Label Here] RPC	Blake Cushing
2/5/15	[For RP DataMaster Simulator Solution Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: ce Label Here] Solution Change	Blake Cushing
2/5/15	[For RPC or Simulator Solution Change Place Label Here] RPC	Blake Cushing

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP-Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
04/07/15	[For RPC **DataMaster Simulator Solution** Lot # 14-14-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 10/16/2014 Expires: 10/16/2015 OPENED: April 07, 2015 [illegible] ; Label Here] Solution Change	Steven Schutt Trooper
6/25/15	[For RP **DataMaster Simulator Solution** Lot # 15-05-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 2/19/2015 Expires: 2/19/2016 OPENED: ace Label Here] RPC	Blake Cushing
8/21/15	Downloaded data to USB drive, uploaded new [For RPC or Simulator Solution Change Place Label Here] software, installed new comm. PCB, and connected to DPS Network.	R. Driscoll DMT Tech.
8/25/15	[illegible] **DataMaster Simulator Solution** Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 OPENED: 8/25/15 RCD [For RP ace Label Here] Installation post TSE	R. Driscoll DMT Tech.

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP-Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
9/27/15	[For RPC ... ce Label Here] **DataMaster Simulator Solution** Lot # 15-09-100 EtOHaq Conc: 0.099 g/210L Range: 0.094 - 0.104 g/210L Preparation Date: 6/24/2015 Expires: 6/24/2016 OPENED: RPC Because of Solution Interference Detected	Blake Cushing Fixed & passed
11/1/15	[For RPC ... ce Label Here] **DataMaster Simulator Solution** Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: RPC	Blake Cush
2/29/16 2/29/16	[For RPC ... Label Here] **DataMaster Simulator Solution** Lot # 15-13-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 09/15/2015 Expires: 09/15/2016 OPENED: RPC	Blake Cushing
6/30/16	[For RPC ... ce Label Here] **DataMaster External Standard Solution** Lot # 16-03-100 EtOHaq Conc: 0.099 Range: 0.094 - 0.104 Preparation Date: 2/2/2016 Expires: 2/2/2017 OPENED: RPC	Blake Cushing

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
8/11/16	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll DMT Tech.
11/2/16	**DMT External Standard Solution** Lot # 16-23-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/19/2016 Expires: 10/19/2017 OPENED: [For RPC ce Label Here] RPC	Blake Cushing
2/5/17	**DMT External Standard Solution** Lot # 16-23-100 EtOHaq Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 10/19/2016 Expires: 10/19/2017 OPENED: 2-5-17 [For RPC ce Label Here] RPC	Steven Schutt
3/17/17	**DMT Simulator Solution** Lot # 17-10-100 EtOHaq Certified Conc: 0.101 Range: 0.096 - 0.106 Preparation Date: 02/06/2017 Expires: 02/06/2018 OPENED: 3/17/17 RD [For RPC ce Label Here] Install post simulator update	R. Driscoll DMT Tech.

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP/Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/15/17	[For RPC **DMT Simulator Solution** Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018 OPENED: 6-15-17 (88) 423 e Label Here] RPC	Steven Schutt VSP - 423
6/21/17	Completed APM and simulator temp check. [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
10/14/17	[For RPC **DMT Simulator Solution** Lot # 17-18-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 05/02/2017 Expires: 05/02/2018 OPENED: 10-14-17 (88) 423 ce Label Here] RPC	Steven Schutt VSP - 423
10/16/17	Replaced RIC chip with version 2.08 [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
3/3/18	[For RPC c **DMT Simulator Solution** Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 Label Here] OPENED: 3-3-18 - SD 423 RPC	Steven Schutt VSP-423
6/20/18	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	R. Driscoll VFL
7/1/18	[For RF **DMT External Std Solution** ace Label Here] Lot: 18-01-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 1/29/18 Exp Date: 01/29/19 Opened: 7/1/18 406	Eden Neary VSP-406
11/17/18	[For RPC **DMT Simulator Solution** Lot # 17-21-100 EtOHaq Certified Conc: 0.100 Range: 0.095 - 0.105 Preparation Date: 09/20/2017 Expires: 09/20/2018 e Label Here] OPENED: SD 11-17-18 423 RPC	Steven Schutt VSP-423

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP- Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
11/22/18	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 18-12-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/11/18 Exp Date: 07/11/19 Opened: 11-22-18 423 RPC	Steven Schott VSP-423
3/18/19	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 19-04-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 3-18-19 423 RPC	Steven Schott VSP-423
5/29/19	Replaced printer [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VFC)
6/11/19	[For RPC or Simulator Solution Change Place Label Here] DMT External Std Solution Lot: 19-06-100 EtOH Certified Conc. 0.098 Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened: 6-11-19 423	Steven Schott VSP-423

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____ DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/14/19	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VRL)
10/13/19	**DMT External Std Solution** [For RPC **Lot:** **19-04-100** **EtOH Certified Conc.** **0.101** e Label Here] Range: 0.096 - 0.106 Prep Date: 2/28/19 Exp Date: 02/28/20 Opened: 10-13-19 88 423 RPC	Steven Schott VSP-423
2/20/20	**DMT External Std Solution** [For RPC **Lot:** **19-06-100** **EtOH Certified Conc.** **0.098** ce Label Here] Range: 0.093 - 0.103 Prep Date: 4/3/19 Exp Date: 4/3/20 Opened: 2-20-20 88 423 RPC	Steven Schott VSP-423
6/8/2020	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (VRL)

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____ DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/28/20	DMT External Std Solution [For RP ... ace Label Here] Lot: 20-05-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 4/27/20 Exp Date: 04/27/21 Opened: 6-28-20 88 423 RPC	Steven Schutt VSP-423
7/14/20	Software Update [For RPC or Simulator Solution Change Place Label Here]	[illegible]
11/6/20	DMT External Std Solution [For RPC ... Label Here] Lot: 20-11-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 7/9/20 Exp Date: 07/09/22 Opened: 11-6-20 88 423 RPC	Steven Schutt VSP-423
2/6/21	DMT External Std Solution [For RPC ... e Label Here] Lot: 20-15-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 11/20/20 Exp Date: 11/20/22 Opened: 2/6/21 88 423 RPC	Steven Schutt VSP-423

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

VSP - Rutland
Department/Agency

100155
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/21/2021	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
6/29/21	DMT External Std Solution [For RPC c Label Here] Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened: 6/29/21 SS 423 RPC	Steven Schott - 423
10/27/21	DMT External Std Solution [For RPC ce Label Here] Lot: 21-06-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 3/25/21 Exp Date: 03/25/23 Opened:	R. Gardner -450
1/30/22	[For RPC DMT External Std Solution ce Label Here] Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: 1-30-22 SS RPC	Steven Schutt - 423

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency _____ DataMaster DMT Serial Number _____

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/1/2022	Completed APM protocol [For RPC or Simulator Solution Change Place Label Here]	S. Stratton (VFL)
7/21	**DMT External Std Solution** Lot: 22-12-100 EtOH Certified Conc. 0.101 Range: 0.096 - 0.106 Prep Date: 5/31/22 Exp Date: 05/31/24 Opened: AB Install post TSI	A. Boldue
7/21	dummy APM done to satisfy [~~For RPC or Simulator Solution Change Place Label Here~~] Software new board, see TSI	AB
10/12/22	[For RPC or Simulator Solution Change Place Label Here] **DMT External Std Solution** Lot: 21-09-100 EtOH Certified Conc. 0.099 Range: 0.094 - 0.104 Prep Date: 6/4/21 Exp Date: 06/04/23 Opened: 10/12/22 423	S. Schutt VSP 423
2/2/23	[For RPC ce Label Here] **DMT External Std Solution** Lot: 22-10-100 EtOH Certified Conc. 0.100 Range: 0.095 - 0.105 Prep Date: 4/1/22 Exp Date: 04/01/24 Opened: 2/2/23 423 DMT 100155 removed from agency / 6/8/23	S. Schutt VSP 423

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Rutland USP
Department/Agency

123106
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
6/8/23	DMT External Std Solution [For RPC or Lot: 23-06-100 EtOH Certified Conc. 0.100 abel Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 850 6/8/23 Install post TSI. New DMT installed. Unable to connect to ethernet, APM also performed	J Dukette (UFL)
10/29/23	DMT External Std Solution [For RPC Lot: 23-06-100 EtOH Certified Conc. 0.100 e Label Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 10-29-23 RPC	S. Schutt 4/23
02/29/24	DMT External Std Solution [For R Lot: 23-06-100 EtOH Certified Conc. 0.100 lace Label Here] Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 02/29/24 RPC	R. Gardner 450
5/22/24	Completed APM Protocol [For RPC or Simulator Solution Change Place Label Here]	J Dukette (UFL)

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

123106
DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
06/22/24	[For R **DMT Quality Control Solution** lace Label Here] Lot: 24-02-100 EtOH Certified Conc. 0.099 Prep Date: 4/9/24 Exp Date: 04/09/26 Opened: 06/22/24 R. Gardner RPC	R. Gardner 450
11/03/2024	[For **DMT Quality Control Solution** Place Label Here] Lot: 24-02-100 EtOH Certified Conc. 0.099 Prep Date: 4/9/24 Exp Date: 04/09/26 Opened:	R. Gardner Trooper 450
2/8/25	RPC [For RPC **DMT External Std Solution** ce Label Here] Lot: 23-06-100 EtOH Certified Conc. 0.100 Prep Date: 5/1/23 Exp Date: 05/01/25 Opened: 2/8/25 SS	S. Schutt 423
5/19/25	Remove for refurb [For RPC or Simulator Solution Change Place Label Here]	A. Bolduc

ALC 803 REV 03 08/2011

DATAMASTER DMT MAINTENANCE LOG

Department/Agency

DataMaster DMT Serial Number

DATE	TYPE OF SERVICE PERFORMED	PERFORMED BY (Name and Title)
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	
	[For RPC or Simulator Solution Change Place Label Here]	

ALC 803 REV 03 08/2011