

VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>



*100156 (VSP-CASTLETON)
RELOCATING TO FAIR
HAVEN PD. DEC 12/31/12

DataMaster DMT Site Evaluation Checklist

Agency or Site Information	
Name:	Fair Haven PD
Address:	3 North Park Place Fair Haven, VT, 05743
Phone:	(802) 265-8293 / 265-7839
Fax:	

Personnel Information	
Head of Site:	William Humphreys
Current DataMaster Supervisor (s):	Blake Cushing
Proposed DataMaster Supervisor(s):	Blake Cushing ^{RED} 12/20/12

(Proposed) DataMaster DMT Environment		
Limited Controlled Access:	☒ Y	N
Low Activity Area:	☒ Y	N
Unlimited Access by Other Law Enforcement Personnel	☒ Y	N
Access by VFL Technical Services	☒ Y	N
Air Conditioned	☒ Y	N
Room Temperature 65° - 85° F	☒ Y	N
Adequate Ventilation	☒ Y	N
Free of Cleaning Supplies	☒ Y	N
Clean	☒ Y	N

(Proposed) DataMaster DMT Connection(s)		
Power:		
Min 15 Amp	☒ Y	N
Limited low power sharing:	☒ Y	N
Phone Line Connection Number:	- - -	
Phone Line Connection Type		

(Proposed) DataMaster DMT Site Maintenance Agreement		
Will notify VFL Alcohol Program staff of maintenance and supply needs promptly:	☒ Y	N
Will notify VFL Alcohol Program staff of changes in building conditions or moves promptly:	☒ Y	N
Will notify VFL Alcohol Program staff of relevant personnel changes at site promptly:	☒ Y	N

VFL Representative:	Robert Driscoll		12/20/12
	Print Name	Signature	Date
Agency Representative:	William Humphreys		12/20/12
	Print Name	Signature	Date