



VERMONT FORENSIC LABORATORY

PO BOX 47
WATERBURY, VERMONT 05676-0047
TEL: 802-244-8788
FAX: 802-241-5557
<http://vfl.vermont.gov>

DATAMASTER TECHNICAL SUPPORT INQUIRY

Agency/Site: Fair Haven PD – Rutland VSP DataMaster Serial #: 100156

Date: 7/15/15 Requestor: VFL Alcohol Program Staff

Instrument Supervisor: Dale Kerber Phone #: (802) 265-7839

Taken By: Robert Driscoll

Problem: Detector voltage at filter one is above 0.600V.

Laboratory Response: Alcohol Program staff will bring instrument to the Lab to optimize the detector voltage.

Work Performed: Optimized detector voltage at filter one to within +/- 0.100V with lamp and cooler voltage adjustments. Installed a new communication PCB and loaded updated software. Data was downloaded to a USB drive and then transferred into DM Host. The instrument has been Calibrated, Certified, and is now ready for installation at the agency.

☐ On Site ☒ In-house

Attachments: 4 pages

Performed By: Robert Driscoll *RD*

Date: 7/16/15

Final Review: *[Signature]*

Date: 7/16/15

CALIBRATION REPORT

DataMaster DMT: 100156
Calibration Date: 07/16/2015
Calibrated by: ROBERT DRISCOLL
Lot: 15-08-100



Ca	=	0.100		
CAL	=	0.987816	0.800	<= CAL < 1.200
b1	=	0.000	0.000	<= b1 < 0.004
b2	=	0.004	0.002	<= b2 < 0.010
b3	=	0.000	0.000	<= b3 < 0.004
Xq	=	0.086	0.050	<= Xq < 0.200
a21	=	1.192566	1.050	<= a21 < 1.300
a31	=	0.456757	0.300	<= a31 < 0.800

Performed by

766 10

Date

07/16/2015



ACCURACY & PRECISION REPORT

STATE OF VERMONT

DataMaster DMT: 100156

Date: 07/16/2015

Time: 08:01:46

SUPERVISOR NAME:

RCD CAL CHECK

SOLUTION LOT #: GUTH 14110


SOLUTION CONCENTRATION: 0.100

BLANK TEST	0.000	08:02
CALIBRATION CHECK	PASSED	08:02
SIMULATOR VAPOR 34.0°C	0.097	08:03
SIMULATOR VAPOR 33.9°C	0.097	08:04
SIMULATOR VAPOR 33.9°C	0.098	08:05
SIMULATOR VAPOR 34.0°C	0.098	08:06
SIMULATOR VAPOR 34.0°C	0.098	08:07
SIMULATOR VAPOR 34.0°C	0.098	08:08
SIMULATOR VAPOR 34.0°C	0.098	08:09
SIMULATOR VAPOR 34.0°C	0.098	08:10
SIMULATOR VAPOR 33.9°C	0.098	08:11
SIMULATOR VAPOR 34.0°C	0.098	08:12
BLANK TEST	0.000	08:13

Average = 0.097

Std Dev = 0.0004

NIST Traceable Thermometer SN
111214512 33.9°C
RCD 7/16/15



CERTIFICATION REPORT

DataMaster DMT: 100156
Calibration Date: 07/16/2015
Certification Date: 07/16/2015
Certified by: ROBERT DRISCOLL



Diagnostic Results

VERSIONS

DMT: 2.00
PIC: 2.05
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 48.9°C
Breath Tube = 45.6°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.96 V
Cooler Voltage = 1.89 V
Bias Voltage = 80 V
Chopper Freq = 538 Hz

PUMP INFO

Flow Rate = 5.975 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.0114	0.0150
MIN (V)	0.0100	0.0136

FILTER INFO

Filter 1	0.014	Zero = true
Filter 2	0.144	Zero = true
Filter 3	1.131	Zero = true

CALIBRATION CHECK

Xq = 0.086 0.10%

Options

OPTIONS

Units

Alcohol..... g/210L

Simulator

Tolerance Check..... yes
Standard Type..... wet
Nominal..... 0.100
Digital Simulator..... Guth34CNP

Subject

Ask Questions..... yes
Number of Tests..... 2
Alcohol Display..... yes
Volume Display..... yes
Query Refusal..... yes
Copies..... 3
Simulator Before..... yes
Simulator Between..... yes
Simulator After..... no
Observation Time..... 0

Supervisor

Number of Tests..... 10

Calibration

Standard Type..... wet
Nominal..... 0.100
Nominal (Dry Gas)..... 0.100
Number of Tests..... 1

Printer

Printer On..... yes

Software Configuration

Data Collection..... yes

Linearity Check Results

True Value	Reported Average	Std Dev
0.020 g/210L	0.019 g/210L	0.0003
Lot # 14-13-020		
0.080 g/210L	0.080 g/210L	0.0003
Lot # 15-04-080		
0.159 g/210L	0.161 g/210L	0.0000
Lot # 15-03-160		
0.398 g/210L	0.395 g/210L	0.0009
Lot # 14-09-400		

$R^2 = 0.9999$

Acetone Interference Test

Lot # 15-04-08A1

Interference Detected

Mouth Alcohol Test

Mouth Alcohol Detected




RF Detection Test

Passed

Sample Acceptance Test

Passed

CERTIFICATION PASSED



Performed by

AS RCD

Date

07/16/2015

Reviewed by

AS

Date

7/16/15

DMT Serial Number #100156

Page 2 of 2

07/16/2015 9:07 AM

5 of 5 RCD