

SOLUTION CHANGE

DataMaster DMT: 100156
Location: FAIR HAVEN PD / RUTLAND VSP
Date: 06/21/2016
Performed by: DALE H KERBER



Accuracy and Precision Check

Concentration = 0.099 g/210L
Lot # = 15-13-100
Range = 0.094 - 0.104
Average = 0.100 g/210L
Std Dev = 0.0005

RCO 6/22/16

Simulator Temperature: 34.0°C

Performed by

Dale H Kerber

Date

06/21/2016