

# ANNUAL PREVENTATIVE MAINTENANCE

DataMaster DMT: 100156  
Location: FAIR HAVEN PD  
Calibration Date: 03/10/2017  
Certification Date: 03/10/2017  
Installation Date: 03/13/2017  
APM Date: 06/20/2018  
Performed by: ROBERT DRISCOLL



## Diagnostic Results

### VERSIONS

DMT: 2.03  
PIC: 2.08  
Modem: 2.6  
Questions: 2.1

### TEMPERATURES

Sample Chamber = 48.7°C  
Breath Tube = 44.9°C  
Digital Sim = 34.0°C

### SETTINGS

Lamp Voltage = 1.64 V  
Cooler Voltage = 1.71 V  
Bias Voltage = 80 V  
Chopper Freq = 537 Hz

### PUMP INFO

Flow Rate = 5.526 L/M

### DETECTOR INFO

PUMP	ON	OFF
MAX (V)	0.1754	0.1795
MIN (V)	0.1732	0.1770

### FILTER INFO

Filter	Value	Zero
Filter 1	0.177	true
Filter 2	0.451	true
Filter 3	1.128	true

### CALIBRATION CHECK

Xq = 0.098 0.27%

APM PASSED

## Simulator Vapor Test


Concentration = 0.100 g/210L  
Lot # = 18-05-100  
Range = 0.095 - 0.105  
Simulator Vapor = 0.100 g/210L

## RF Detection Test

Passed

## Sample Acceptance Test

Passed



Performed by

RD

Date

06/20/2018

Reviewed by

RD

Date

6/27/18

VERMONT FORENSIC LABORATORY

Simulator Temperature Check

Simulator Serial Number:

MP3474

DMT Serial Number / Simulator Location:

100156 / Fair Haven PD

NIST Traceable Thermometer Serial Number:

160294076

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	34.01	34.00	0.01	Y

\* 34.0 +/- 0.2 degree Celsius acceptable range

Was temperature adjustment performed?

Y / (N)

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By:



Date:

6/20/18

Reviewed by:



Date:

6/21/18