

ANNUAL PREVENTATIVE MAINTENANCE

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Calibration Date: 03/10/2017
Certification Date: 03/10/2017
Installation Date: 03/13/2017
APM Date: 06/14/2019
Performed by: JEFFERY DUKETTE



Diagnostic Results

VERSIONS

DMT: 2.05
PIC: 2.08
Modem: 2.6
Questions: 2.1

TEMPERATURES

Sample Chamber = 49.1°C
Breath Tube = 46.5°C
Digital Sim = 34.0°C

SETTINGS

Lamp Voltage = 1.64 V
Cooler Voltage = 1.71 V
Bias Voltage = 80 V
Chopper Freq = 539 Hz

PUMP INFO

Flow Rate = 5.454 L/M

DETECTOR INFO

PUMP	ON	OFF
MAX(V)	0.2587	0.2624
MIN(V)	0.2569	0.2602

FILTER INFO

Filter	Value	Zero
Filter 1	0.262	true
Filter 2	0.532	true
Filter 3	1.191	true

CALIBRATION CHECK

Xq = 0.097 0.54%

APM PASSED

Simulator Vapor Test


Concentration = 0.100 g/210L
Lot # = 18-12-100
Range = 0.095 - 0.105
Simulator Vapor = 0.097 g/210L

RF Detection Test

Passed

Sample Acceptance Test

Passed



Performed by

 JSD

Date

06/14/2019

Reviewed by



Date

6/14/19

DMT Serial Number #100156

Page 1 of 1

06/14/2019 9:45 AM

AS 6/14/19 JSD

page 1 of 2

Simulator Temperature Check

Simulator Serial Number:

MP3474

DMT Serial Number / Simulator Location:

100156 / Fair Haven PD

NIST Traceable Thermometer Serial Number:

160294076

As Found

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator	33.99	34.01	0.02	Y

* 34.0 +/- 0.2 degree Celsius acceptable range

Was temperature adjustment performed?

Y / N

As Left (if applicable)

Item	Thermometer Temperature	Simulator Temperature	Difference	Within Tolerance?*
Guth 12V500 Simulator				

Notes:

Check Performed By:



Date:

6/14/2019

Reviewed by:



Date:

6/14/19 ~~6/14/19~~ 6/19/19 ^{JS}