

SOLUTION CHANGE

DataMaster DMT: 100156
Location: FAIR HAVEN PD
Date: 01/17/2020
Performed by: DALE H KERBER



Accuracy and Precision Check

Concentration = 0.098 g/210L
Lot # = 19-06-100
Range = 0.093 - 0.103
Average = 0.096 g/210L
Std Dev = 0.0004

SDS
01/21/2020

Simulator Temperature: 34.0°C

Performed by

Dale Kerber

Date

01/17/2020